可溶性转铁蛋白受体质控品说明书

【产品名称】
通用名称：可溶性转铁蛋白受体质控品
英文名称：Access sTfR QC

【包装规格】
水平 1（QC1）：2×2.5mL/瓶；水平 2（QC2）：2×2.5mL/瓶；水平 3（QC3）：2×2.5mL/瓶。

【预期用途】
本产品用于体外定量测定可溶性转铁蛋白受体时的质量控制。
用于监控Access sTfR测定的系统性能。

【检验原理】
质控品可模拟病人样本的诸多特性，且在监控Access sTfR免疫分析的系统性能时，显得十分重要。此外，它们是规范的实验室操作流程的一个组成部分。[45,51,52,53,54,55]在对sTfR的Access试剂进行检测时，需包括质控品以确认检测结果的准确性。若测试所使用的系统运行正常，那么所测试的值应在可接受的范围内。

【主要组成成分】

水平 1（QC1）:	人体可溶性转铁蛋白受体水平大约为 10nmol/L、牛血清白蛋白(BSA)缓冲基质、人血浆、0.1%叠氮钠以及 0.15%ProClin 300。
水平 2（QC 2）:	人体可溶性转铁蛋白受体水平大约为 25 nmol/L 、牛血清白蛋白(BSA)缓冲基质、人血浆、0.1% 叠氮钠以及 0.15% ProClin 300 。
水平 3（QC 3）:	人体可溶性转铁蛋白受体水平大约为 90 nmol/L 、牛血清白蛋白(BSA)缓冲基质、人血浆、0.1% 叠氮钠以及 0.15% ProClin 300 。
QC 值卡:	1

溯源性：

Access sTfR QC 内的分析物，可溯源至生产商的工作用校准品。溯源性的建立过程是按ENISO17511 进行的。

使用这一批次QC中具有代表性的样本，建立指定的值，该值对于Access试剂的测定方法是特定的。使用其它方法确定的值可能会存在差异。若此差异存在，可能是不同方法之间的偏倚造成的。

【储存条件及有效期】

- 在-20℃下保存，有效期 8 个月。
- 在-20℃环境下保存时，稳定期可持续至标签上所指明的有效期。
- 首次使用后，在 2~10℃环境下保存的稳定期可维持 90 天。
- 可能存在变质情况的表现为质控值超出范围。
- -20℃的环境下保存。仅可解冻一次。
- 生产日期和失效日期详见标签。

【适用仪器】

Access 2、UniCel DxI 800 Access免疫分析仪。

【检验方法】

1. 与测定病人样本一致，使用Access免疫分析系统来测定AccesssTfRQC品内sTfR的浓度。因为可在任意时间以“随机存取”形式，而非“批”形式对样品进行处理，则应将质控材料纳入各个24小时时间周期内。45由使用者按照非临床研究质量管理规范或实验室认证要求及适用法规，自行判断是否增加质控品的使用频率或使用额外的质控品。参考相应的系统操作手册和 / 或帮助系统，获取有关质控原理、设置质控品、质控样本测试申请的输入和复查质控数据的信息。
2. 对于Access sTfR QC 品值的指定，需选择整批内一些具有代表性的样本，并对它们进行检测以对均值做出可靠的预估。均值和标准差均列于QC值卡片上。但由于技术、设备或试剂的差异，可能会导致与所列的值存在差距。因此，各实验室需确立个别适用的均值和标准差（SD）。
3. 即用型。
4. 在室温下解冻，并在使用前轻轻倒转，混匀成份。防止形成气泡。
5. 参见QC值卡，以获取均值和标准差（SD）信息。

【检验方法的局限性】

与Access免疫测定系统及配套试剂使用。

试剂内如有微生物污染的迹象或显著浑浊，应弃去。

【注意事项】

- 供体外诊断使用。
- 按所描述的步骤进行操作，可使日常处理病人样本和血制品中所带来的风险降到最低。然而，无论这些产品的来源、处理方式或先前获得的证明，使用这些具有潜在传染性的物品均须按统一的注意事项和完善的临床实验室操作规范来进行。应使用合适的消毒剂来消除污染。遵循当地的操作规章和导则来保存并处理这些物品以及盛放这些物品的容器。
- 经过测试发现用于制备试剂的人源性材料，对于乙型肝炎、丙型肝炎（HCV）、和人体免疫缺陷病毒（HIV-1 和 HIV-2），均显示为阴性或无反应。由于任何已知的测试方法均无法完全确保不存在感染性物质，因此处理试剂与病人样本时均视其具有传染疾病的可能。[50]
- 叠氮钠可能会与铅以及铜制的管件发生反应，形成具有高危爆炸性的金属叠氮物。在倾

弃这种液体时，应以大量的水进行冲洗，防止叠氮物的集结。[43]

- Xi.刺激物：0.15%ProClin 300。

R43：如接触皮肤可引起致敏作用。

S28-37：接触到皮肤后应立即用大量肥皂和水冲洗。应戴适用的手套。

- Xn.有害物：≥0.1%NaN_3。

R22：如吞入，对身体有害。

S28：如果接触到皮肤，请立即用大量水冲洗。

- 材料安全数据资料（MSDS）可按需索取。

【标识的解释】

制造商

欧盟授权代表

生产日期

批次代码

产品编号

温度极限

温度上限

温度下限

体外诊断医疗器械

警告

阴性对照

阳性对照

有效期

生物风险

含量足够测试<n>次

查阅使用说明

怕晒

【参考文献】

1. Harford JB. Rouault, TA, Huebers, HA, and Klausner RD. Molecular mechanisms of iron metabolism in The Molecular Basis of Blood Diseases, 2nd edition, Stamatoyannopoulos, G. et al.

eds., Saunders, Philadelphia, 1994; 351.
2. BeguinY. The soluble transferrin receptor: biological aspects and clinical usefulness as quantitative measure of erythropoiesis. Haematologica 1992; 77: 1.
3. Omary MB, and Trowbridge IS. Biosynthesis of the human transferrin receptor in cultured cells. J Biol Chem.1981; 256:12888.
4. Schneider C, Owen MJ, Banville D, Williams JG. Primary structure of human transferrin receptor deduced from the mRNA sequence. Nature 1984; 311: 675.
5. Seligman PA, Schleicher RB, Allen RH. Isolation and Characterization of the Transferrin Receptor From Human Placenta J Biol Chem 1979; 254: 9943.
6. McClelland A, Kühn LC, Ruddle FH. The human transferrin receptor gene: Genomic organization, and the complete primary structure of the receptor deduced for a cDNA sequence. Cell 1984; 39: 267.
7. Jing SQ and Trowbridge IS. Identification of the intermolecular disulfide bonds of the human transferrin receptor and its lipid-attachment site, EMBO J. 1987; 6:327.
8. Rao KK, Shapiro D, Mattia E, Bridges K, Klausner R. (1985) Effects of alterations in cellular iron on biosynthesis of the transferrin receptor in K562 cells, Mol. Cell. Biol. 1985; 5: 595.
9. Müllner EW and Kuhn LC. A stem-loop in the 3' untranslated region mediates iron-dependent regulation of transferrin receptor mRNA stability in the cytoplasm, Cell 1988; 53: 815.
10. Koeller DM, Casey JL, Hentze MW, Gerhardt EM, Chan LN, Klausner RD, Harford JB. A cytosolic protein binds to structural elements within the iron regulatory region of the transferrin receptor mRNA, Proc. Natl. Acad. Sci. USA 1989; 86: 3574.
11. Thorstensen K and Romslo I. The transferrin receptor: its diagnostic value and its potential as therapeutic target, Scand. J. Lab. Invest. 1993; 53(Suppl. 215): 113.
12. Shih YJ, Baynes RD, Hudson BG, Flowers CH, Skikne BS, Cook JD. Serum transferrin receptor is a truncated form of tissue receptor, J. Biol. Chem. 1990; 265: 19077.
13. Baynes RD, Shih YJ, Cook JD. Production of soluble transferrin receptor by K562 erythroleukemia cells, Proc. Soc. Exp. Med. 1993; 204: 65.
14. Beguin Y, Huebers HA, Josephson B, Finch CA. Transferrin receptors in rat plasma, Proc. Natl. Acad. Sci. USA 1988; 85: 637.
15. Cook JD, Skikne BS, Baynes RD. Serum transferrin receptor, Annu. Rev. Med. 1993; 44: 63.
16. Kohgo Y, Niitsu Y, Kondo H, Kato J, Tsushima N, Sasaki K, Hirayama M, Numata T, Nishisato T, Urushizaki I. Serum transferrin receptor as a new index of erythropoiesis, Blood 1987; 70: 1955.
17. Flowers CH, Skikne BS, Covell AM, Cook JD. The clinical measurement of serum transferrin receptor, J. Lab. Clin. Med. 1989; 114: 368.
18. Ferguson BJ, Skikne BS, Simpson KM, Baynes RD, Cook JD. Serum transferrin receptor distinguishes the anemia of chronic disease from iron deficiency anemia, J. Lab. Clin. Med. 1992;119: 385.
19. Huebers HA, Beguin Y, Pootrakul P, Einspahr D, Finch CA. Intact transferrin receptors in human plasma and their relation to erythropoiesis, Blood 1990; 75: 102.
20. Ahluwalia N, Lammi-Keefe CJ, Bendel RB, Morse EE, Beard JL, Haley NR. Iron deficiency and anemia of chronic disease in elderly women: a discriminate-analysis approach for differentiation, Am. J. Clin. Nutr. 1995; 61: 590.
21. Skikne BS, Flowers CH, Cook JD. Serum transferrin receptor: a quantitative measure of tissue iron deficiency, Blood 1990; 75: 1870.
22. Ragab L, Ibrahim HA, Eid AS, Kotb T, Konsowa MF. Suitability of soluble transferrin receptor for the clinical diagnosis of different types of anaemia in children. East Mediterr Health J 2002; 8: 298.
23. Erslev AJ. in Hematology, 5th edition, Beutler, E. et al.eds., McGraw-Hill, New York, 1995: 490.
24. Baer AN, Dessypris EN, Krantz SB. The pathogenesis of anemia in rheumatoid arthritis: a clinical and laboratory analysis, Sem. Arth. Rheum. 1990; 19: 209.
25. Punnonen K, Irjala K, Rajamaki A. Serum transferrin receptor and it ratio to serum ferritin in the diagnosis of iron deficiency, Blood 1997; 89: 1052.
26. O'Broin S, Kelleher B, Balfe A, Mc Mahon C. Evaluation of serum transferrin receptor assay in a centralized iron screening service. Clin. Lab. Haem. 2005; 27: 190–194.

27. Allen J, Backstrom KR, Cooper JA, Cooper MC, Detwiler TC, Essex DW, Fritz RP, Means, Jr. RT, Meier PB, Pearlman SR, Roitman-Johnson BR, Seligman PA. Measurement of soluble transferrin receptor in serum of healthy adults. Clinical Chemistry 1998; 44: 135–39.
28. Baillie FJ, Morrison AE, Fergus I. Soluble transferrin receptor: a discriminating assay for iron deficiency. Clin. Lab. Haem. 2003; 25: 353–357.
29. Markovic M, Majkic N, Singh, Subota V. Usefulness of soluble transferrin receptor and ferritin in iron deficiency and chronic disease. Scand J Clin Lab Invest 2005; 65:571–576.
30. Weiss G, Goodnough LT. Anemia of Chronic Disease. N Engl J Med 2005; 352:1011-23.
31. Suominen P, Punnonen K, Rajamäki A, Irjal K. Serum Transferrin Receptor and Transferrin Receptor-Ferritin Index Identify Healthy Subjects With Subclinical Iron Deficits. Blood 1998; 92(8): 2934-2939.
32. Genc S, Erten N, Karan MA, Besisik SK, Saka B, Tascioglu C, Sivas A. Soluble Transferrin Receptor and Soluble Transferrin receptor-Ferritin Index for Evaluation of the Iron Status in Elderly Patients. Tohoku J. Exp. Med. 2004; 202;135-142.
33. Verstraelen H, Delanghe J, Roelens K, Blot S, Claeys G, Temmerman M. Subclinical iron deficiency is a strong predictor of bacterial vaginosis in early pregnancy. BMC Infectious Diseases. 2005; 5: 55.
34. Means RT. Allen J. Sears DA, Schuster SJ. Serum soluble transferrin receptor and the prediction of marrow aspirate iron results in a heterogeneous group of patients. Clin. Lab. Haem. 1999; 21: 161-167.
35. Margetic S, Topic E, Ruzic DF, Kvaternik M. Soluble transferrin receptor and transferrin receptor-ferritin index in iron deficiency anemia and anemia in rheumatoid arthritis.Clin Chem Lab Med 2005; 43(3): 326–331.
36. Rimon E, Levy S, Sapir A, Gelzer G, Peled R, Sthoeger Z. Diagnosis of Iron Defiency Anemia in the Elderly By Transferrin Receptor-Ferritin Index. Arch Internal Med 2002; 162: 445-449.
37. Cook J. Diagnosis and Management of Iron Deficiency Anemia. Best Practice & Research Clinical Haematology 2005; 18:319-22.
38. Cermak J, Brabec V. Transferrin receptor-ferritin index: a useful parameter in differential diagnosis of iron deficiency and hyperplastic erythropoiesis. Eur J Haematol, 1998; 61: 210-212.
39. De Caterina1 M, Grimaldi E, Di Pascale G, Salerno G, Rosiello A, Passaretti1 M, Scopacasa F. The soluble transferrin receptor (sTfR)-ferritin index is a potential predictor of celiac disease in children with refractory iron deficiency anemia. Clin Chem Lab Med 2005; 43 (1): 38–42.
40. Chen YC, Hung SC, Tarng DC. Association Between Transferrin Receptor-Ferritin Index and Conventional Measures of Iron Responsiveness in Hemodialysis Patients. American Journal of Kidney Diseases 2006; 47: 1036-1044.
41. Agorasti A, Trivellas T, Papadopoulos V, Konstantinidou D. Innovative Parameters RET-Y, sTfR, and sTfR-F Index in Patients with Microcytic, Hypochromic Anemia—Their Special Value for Hemoglobinopathies, Laboratory Hematology 2007;13 (2): 63-68
42. Beguin Y. Soluble transferrin receptor for the evaluation of erythropoiesis and iron status. Clinica Chimica Acta 2003; 329: 9 –22.
43. DHHS (NIOSH) Publication No. 78-127, August 1976. Current Intelligence Bulletin 13 - Explosive Azide Hazard. Available http://www.cdc.gov/niosh.
44. Approved Guideline – Procedures for the Handling and Processing of Blood Specimens, H18-A3. 2004. Clinical and Laboratory Standards Institute.
45. embrowski GS, Carey RN. Laboratory quality management: QC→QA. ASCP Press, Chicago, IL, 1989.
46. Kricka L. Interferences in immunoassays – still a threat. Clin Chem 2000; 46: 1037–1038.
47. Bjerner J, et al. Immunometric assay interference: incidence and prevention. Clin Chem 2002; 48: 613–621.
48. Approved Guideline – Evaluation of Precision Performance of Quantitative Measurement Methods, EP5-A2. August 2004. National Committee for Clinical Laboratory Standards.
49. Approved Guideline – Interference Testing in Clinical Chemistry, EP7-A2. November 2005. Clinical and Laboratory Standards Institute.

50. HHS Publication, 4th ed., April 1999. Biosafety in Microbiological and Biomedical Laboratories. Available http://www.cdc.gov/od/ohs/biosfty/bmbl4/bmbl4toc.htm.
51. Broome HE, Cembrowski GS, Kahn SN, Martin PL, Patrick CA. Implementation and use of a manual multi-rule quality control procedure. Lab Med 1985; 16: 533-537.
52. Westgard JO, Barry PL, Hunt MR, Groth T. A multi-rule Shewhart chart for quality control in clinical chemistry. Clin Chem 1981; 27: 493-501.
53. Koch DD, Oryall JJ, Quam EF, Feldbruegger DH, et al. Selection of medically useful QC procedures for individual tests done in a multitest analytical system. Clin Chem 1990; 36: 230-233.
54. Mugan K, Carlson IH, Westgard JO. Planning QC procedures for immunoassays. J Clin Immunoassay 1994;17: 216-222.
55. Approved Guideline –Statistical Quality Control for Quantitative Measurement Procedures: Principles and Definitions, C24-A3. June 2006. Clinical and Laboratory Standards Institute.

【基本信息】
注册人/生产企业名称： Beckman Coulter, Inc. 贝克曼库尔特（美国）股份有限公司
住所：250 South Kraemer Blvd. Brea, CA 92821, USA
生产地址：1000 Lake Hazeltine Drive Chaska, MN 55318-1084 USA
联系方式： 001-800-742-2345
售后服务单位名称：贝克曼库尔特商贸（中国）有限公司
联系方式：021-38651000, 4008855355
代理人名称：贝克曼库尔特商贸（中国）有限公司
住所：中国（上海）自由贸易试验区峨山路 91 弄 98 号 3 层 01 单元
联系方式：021-38651000, 4008855355

【医疗器械注册证编号/产品技术要求编号】
国械注进20172400290

【说明书核准日期及修改日期】
2021-05-19
2022-06-30