



- GENERAL NOTES:
1. MATERIAL: **ALUMINUM, 6061-T651**
 2. SPECIAL FINISH: **NONE**
 3. SURFACE ROUGHNESS (UNLESS SPECIFIED OTHERWISE): **32 Ra** (AVERAGE MICRO-INCHES)
 4. INTERPRET DIMENSIONS AND TOLERANCES PER ASME Y14.5M-1994
 5. DEBURR AND BREAK ALL SHARP EDGES, MAX 0.010" (UNLESS SPECIFIED OTHERWISE)
 6. PARTS ARE TO BE CLEAN AND FREE OF OIL, GREASE, AND OTHER CONTAMINANTS
 7. DIMENSIONS INCLUDE CHEMICALLY APPLIED FINISHES IF APPLICABLE

- ALTERED ITEM NOTES:
1. THIS IS A PART MANUFACTURED BY ANOTHER VENDOR (DETAILS BELOW). THE DIMENSIONS SHOWN IN THIS DRAWING ARE REQUIRED MODIFICATIONS.
 2. ALL REFERENCE/INSPECTION DIMENSIONS SHALL BE VERIFIED PRIOR TO ALTERATIONS.
 2. ALTERED ITEM VENDOR: **THORLABS**
 3. ALTERED ITEM PART NUMBER: **LCP02**
 4. ALTERED ITEM DESCRIPTION: **30 mm to 60 mm Cage Plate Adapter**



VIEW D-D

NOTICE:
INFORMATION CONTAINED IN THIS DOCUMENT OR ANY REPRODUCTION THEREOF, IS PROPRIETARY INFORMATION AND PROPERTY OF HOWARD HUGHES MEDICAL INSTITUTE. IT SHALL NOT BE DISCLOSED, COPIED, DUPLICATED OR USED FOR MANUFACTURE, PRODUCTION OR PROCUREMENT, WITHOUT THE EXPRESS WRITTEN PERMISSION OF HOWARD HUGHES MEDICAL INSTITUTE.

(UNLESS SPECIFIED OTHERWISE)
PRIMARY UNITS: INCHES
[SECONDARY UNITS]: MILLIMETERS
PRIMARY TOLERANCES:
X.X ± 0.020
X.XX ± 0.010
X.XXX ± 0.005
X.XXXX ± 0.0005
ANGULAR ± 0.5 DEG
- DO NOT SCALE DRAWING -
THIRD ANGLE PROJECTION:

HHMI

HOWARD HUGHES MEDICAL INSTITUTE

HHMI

janelia farm

research campus

LCP02_30mmTo60mmCagePlateAdapter_ModForMsrmntBeam.ipt

SIZE	C	PART NUMBER	J001173	REV	-	SHEET	1 OF 1
------	---	-------------	---------	-----	---	-------	--------