StratX™ Lung Report



Patient ID 573

Upload Date

Oct. 24, 2024

Scan ID 075.64

Report Date

Oct. 26, 2024

CT Scan Date Oct. 17, 2024

Scan Comments

Possible signs of Pulmonary Arterial Hypertension

ATTENTION: This CT scan contains motion artifacts that may affect the accuracy of voxel density, fissure completeness and inspiratory volume data present in the report. Please obtain a CT scan without motion artifacts and resubmit.

SUMMARY



KEY

≥ 70% Voxel Density Less Than -910 HU

 \geq 60 – < 70% Voxel Density Less Than -910 HU

> ≥ 50 - < 60% Voxel Density Less Than -910 HU

< 50% Voxel Density Less Than -910 HU

≥ 95% Fissure Completeness

≥ 80 – < 95% Fissure Completeness

< 80% Fissure Completeness

RESULTS

REGOLIO	RIGHT LUNG				LEFT LUNG	
	RUL	RUL+RML	RML	RLL	LUL	LLL
% Fissure Completeness	95	95	95	95	95	95
% Voxel Density Less Than -910 HU	11	15	21	20	12	16
% Voxel Density Less Than -950 HU	3	4	5	6	3	4
Inspiratory Volume (mI)	878	1328	450	630	984	570