

MN242, August 5th, 2024
(After wire replacement)

Short wire : ①

Missing anode pin : ⑤⑤

Sparks : ⑥ ⑦

After 1st build

MN 242

+ wire 1 broken

+ wire 69 spark at cal side

+ Straws 3 still blocked a little

→ need to drill again

+ wire 55 HV side anode pin is pulled out, wire 55 drop

+ econ card problem → test again → OK

+ No pass AC test