

## Pre Plane Assembly Summary

Panel No	Plane No	Cleaned/ Prepped	Blocked straws	Missing wires	Missing wires confirmed	PFN & std. off epoxy	Ear install	Ear leak test	Prep Finished	Panel Pre-Assembly gas leak test
MN277	26	✓	none			✓	grey ✓ clear ✓	✓		✓
Repairs	Added epoxy to Rib Leak replaced 2 clips (40/41) (2/3)									
HV issues	2 piece ch 20 touch ch 21 <del>for</del> Cal side ; 2 piece ch 1 cal side not stable  Fe55 test - ① No block straw HV test - ① Pen & C test gain test - ①									
Notes	air test for blocked straws - ✓ OMB gap test w/ 5 mil kapton 3 air - ✓ Rib leak - Rt side corner by ear backside PFN check - ✓ ear flooding - ✓									