

Pre Plane Assembly Summary

[illegible]

N215 (July 22, 2024)

wire broken: 0

wire missing (straw): 62

problems resolved with KAPTON

~~81, 87~~ 81, 87]

high current wire: 29

passed
QC test.

NOTE 6-13-24 Bob Wagner

1) Treated with IPA

2) Screw 29 disconnected (leave disconnected)

3) AT .5 SCFH, only channel 13 > 1 nA in room
13 is usually ≈ 1.5 nA.

4) AT RH $\approx 47\%$, ALL CURRENTS < 1 nA AT .16 SCFH