

## Pre Plane Assembly Summary

Panel No	Plane No	Cleaned/ Prepped	Blocked straws	Missing wires	Missing wires confirmed	PFN & stnd. off epoxy	Ear install	Ear leak test	Prep Finished	Pre Assy gas leak test
MN/77	26	✓		45, 65, 88	45 broken wire		grey ✓ clear ✓	✓ ✓		✓
Repairs	ear flooded PFN cleared drilled (25)(26)									LT post PFN repair ✓
HV issues	<del>wire 80 spark</del> blocked straws: 25, 26 Fe55 test - ✓ HV test - ✓ Gain test - ✓ passes the QC test but need to repair the blocked straws [07-27-2023 - no blocked straws]									
Notes	Air tested for blocked straws - ✓ DMB gap test w/5 mil Kapton 3, air - ✓ backside PFN check - ✓ #2 has epoxy init Ear Flooding - ✓									

Inflation: 1