

## Pre Plane Assembly Summary

Panel No	Plane No	Cleaned/ Prepped	Blocked straws	Missing wires	Missing wires confirmed	PFN & std. off epoxy	Ear install	Ear leak test	Prep Finished	Panel pre assy. gas leak test
MM 283	19.	✓				✓ ✓	CIR. Gray ✓	CIR. Gray ✓	✓	✓
Repairs	cleared PFNs									LT Post PFN repair
HV issues	<p>Wire short: 1</p> <p>FeSS test - ①</p> <p>HV test - ②</p> <p>gain test - ③</p> <p>pass QC test</p>									
Notes	<p>Replaced for gaps w/ 5mil Kapton + air - ✓</p> <p>DmB test ↑</p> <p>Backside pfn test - #2, 3, 4, 5, 7, 8 need to be cleared</p> <p>Ear flooding - ✓</p> <p>Left end Backside MIR screw head high, needs to be filed down</p>									