

stand off

Panel No	Plane No	Cleaned/ Prepped	Air test	Blocked straws	Missing wires	Missing wires confirmed	PFN + epoxy	DMB Id	ear install	
MNO71	11	✓	✓	0			✓	7 installed - already	PASS	Leak ✓
Repairs	Replaced omega - 46.65 to FRAME TO BP LEAK LEFT AND OR 2-18-22 (RW) 78/79 • ear flooded Removed epoxy from far left (inner) pfn; replaced cap & re epoxied * Smoothed BOP counter bores before installing in 11 plane 12/23/2024 Kapton washers at all pins High current wire: 29 Gain ✓ HV ✓ FOSS ✓									2nd gas leak test
HV issues										2-24-22 PASS LEAK TEST
Notes	* Re Epoxied Stand offs / threads * Damage to Sealing screw counter bores needs smoothing Omega missing L-78/79 Removed from plane 4 (Rib Leak) 2/8/22 Replaced w/ 105									