

Panel No	Plane No	Cleaned/ Prepped	Air test	Blocked straws	Missing wires straw/#58	Missing wires confirmed	PFN + epoxy	DMB Id	ear install	
MN074	4	✓	✓	1 #61		58	✓		pass	Leak test ✓
Repairs	ear flooded. cleaned 42, 43, 48, 49									Pre Assy. complete ✓
HV issues	+ blocked straw at cathode channel 14, 30 + cathode channel 24, high current 24, 21 high current - HV test ⓐ + gain test ⓑ + pass AC test but need to repair blocked straws and wire high current current - FCSS 1st Ⓐ After 1st repaired + blocked straw: 61 + wire high current: 42 + speak at channel 4									
Notes	MISSING straw - # 58 MISSING Omega - 10/11 / Right side * Re epoxied stand off threads • epoxy in DMB conn cleaned out / from hole in jumper conn.									+ pass AC test