

Pre Plane Assembly Summary

Panel No	Plane No	Cleaned/ Prepped	Blocked straws	Missing wires	Missing wires confirmed	PFN & std. off epoxy	Ear install	Ear leak test	Prep Finished	Panel Pre- Assy. gas leak test
MN235	21	✓	Same as →	54, 62		✓	grey ✓ clear ✓	✓		✓
Repairs										
HV issues	<ul style="list-style-type: none"> • Fe55 test - ✓ • HV test - (✓) • Gain test - (✓) 									
Notes	<ul style="list-style-type: none"> • Air tested for blocked straws - OK • DMB gap test w/ 5 mil kapton and air - OK • backside PFN check - OK • Ear flooding - ✓ 									