
(A) RSNA External Clinical Validation Subgroup Performance
AUROC (%)
Sensitivity/Recall (%)
Specificity (%)
Metric Value (%)
0
20
40
60
80
100
93.5%
91.3%
79.7%
81.0%
91.3%
51.7%
71.7%
91.3%
30.8%
Pneumonia vs. Normal
(N = 14,863)
Overall Cohort
(All N = 26,684)
Pneumonia vs. Other Abnormalities
(N = 17,833)
(B) Full External RSNA Confusion Matrix (N = 26,684)
Radiologist Ground-Truth Annotation
Actual No Opacity
Actual Lung Opacity
TN: 10,688
(40.1%)
FP: 9,984
(37.4%)
FN: 520
(1.9%)
TP: 5,492
(20.6%)
Predicted No Opacity
Predicted Lung Opacity
Model Diagnostic Classification