

**CHROMATOGRAPHIC CONDITIONS:****ASSAY**

Column No:	<u>183</u>	Type of Column:	<u>Nucleosil C18, 12.5cm , 15cm</u>
Column Temp (°C):	<u>35</u>		
Detection λ (nm):	<u>255</u>	Injection Vol (μL):	<u>10</u>

Mobile Phase: Composition (% v/v) &amp; Ratios

<u>ACN:BUFFER()2:98</u>	Flow Rate (mL/min):	<u>0.8</u>
	Pump Pressure (bars):	<u>98-120</u>

**DISSOLUTION**

Column No:	<u>183</u>	Type of Column:	<u>Nucleosil C18, 12.5cm , 15cm</u>
Column Temp (°C):	<u>35</u>		
Detection λ (nm):	<u>255</u>	Injection Vol (μL):	<u>10</u>

Mobile Phase: Composition (% v/v) &amp; Ratios

<u>ACN:BUFFFER()2:98</u>	Flow Rate (mL/min):	<u>0.8</u>
	Pump Pressure (bars):	<u>98-120</u>

**REFERENCE SUBSTANCES:**

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.			
2.			
3.			
4.			
5.			