

CHROMATOGRAPHIC CONDITIONS:

ASSAY

Column No: 256 Type of Column: XTERRA C18 4.6X250MM
Column Temp (°C): 30
Detection λ (nm): 254 Injection Vol (μ L): 10

Mobile Phase: Composition (% v/v) & Ratios

H2O: ACN: TEA Flow Rate (mL/min): 1.0
1400:400:2 Pump Pressure (bars): 149-151

DISSOLUTION

Column No: 256 Type of Column: XTERRA WATERS C18 4.6X250MM
Column Temp (°C): 30
Detection λ (nm): 254 Injection Vol (μ L): 10

Mobile Phase: Composition (% v/v) & Ratios

H2O: ACN: TEA Flow Rate (mL/min): 1
1400:400:2 Pump Pressure (bars): 149-151

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.	SULFAMETHOXAZOLE	J1F148	99.02
2.	TRIMETHOPRIM	L0M053	99.75
3.			
4.			
5.			