

CHROMATOGRAPHIC CONDITIONS:**ASSAY**

Column No: 118 Type of Column: Gemini C18, 25 cm
Column Temp (°C): 40
Detection λ (nm): 297 Injection Vol (μL): 50

Mobile Phase: Composition (% v/v) & Ratios

Methanol:Water:Triethylamine:75:25:0.4 Flow Rate (mL/min): 1.0
Pump Pressure (bars): 160-190

DISSOLUTION

Column No: _____ Type of Column: _____
Column Temp (°C): _____
Detection λ (nm): _____ Injection Vol (μL): _____

Mobile Phase: Composition (% v/v) & Ratios

Flow Rate (mL/min): _____
Pump Pressure (bars): _____

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.			
2.			
3.			
4.			
5.			