CHROMATOGRAPHIC CONDITIONS:

ASSAY

Column No: ___223___ Type of Column: ___Xterra RP 18, 25cm___

Column Temp (°C): ___30___

Detection λ (nm): ___220___ Injection Vol (µL): ___20___

Mobile Phase: Composition (% v/v) & Ratios

BUFFER:MEOH (19:1) Flow Rate (mL/min): ___2___

Pump Pressure (bars): ___135-160___

DISSOLUTION

Column No: _____ Type of Column: ___,___

Column Temp (°C): _____

Detection λ (nm): _____ Injection Vol (µL): _____

Mobile Phase: Composition (% v/v) & Ratios

_____ Flow Rate (mL/min): _____

Pump Pressure (bars): _____

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.			
2.			
3.			
4.			
5.			