

CHROMATOGRAPHIC CONDITIONS:**ASSAY**

Column No: _____ Type of Column: _____,
Column Temp (°C): _____
Detection λ (nm): _____ Injection Vol (μ L): _____

Mobile Phase: Composition (% v/v) & Ratios

_____ Flow Rate (mL/min): _____
Pump Pressure (bars): _____

DISSOLUTION

Column No: 196 Type of Column: Xterra RP 18, 25cm
Column Temp (°C): 30
Detection λ (nm): 238 Injection Vol (μ L): 20

Mobile Phase: Composition (% v/v) & Ratios

~~M.P.A: Buffer: ACN: 96:4~~
~~M.P.B: ACN: Buffer: 55:45~~

_____ Flow Rate (mL/min): 1.5
Pump Pressure (bars): 205-220

REFERENCE SUBSTANCES:

| NO | Reference Substances/Related Substances | NQCL Code/Batch | Purity (%) |
|----|---|-----------------|------------|
| 1. | | | |
| 2. | | | |
| 3. | | | |
| 4. | | | |
| 5. | | | |