

CHROMATOGRAPHIC CONDITIONS:**ASSAY**

Column No:	180	Type of Column:	Hyperclone BDS,C18,4.6*150mm
Column Temp (°C):	30		
Detection λ (nm):	214	Injection Vol (μL):	10

Mobile Phase: Composition (% v/v) & Ratios

ACN:H2O	Flow Rate (mL/min):	1
23:77	Pump Pressure (bars):	70

DISSOLUTION

Column No:	180	Type of Column:	Hyperclone BDS,C18,4.6*150MM
Column Temp (°C):	30		
Detection λ (nm):	214	Injection Vol (μL):	10

Mobile Phase: Composition (% v/v) & Ratios

ACN:H2O	Flow Rate (mL/min):	1
23:77	Pump Pressure (bars):	71

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.	NEVIRAPINE	N1-6	99.7
2.			
3.			
4.			
5.			