

CHROMATOGRAPHIC CONDITIONS:**ASSAY**

Column No: 155 Type of Column: Eclipse plus C8, 4.6 * 250mm
Column Temp (°C): 30
Detection λ (nm): 226 Injection Vol (μL): 10

Mobile Phase: Composition (% v/v) & Ratios

BUFFER:ACN Flow Rate (mL/min): 1.0
73:27 Pump Pressure (bars): 150-160

DISSOLUTION

Column No: 0 Type of Column: _____
Column Temp (°C): 0
Detection λ (nm): _____ Injection Vol (μL): _____

Mobile Phase: Composition (% v/v) & Ratios

_____ Flow Rate (mL/min): _____
_____ Pump Pressure (bars): _____

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.	Imiquad	I26-2	99.8
2.			
3.			
4.			
5.			