MICROBIAL COUNT

REFERENCE DOCUMENT: SOP NO. BIOL 007

MICROBIOLOGY LAB NO.	DATE RECEIVED	DATE TEST SET	DATE OF RESULTS
BIOL/001/2016	2016-06-27 09:49:24	05-Jul-2016	10-Jul-2016

SAMPLE PREPARATION

$$\frac{10g}{100ml\ Pepone\ Wat} \times \frac{ml}{ml\ Pepone\ Wat} \times \frac{1ml}{1ml\ Plating}$$ Replicates: 2

0   0

RESULTS

0   0

		10¹ CFU	10² CFU	10³ CFU	Negative Control
		<10		<10	
Nutrient Agar	Plate 1				
	Plate 2	0		0	
	Average (A): CFU (Total Aerobic Microbial Count)	0		0	
		<10		<10	Negative Control
Sabourauds Dextrose Agar	Plate 1				
	Plate 2				
	Average (B): CFU (Total Yeast Microbial Count)				

NB: Acceptance Criteria is interpreted as follows depending on route of administration

— 10¹ cfu: maximum acceptable count = 20; 10² cfu: maximum acceptable count = 200; 10³ cfu: maximum acceptable count = 2000; and so forth.

Yes

CONCLUSION: The Product		**Complies**	With the requirements of the Microbial Enumeration Test.
		Does Not Comply	
Analyst:		**Head, Biological Analysis Unit:**	
Date:		Date:	
Analyst:		Signature:	
Date:			

TEST FOR SPECIFIED MICROORGANISMS

REFERENCE DOCUMENT: SOP NO. BIOL 007

MICROBIOLOGY LAB NO.	DATE RECEIVED	DATE TEST SET	DATE OF RESULTS
BIOL/001/2016	2016-06-27 09:49:24	05-Jul-2016	10-Jul-2016

SAMPLE PREPARATION

$$\frac{10g}{100ml\ Pepone\ Wat} \quad X \quad \frac{ml}{ml\ Pepone\ Wat} \quad X \quad \frac{1ml}{1ml\ Plating} \qquad Replicates:\ 2$$

0   0

RESULTS

Microorganism	Test Media	Observation	Negative Control
		0 0	
		<10 <10	
		0	0
		0	0
		<10	<10

Observation – Indicate whether there is growth/turbidity/colour change in the test media or Not.

Yes

CONCLUSION: The Product		**Complies**	With the requirements of the Test for Specified Microorganisms.
		Does Not Comply	
Analyst:		**Head, Biological Analysis Unit:**	
Date:		Date:	
Analyst:		Signature:	
Date:			