



WMS 05-3
CORAL LABORATORIES LTD

AN ISO 9001:2008 CERTIFIED ORGANIZATION



FACTORY : PLOT NO. 57 / 1 (16), BHENSLORE, DUNETHA, NANI DAMAN - 396 210.
TEL.: 0260-2263 923 • FAX : 0260-2263 489 • EMAIL : daman@corallab.com

WORKING STANDARD
CERTIFICATE OF ANALYSIS

Name of the standard : Dihydroartemisinin IH				Working Standard No: WS- 11/15-16	
Manufactures/ Suppliers : Micro Orgochem / Micro Orgochem					
S. No.:	DHA-005-13	Mfg. Date	11/2013	Exp. Date	11/2017
Date of Preparation	17/01/2015	Valid up to	16/01/2016	Qty of W.S.	1 g
Storage Condition: Store in a cool and dry place, protected from light.					
Reference for Analysis : As per IH					

Sr. No.	TEST	Specification/Limit	Result
1	Appearance	White to off white crystalline powder	Off white crystalline powder
2	Water	NMT 1.0%w/w	0.56 %
3	Assay	Between: 98.0% to 102.0 % on anhydrous basis	98.70 %

Remark:

Above material is recommended to be used as a working standard.

Analysed by: *[Signature]*
Date: 12/01/15

Checked by: *[Signature]*
Date: 17/01/15



Approved by: *[Signature]*
Date: 17/01/15