

CHROMATOGRAPHIC CONDITIONS:**ASSAY**

Column No: 223 Type of Column: Xterra RP 18, 25cm
Column Temp (°C): 30
Detection λ (nm): 220 Injection Vol (μ L): 20

Mobile Phase: Composition (% v/v) & Ratios

BUFFER:ACN (19:1) Flow Rate (mL/min): 2
 Pump Pressure (bars): 135-160

DISSOLUTION

Column No: Type of Column: ,
Column Temp (°C):
Detection λ (nm): Injection Vol (μ L):

Mobile Phase: Composition (% v/v) & Ratios

Flow Rate (mL/min):
 Pump Pressure (bars):

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.			
2.			
3.			
4.			
5.			