

CHROMATOGRAPHIC CONDITIONS:**ASSAY**

Column No:	<u>190</u>	Type of Column:	<u>Xterra RP 18, 25cm</u>
Column Temp (°C):	<u>40</u>		
Detection λ (nm):	<u>254</u>	Injection Vol (μL):	<u>20</u>

Mobile Phase: Composition (% v/v) & Ratios

	Flow Rate (mL/min):	<u>1.5</u>
	Pump Pressure (bars):	<u>185-187</u>

DISSOLUTION

Column No:	<u>190</u>	Type of Column:	<u>Xterra RP 18, 25cm</u>
Column Temp (°C):	<u>40</u>		
Detection λ (nm):	<u>254</u>	Injection Vol (μL):	<u>20</u>

Mobile Phase: Composition (% v/v) & Ratios

<u>same as assay</u>	Flow Rate (mL/min):	<u>1.5</u>
	Pump Pressure (bars):	<u>185-187</u>

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.			
2.			
3.			
4.			
5.			