

## MICROBIAL COUNT

REFERENCE DOCUMENT: SOP NO. BIOL 007

MICROBIOLOGY LAB NO.		DATE RECEIVED		DATE TEST SET		DATE OF RESULTS	
BIOL/001/2017		2017-02-03 10:35:20		08-Feb-2017		11-Feb-2017	
SAMPLE PREPARATION							
<div> <div> <div>10000mg</div> <div>_____</div> <div>100ml BPW</div> </div> <div>X</div> <div> <div>10ml</div> <div>_____</div> <div>100ml BPW</div> </div> <div>X</div> <div> <div>1ml</div> <div>_____</div> <div>1ml Plating</div> </div> <div>Replicates: 2</div> </div>							
				0	0		
RESULTS							
				0	0		
		10 <sup>1</sup> CFU	10 <sup>2</sup> CFU	10 <sup>3</sup> CFU	Negative Control		
Nutrient Agar	Plate 1						
	Plate 2						
	Average (A): CFU (Total Aerobic Microbial Count)	0		0			
		0		0	Negative Control		
Sabourauds Dextrose Agar	Plate 1						
	Plate 2						
	Average (B): CFU (Total Yeast Microbial Count)						
<b>NB: Acceptance Criteria is interpreted as follows depending on route of administration</b> – 10 <sup>1</sup> cfu: maximum acceptable count = 20; 10 <sup>2</sup> cfu: maximum acceptable count = 200; 10 <sup>3</sup> cfu: maximum acceptable count = 2000; and so forth.							
CONCLUSION: The Product			Complies		With the requirements of the Microbial Enumeration Test.		
			Does Not Comply				
Analyst:				Head, Biological Analysis Unit:			
Date:				Date:			
Analyst:				Signature:			
Date:							

# TEST FOR SPECIFIED MICROORGANISMS

REFERENCE DOCUMENT: SOP NO. BIOL 007

MICROBIOLOGY LAB NO.		DATE RECEIVED		DATE TEST SET		DATE OF RESULTS	
BIOL/001/2017		2017-02-03 10:35:20		08-Feb-2017		11-Feb-2017	
SAMPLE PREPARATION							
<div> <div> <div>10000mg</div> <div>100ml BPW</div> </div> <div>X</div> <div> <div>10ml</div> <div>100ml BPW</div> </div> <div>X</div> <div> <div>1ml</div> <div>1ml Plating</div> </div> <div>Replicates: 2</div> </div>							
<div>0</div> <div>0</div>							
RESULTS							
Microorganism	Test Media	Observation	Negative Control				
		0	0				
		0	0				
		0	0				
		0	0				
		< 100	0				
<b>Observation</b> - Indicate whether there is growth/turbidity/colour change in the test media or Not.							
CONCLUSION: The Product		Complies	With the requirements of the Test for Specified Microorganisms.				
		Does Not Comply					
Analyst:			Head, Biological Analysis Unit:				
Date:			Date:				
Analyst:			Signature:				
Date:							