



Kotra Pharma (M) Sdn Bhd  
1, Jalan TTC 12, (90082-V)  
Cheng Industrial Estate,  
75250 Melaka, Malaysia  
Tel : +606 336 2222  
Fax: +606 336 6122

Wholly-owned subsidiary of Kotra Industries Berhad  
(Listed in BSM Main Board)  
Member of the Malaysian Organization of Pharmaceutical Industries  
www.kotrapharma.com

Sales and Marketing Office:  
B212, Level 2, Block B, Kelana Square,  
17, Jalan SS7/26 Kelana Jaya,  
47301 Petaling Jaya,  
Selangor Darul Ehsan, Malaysia.  
Tel: +603 7494 1288 Fax: +603 7494 1299

Document no: FR/QC/016  
SOP Reference: Q/S/009

Rev : 01

Effective Date : 23 May 2013  
Supersedes : Rev:00;15/07/08


### CERTIFICATE OF ANALYSIS Secondary Reference Standard

Substance Name : BENZYL ALCOHOL  
Working Standard No. : WS14022  
Standardization Date : 15 May 2014      Expiry Date : 30 Apr 2015


Description : A clear, colourless and oily liquid  
Loss On Drying : N/A  
Assay : 100.23% of Benzyl Alcohol calculated on as is basis.

Source of Working Standard : Benzyl Alcohol, Batch No: BZA 1405030  
Source of Primary Standard : Benzyl Alcohol (PS13005/ I0J228)  
Intended Use : For test and assay use only.  
Direction for Use : Do not dry before use.  
Storage : Keep in a tightly closed container and protected from light.  
Store container in refrigerator.

Prepared By :

  
Name: Leong How Yee  
QC Chemist  
Date: 20/05/14

Checked By :

  
Name: Simon Piong  
QC Manager  
Date: 22/06/14

