

**CHROMATOGRAPHIC CONDITIONS:****ASSAY**

Column No:	96	Type of Column:	XTerra waters,4.6*250mm,C18
Column Temp (°C):	30		
Detection λ (nm):	270	Injection Vol (μL):	5

Mobile Phase: Composition (% v/v) &amp; Ratios

BUFFER:MEOH	Flow Rate (mL/min):	0.8
60:40	Pump Pressure (bars):	195-200

**DISSOLUTION**

Column No:	96	Type of Column:	XTerra waters,4.6*250mm,C18
Column Temp (°C):	30		
Detection λ (nm):	270	Injection Vol (μL):	5

Mobile Phase: Composition (% v/v) &amp; Ratios

BUFFER:MEOH	Flow Rate (mL/min):	0.8
60:40	Pump Pressure (bars):	195-200

**REFERENCE SUBSTANCES:**

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.	Lamivudine	L3-10	99.39
2.	Nevirapine	DBH027-C16A-160912	99.7
3.	Zidovudine	Z1-3	99.65
4.			
5.			