

CHROMATOGRAPHIC CONDITIONS:

ASSAY

Column No:	116	Type of Column:	Gemini C18, 25 cm
Column Temp (°C):	30		
Detection λ (nm):	254	Injection Vol (μL):	20

Mobile Phase: Composition (% v/v) & Ratios

H2O:ACN:TEA	Flow Rate (mL/min):	1.5
1598:400:2	Pump Pressure (bars):	122-125
PH 5.9		

DISSOLUTION

Column No:	116	Type of Column:	Gemini C18, 25 cm
Column Temp (°C):	30		
Detection λ (nm):	254	Injection Vol (μL):	20

Mobile Phase: Composition (% v/v) & Ratios

H2O:ACN:TEA	Flow Rate (mL/min):	1.5
1598:400:2	Pump Pressure (bars):	122-125
PH 5.9		

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.			
2.			
3.			
4.			
5.			