

## MICROBIAL COUNT

REFERENCE DOCUMENT: SOP NO. BIOL 007

MICROBIOLOGY LAB NO.		DATE RECEIVED		DATE TEST SET		DATE OF RESULTS	
BIOL/001/2015		2015-10-27 07:17:28		04-Nov-2015		09-Nov-2015	
SAMPLE PREPARATION							
<div style="display: flex; justify-content: space-around; align-items: center;"> <div style="text-align: center;">             10ml              _____              100ml 90           </div> <div>X</div> <div style="text-align: center;">             10ml              _____              100ml 90           </div> <div>X</div> <div style="text-align: center;">             1ml              _____              1 mlml Plating           </div> <div style="margin-left: 20px;">Replicates: 2</div> </div>							
				>250		0	
RESULTS							
				>250		0	
		10 <sup>1</sup> CFU	10 <sup>2</sup> CFU	10 <sup>3</sup> CFU	Negative Control		
Nutrient Agar	Plate 1						
	Plate 2						
	Average (A): CFU	>250		0			
	(Total Aerobic Microbial Count)	>250		0			
					Negative Control		
Sabourauds Dextrose Agar	Plate 1						
	Plate 2						
	Average (B): CFU	>250		0			
	(Total Yeast Microbial Count)						
<b>NB: Acceptance Criteria is interpreted as follows depending on route of administration</b> – 10 <sup>1</sup> cfu: maximum acceptable count = 20; 10 <sup>2</sup> cfu: maximum acceptable count = 200; 10 <sup>3</sup> cfu: maximum acceptable count = 2000; and so forth. <div style="text-align: center;"><b>No</b></div>							
CONCLUSION: The Product			Complies			With the requirements of the Microbial Enumeration Test.	
			Does Not Comply				
Analyst:				Head, Biological Analysis Unit:			
Date:				Date:			
Analyst:				Signature:			
Date:							

# TEST FOR SPECIFIED MICROORGANISMS

REFERENCE DOCUMENT: SOP NO. BIOL 007

MICROBIOLOGY LAB NO.		DATE RECEIVED		DATE TEST SET		DATE OF RESULTS	
BIOL/001/2015		2015-10-27 07:17:28		04-Nov-2015		09-Nov-2015	
SAMPLE PREPARATION							
<div> <div>10ml</div> <div>X</div> <div>10ml</div> <div>X</div> <div>1ml</div> <div>Replicates: 2</div> </div> <div> <div>100ml 90</div> <div>100ml 90</div> <div>1 mlml Plating</div> </div>							
<div> <div>&gt;250</div> <div>0</div> </div>							
RESULTS							
Microorganism	Test Media	Observation	Negative Control				
		>250	0				
		>250	0				
		>250	0				
		>250	0				
		>250	0				
		>250	0				
<b>Observation</b> - Indicate whether there is growth/turbidity/colour change in the test media or Not.							
CONCLUSION: The Product	Complies		With the requirements of the Test for Specified Microorganisms.				
	Does Not Comply						
Analyst:				Head, Biological Analysis Unit:			
Date:				Date:			
Analyst:				Signature:			
Date:							