## **CHROMATOGRAPHIC CONDITIONS:**

		<b>ASSAY</b>		
Column No:	96	Type of Column:	xterra c18 250*46 mm	
Column Temp (°C):	30			
Detection $\lambda$ (nm):	254	Injection Vol (μL):	20	
Mobile Phase: Compos	sition (% v/v	v) & Ratios		
h20:acn:tea			Flow Rate (mL/min):	2
1598:400:2 ph 5.9			Pump Pressure (bars):	173-176
		DISSOLUTIO	<u>ON</u>	
Column No:	96	Type of Column:	xterra c18 250*46 mm	
Column Temp (°C):	30			
Detection $\lambda$ (nm):	254	Injection Vol (μL):	20	
Mobile Phase: Compos	sition (% v/v	v) & Ratios		
h20:acn:tea			Flow Rate (mL/min):	2
1598:400:2 ph 5.9			Pump Pressure (bars):	175-177

## **REFERENCE SUBSTANCES:**

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.	Sulfamethoxazole	SMX15082073	99.8
2.	Trimethoprim	VWS/135/15	99.3
3.			
4.			
5.			