

CHROMATOGRAPHIC CONDITIONS:

ASSAY

Column No: 96 Type of Column: Xterra RP 18, 25cm
Column Temp (°C): 40
Detection λ (nm): 254 Injection Vol (μ L): 10

Mobile Phase: Composition (% v/v) & Ratios

Water:ACN:TEA (1598:400:2) pH 5.9 Flow Rate (mL/min): 1.2
 Pump Pressure (bars): 158-160

DISSOLUTION

Column No: 96 Type of Column: Xterra RP 18, 25cm
Column Temp (°C): 40
Detection λ (nm): 254 Injection Vol (μ L): 10

Mobile Phase: Composition (% v/v) & Ratios

Water:ACN:TFA (1598:400:2) pH 5.9 Flow Rate (mL/min): 1.2
 Pump Pressure (bars): 158-160

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.			
2.			
3.			
4.			
5.			