

CHROMATOGRAPHIC CONDITIONS:**ASSAY**

Column No: 99 Type of Column: XTERRA C18 4.5X250MM
Column Temp (°C): 30
Detection λ (nm): 248 Injection Vol (μL): 10

Mobile Phase: Composition (% v/v) & Ratios

ACN:BUFFER Flow Rate (mL/min): 1.0
1:1 Pump Pressure (bars): 125-130

DISSOLUTION

Column No: 0 Type of Column: _____
Column Temp (°C): 0
Detection λ (nm): _____ Injection Vol (μL): _____

Mobile Phase: Composition (% v/v) & Ratios

_____ Flow Rate (mL/min): _____
_____ Pump Pressure (bars): _____

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.	AMBROXOL HYDROCHLORIDE	SAM/AMH/0416001	99.39
2.			
3.			
4.			
5.			