

CHROMATOGRAPHIC CONDITIONS:**ASSAY**

Column No: 206 Type of Column: _____
Column Temp (°C): 30 _____
Detection λ (nm): 265 Injection Vol (μL): 10

Mobile Phase: Composition (% v/v) & Ratios

TRIETHYLAMINE PH 3.2:ACN(98:2) Flow Rate (mL/min): 1

Pump Pressure (bars): 165-220

DISSOLUTION

Column No: 259 Type of Column: C8,NUCLEODUR 100-5
Column Temp (°C): 35 _____
Detection λ (nm): 265,295 Injection Vol (μL): 20

Mobile Phase: Composition (% v/v) & Ratios

PHOSPHORIC ACID PH 3.0:METHANOL(97:3) Flow Rate (mL/min): 2.2

Pump Pressure (bars): 65-120

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.	EFFAVIRENZ	E15-6	97.21
2.	LAMIVUDINE	NDQE201607029	99.39
3.	TENOFOVIR	T11-10	99.54
4.			
5.			