

**CHROMATOGRAPHIC CONDITIONS:****ASSAY**

Column No: 99 Type of Column: Xterra RP 18, 25cm  
Column Temp (°C): 30  
Detection  $\lambda$  (nm): 270 Injection Vol ( $\mu$ L): 20

Mobile Phase: Composition (% v/v) & Ratios

BUFFER:MEOH Flow Rate (mL/min): 1  
50:50 Pump Pressure (bars): 135-140

**DISSOLUTION**

Column No: 99 Type of Column: Xterra RP 18, 25cm  
Column Temp (°C): 30  
Detection  $\lambda$  (nm): 270 Injection Vol ( $\mu$ L): 20

Mobile Phase: Composition (% v/v) & Ratios

BUFFER:MEOH Flow Rate (mL/min): 1  
50:50 Pump Pressure (bars): 135-140

**REFERENCE SUBSTANCES:**

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.			
2.			
3.			
4.			
5.			