

## STERILITY TEST

**REFERENCE DOCUMENT: SOP NO: NQCL/SOP/BIOL – 006**

MICROBIOLOGY LAB NO.	DATE RECEIVED 2016-08-10	DATE TEST SET 09-Sep-2016	DATE OF RESULTS 23-Sep-2016	
	<b>MF</b>			
SAMPLE PREPARATION				
METHODOLOGY: Method Used	<b>Membrane Filtration</b>	<b>6 Vials</b>	(Tick as appropriate)	
	<b>Direct Inoculation</b>			
Quantity Used per filtration/per media:				
Any other Procedures Done:				
RESULTS				
	<b>Sample</b>	<b>Positive Control</b>	<b>Negative Control</b>	<b>Positive Sample Control</b>
<b>Fluid Thioglycolate Medium</b>	<b>No Growth</b>	<i>C. sporogenes</i> (NC00532) <b>Growth</b>	<b>Growth</b>	<i>C. sporogenes</i> (NC00532) <b>Growth</b>
		<i>P. aeruginosa</i> (NC12924) <b>Growth</b>		<i>P. aeruginosa</i> (NC12924) <b>Growth</b>
<b>Soya Bean Digest Medium</b>	<b>No Growth</b>	<i>A. brasiliensis</i> (NC02275) <b>Growth</b>	<b>Growth</b>	<i>A. brasiliensis</i> (NC02275) <b>Growth</b>
		<i>C. albicans</i> (NCPF3179) <b>Growth</b>		<i>C. albicans</i> (NCPF3179) <b>Growth</b>
<b>Key:</b> (Tick: √) - Indicates turbidity, hence microbial growth; (Cross: X) - Indicates clear, hence no microbial growth.				
<b>REMARKS</b>				
*Inoculation of Soya Bean Digest Agar & Sabourauds Dextrose Agar				
		<b>Sample</b>	<b>Negative Control</b>	
<b>Soya Bean Digest Agar</b>				
<b>Sabourauds Dextrose Agar</b>				
<b>Yes</b>				
<b>CONCLUSION:</b> The Product		<b>Complies</b>		<b>With the requirements of the Sterility Test.</b>
		<b>Does Not Comply</b>		
<b>Analyst:</b>		<b>Head, Biological Analysis Unit:</b>		
Date:		Date:		
<b>Analyst:</b>		<b>Signature:</b>		
Date:				

\* Done as a confirmation test where any turbidity observed is suspected to be due to particles from the sample, or due to a reaction between the sample and the media.