

CHROMATOGRAPHIC CONDITIONS:**ASSAY**

Column No: 155 Type of Column: Eclipse plus C8,4.6*250mm
Column Temp (°C): 40
Detection λ (nm): 215 Injection Vol (μ L): 50

Mobile Phase: Composition (% v/v) & Ratios

ACN:MEOH:TETRAHYDROFURAN:BUFFER 1 Flow Rate (mL/min): 1.0
17.5:10:10:62.5 Pump Pressure (bars): 157-159

DISSOLUTION

Column No: 0 Type of Column: _____
Column Temp (°C): 0
Detection λ (nm): _____ Injection Vol (μ L): _____

Mobile Phase: Composition (% v/v) & Ratios

_____ Flow Rate (mL/min): _____
_____ Pump Pressure (bars): _____

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.	RITONAVIR	R9-2	99.4
2.	LOPINAVIR	L20-2	99.8
3.			
4.			
5.			