

CHROMATOGRAPHIC CONDITIONS:**ASSAY**

Column No: 243 Type of Column: Symmetry C18, 25cm
Column Temp (°C): 30
Detection λ (nm): 200 Injection Vol (μ L): 200

Mobile Phase: Composition (% v/v) & Ratios

WATER:ACETONITRILE , 520:480 Flow Rate (mL/min): 1.5
 Pump Pressure (bars): 170-175

DISSOLUTION

Column No: 238 Type of Column: Xterra RP 18, 25cm
Column Temp (°C): 30
Detection λ (nm): 200 Injection Vol (μ L): 200

Mobile Phase: Composition (% v/v) & Ratios

WATER:ACETONITRILE , 520:480 Flow Rate (mL/min): 1.5
 Pump Pressure (bars): 170-175

REFERENCE SUBSTANCES:

NO	Reference Substances/Related Substances	NQCL Code/Batch	Purity (%)
1.			
2.			
3.			
4.			
5.			