MICROBIAL COUNT

REFERENCE DOCUMENT: SOP NO. BIOL 007

MICROBIOLOGY LAB NO.	DATE RECEIVED	DATE TEST SET	DATE OF RESULTS
BIOL/001/2016	2016-05-10 12:01:59	28-Jun-2016	04-Jul-2016

SAMPLE PREPARATION

$$\frac{10mg}{100ml\ Buffered\ P} \quad X \quad \frac{10ml}{100ml\ Buffered\ P} \quad X \quad \frac{1ml}{1ml\ Plating} \qquad Replicates: 2$$

0   0

RESULTS

0   0

		10^1 CFU	10^2 CFU	10^3 CFU	Negative Control
		0		**0**	
Nutrient Agar	Plate 1				
	Plate 2	0		0	
	Average (A): CFU (Total Aerobic Microbial Count)	0		0	
		0		0	**Negative Control**
Sabourauds Dextrose Agar	Plate 1	<100		0	
	Plate 2				
	Average (B): CFU (Total Yeast Microbial Count)				

NB: Acceptance Criteria is interpreted as follows depending on route of administration

— 10^1 cfu: maximum acceptable count = 20; 10^2 cfu: maximum acceptable count = 200; 10^3 cfu: maximum acceptable count = 2000; and so forth.

Yes

CONCLUSION: The Product		Complies	With the requirements of the Microbial Enumeration Test.
		Does Not Comply	
Analyst:		**Head, Biological Analysis Unit:**	
Date:		Date:	
Analyst:		Signature:	
Date:			

TEST FOR SPECIFIED MICROORGANISMS

REFERENCE DOCUMENT: SOP NO. BIOL 007

MICROBIOLOGY LAB NO.	DATE RECEIVED	DATE TEST SET	DATE OF RESULTS
BIOL/001/2016	2016-05-10 12:01:59	28-Jun-2016	04-Jul-2016

SAMPLE PREPARATION

$$\frac{10mg}{100ml\ Buffered\ P} \times \frac{10ml}{100ml\ Buffered\ P} \times \frac{1ml}{1ml\ Plating}$$

Replicates: 2

0   0

RESULTS

Microorganism	Test Media	Observation		Negative Control
		0	0	
		0		
		0	0	
		0	0	
		0	0	
		<100	**0**	

Observation – Indicate whether there is growth/turbidity/colour change in the test media or Not. **Yes**

CONCLUSION: The Product		Complies	With the requirements of the Test for Specified Microorganisms.
		Does Not Comply	

Analyst:		Head, Biological Analysis Unit:	
Date:		Date:	
Analyst:		Signature:	
Date:			