MICROBIAL COUNT

REFERENCE DOCUMENT: SOP NO. BIOL 007

MICROBIOLOGY LAB NO.	DATE RECEIVED	DATE TEST SET	DATE OF RESULTS
BIOL/001/2017	2017-05-22 10:53:25	17-May-2017	22-May-2017

SAMPLE PREPARATION

$$\frac{10ml}{100ml\ BPW} \quad X \quad \frac{10ml}{100ml\ BPW} \quad X \quad \frac{1ml}{1ml\ Plating}$$

Replicates: 2

0   0

RESULTS

0   0

		10^1 CFU 0	10^2 CFU	10^3 CFU 0	Negative Control
Nutrient Agar	Plate 1				
	Plate 2	0		0	
	Average (A): CFU (Total Aerobic Microbial Count)	0		0	
		0		0	**Negative Control**
Sabourauds Dextrose Agar	Plate 1	<10		0	
	Plate 2				
	Average (B): CFU (Total Yeast Microbial Count)				

NB: Acceptance Criteria is interpreted as follows depending on route of administration

— 10^1 cfu: maximum acceptable count = 20; 10^2 cfu: maximum acceptable count = 200; 10^3 cfu: maximum acceptable count = 2000; and so forth.

Yes

CONCLUSION: The Product		Complies	With the requirements of the Microbial Enumeration Test.
		Does Not Comply	
Analyst:		**Head, Biological Analysis Unit:**	
Date:		Date:	
Analyst:		Signature:	
Date:			

TEST FOR SPECIFIED MICROORGANISMS

REFERENCE DOCUMENT: SOP NO. BIOL 007

MICROBIOLOGY LAB NO.	DATE RECEIVED	DATE TEST SET	DATE OF RESULTS
BIOL/001/2017	2017-05-22 10:53:25	17-May-2017	22-May-2017

SAMPLE PREPARATION

$$\frac{10ml}{100ml\ BPW} \quad X \quad \frac{10ml}{100ml\ BPW} \quad X \quad \frac{1ml}{1ml\ Plating}$$

Replicates: 2

0   0

RESULTS			
		0	0
Microorganism	Test Media	**Observation**	**Negative Control**
		0	**0**
		0	0
		0	0
		0	0
		<10	**0**

Observation – Indicate whether there is growth/turbidity/colour change in the test media or Not.

Yes

CONCLUSION: The Product		Complies	With the requirements of the Test for Specified Microorganisms.
		Does Not Comply	
Analyst:		**Head, Biological Analysis Unit:**	
Date:		Date:	
Analyst:		Signature:	
Date:			