

CHROMATOGRAPHIC CONDITIONS:**ASSAY**

Column No: _____ Type of Column: _____,
Column Temp (°C): _____
Detection λ (nm): _____ Injection Vol (μ L): _____

Mobile Phase: Composition (% v/v) & Ratios

_____ Flow Rate (mL/min): _____
Pump Pressure (bars): _____

DISSOLUTION

Column No: 100 Type of Column: Xterra RP 18, 25cm
Column Temp (°C): 45
Detection λ (nm): 240 Injection Vol (μ L): 10

Mobile Phase: Composition (% v/v) & Ratios

ph 7:ACN _____ Flow Rate (mL/min): 1

_____ Pump Pressure (bars): 107-110

REFERENCE SUBSTANCES:

| NO | Reference Substances/Related Substances | NQCL Code/Batch | Purity (%) |
|----|---|-----------------|------------|
| 1. | | | |
| 2. | | | |
| 3. | | | |
| 4. | | | |
| 5. | | | |