

TESTING AND MEASURING EQUIPMENT/ALLOWED SUBCONTRACTING IEC 81001-5-1:2021

Health software and health IT systems safety, effectiveness and security – Part 5-1: Security – Activities in the product life cycle

Status:

- ☒ **Interim**
☐ **Provisional**
☐ **Approved**

Abbreviations to be used at the column “Equipment Classification”:

- “R” Required
“S” May be subcontracted, see OD 2012
“SP” Specialized Facility, see IECEE 02-2
“W” Test may be witnessed at a CTF, another CBTL, or an outside subcontractor
“3PPS” Three Phase Power Supply required

Clause	Measurement/testing	Testing / measuring equipment / material needed	Equipment Classification
All	None	As this standard specifies requirements for secure development process for a product manufacturer, there is no actual measurement to take. Consequently, no measurement equipment is needed at all.	NA