

EMN-100 Event Marker

Overview

The EMN-100 is a hand-held, push button switch. When the button is pressed, the EMN-100 delivers a TTL pulse which is displayed on the selected NI myDAQ channel.

Photo



Specifications:

<i>Specifications</i>	
Voltage output:	0 - 5V

Operating Instructions:

1. Plug the IX-MYDAQ into the NI mydaq. Plug the mini din 7 end of the EMN-100 into CH 1 or CH 2 of the IX-MYDAQ.
2. Plug in the USB cable and set the input range of the channel being used to +/-5V.
3. Record and press the button when a desired event occurs.



