

# 600P

SERIES

# Oxygen



## USER'S

## MANUAL

**Note: For Serial Numbers After  
U06081 Please See Addendum  
Starting After Page 95 of This  
Manual**



1312 West Grove Avenue  
Orange, CA 92865-4134  
Phone: 714-974-5560 Fax: 714-921-2531  
[www.gasanalyzers.com](http://www.gasanalyzers.com)



Safety Alert  
Caution or Warning



Temperature Hazard  
Caution or Warning



Electrical Shock Hazard  
Caution or Warning

## Safety Information in this Manual

Note, caution and warning symbols appear on the instrument and throughout this manual to draw your attention to important operational and safety information.

A “**NOTE**” marks a short message to alert you to an important detail.

A “**CAUTION**” safety alert appears with information that is important for protecting your equipment and performance.

A “**WARNING**” safety alert appears with information that is important for protecting you, others and equipment from damage. Pay very close attention to all warnings that apply to your application.



The symbol (an exclamation point in a triangle) precedes a general CAUTION or WARNING statement.



The symbol (wavy vertical lines with an under score in a triangle) precedes an elevated temperature hazard CAUTION or WARNING statement.



The symbol (a lightning bolt in a triangle) precedes an electric shock hazard CAUTION or WARNING statement.

Some or all of the above symbols may appear in this manual or on the equipment. This manual should be consulted whenever one of these symbols is encountered on the equipment.

**ALWAYS REMOVE POWER BEFORE CONNECTING OR DISCONNECTING SIGNAL CABLES OR WHEN SERVICING THE EQUIPMENT.**

**The 600 series Paramagnetic Oxygen instruments meet or exceed the following directives and standards.**

Application of Council Directive(s):

*Electrical Safety:*

Low Voltage Directive 73/23/EEC

*Electromagnetic Compatibility:*

EMC Directive 89/336/EEC

Standard(s) to which Conformity is Declared:

*Electrical Safety:*

*Standard for Electrical Equipment for Measurement, Control, and Laboratory Use [EN 61010-1:2001 (2nd Edition)]*

*Electromagnetic Compatibility:*

*EN 61326:1997 Electrical equipment for measurement, control and laboratory use - EMC requirements (Amendment A1: 1998 to EN 61326:1997; Amendment A2:2001 to EN 61326:1997)*

## **600P Paramagnetic Oxygen Quick Start Guide**

- 1) Plug in the analyzer and turn the power on.
- 2) Connect the appropriate gas lines and vents to the analyzer.
- 3) Allow the analyzer to stabilize for at least one (1) hour.
- 4) During the analyzer's stabilization period, setup the analyzer to the desired configuration.
  - a) Set the analyzer's output as desired.
    - i) From the Main Menu press F5 (Setup) then F7 (System Settings.)
    - ii) Press F3 (output assignments) to setup the output channels.
      - (1) Set to real time.
      - (2) AUX 1
      - (3) AUX 2
      - (4) AUX 3
      - (5) Press back to return to the system settings menu.
    - iii) Press F4 (output range) to setup the output ranges
      - (1) Set the Min and Max to zero (0). This will cause the outputs to default to the current ranges.
      - (2) Set the outputs for milliamps (mA) or voltage (V) as desired.
        - (a) The mA setting will give 4-20 mA or 2-10V (if 500 ohm resistors are installed).
        - (b) The V setting will give 0-20mA or 0-10V (if 500 ohm resistors are installed).

**Disconnect power before proceeding**

- 5) Connect all appropriate analog outputs.
  - a) Pin 1 is the output common (ground).
  - b) Pins 2, 3, 4, and 5 are output channels 1, 2, 3, and 4 as setup in step 4.a.ii.
  - c) If the output is set for voltages but there is no voltage output, you will need to install a 500 ohm resistor between pin 1 and the channel you are trying to measure. You will have to measure the voltage drop across this resistor.
- 6) Reconnect the power and turn on the analyzer. Press F1 from the analyzer's main Menu to get to the measurement screen.
- 7) Supply sample gas to the analyzer.
- 8) The measurement screen should indicate the concentration of the sample gas. If the concentration is incorrect, the analyzer will need to be calibrated using zero and span calibration gases. Please reference

the appropriate section of the manual for a description on how to zero and span the analyzer.

**TABLE OF CONTENTS**

Low Voltage Directive 73/23/EEC .....	4
EMC Directive 89/336/EEC .....	4
<b>1. Introduction.....</b>	<b>16</b>
1.1. Overview.....	16
1.2. Unpacking Instructions .....	16
1.3. Reporting Damage .....	16
1.4. Contact Information .....	16
1.5. Warranty Certificate.....	17
1.6. Proper Operation.....	18
1.7. Possible Explosion Hazard.....	18
1.8. Electrical Shock Hazard .....	18
1.9. Plug Removal .....	18
<b>2. Features.....</b>	<b>19</b>
2.1. Description .....	19
2.2. Features-General .....	19
2.3. Principle of Operation.....	19
2.4. Model 600P - Specifications.....	20
<b>3. Installation.....</b>	<b>22</b>
3.1. General.....	22
3.2. Site and Mounting.....	22
3.3. Electrical .....	22
3.4. Analog Output Connections (Appendix) .....	23
3.5. Gases .....	23
3.5. 1. Sampling Systems .....	23
3.6. Gas Handling Equipment .....	24
3.7. Gas Connections .....	24
3.8. Sampling Requirements .....	24
3.8.1. Filtration .....	24
3.8.2. Condensation.....	24

3.8.3.	Presence of Corrosive Gases.....	24
3.8.4.	Gas Temperature.....	24
3.8.5.	Sample Gas Bypass Outlet (Vent).....	24
<b>4.</b>	<b>Basic Operation .....</b>	<b>25</b>
4.1.	Menu Trees .....	27
4.2.	Keyboard .....	30
4.2.1.	Operation with the Cursor Keys and the Enter Key .....	30
4.2.2.	Operation with the Function Keys.....	30
4.2.3.	Read/Change Parameters .....	30
<b>5.</b>	<b>Operating Structure.....</b>	<b>31</b>
5.1.	The Main Menu.....	32
<b>6.</b>	<b>Menu Structure .....</b>	<b>33</b>
6.1.	User Functions (Level 1) .....	33
6.2.	Advanced User Functions (Level 2) .....	33
6.3.	Maintenance Functions (Level 3) .....	33
6.4.	System User Functions (Level 4) .....	33
<b>7.</b>	<b>Main Menu Function Descriptions .....</b>	<b>34</b>
7.1.	F1 Measurements.....	34
7.1.1.	F1 Measurement.....	34
7.1.2.	Range Select .....	35
7.2.	F2 Purge Analyzer.....	36
7.3.	F3 Diagnostics.....	36
7.4.	F4 Calibrations .....	37
7.4.1.	F1 Automatic Calibration .....	37
7.4.2.	F2 Manual Calibration.....	38
7.4.3.	Display Deviations – from Calibration menu F3.....	39
7.4.4.	Absolute Zero Gas Deviation.....	39
7.4.5.	Relative Zero Gas Deviation.....	39
7.4.6.	Absolute Span Gas Deviation.....	40
7.4.7.	Relative Span Gas Deviation .....	40
7.4.8.	F4 Check Calibration .....	40
7.4.9.	F5 Reset Calibration Values .....	40
7.4.10.	F6 Range Select .....	40

7.5. F5 Setup .....	41
7.5.1. F1 Span Gas Concentration .....	41
7.5.2. Calibration Settings.....	42
7.5.3. F1 Times.....	42
7.5.4. F2 Measuring Deviations .....	42
7.5.5. F3 Deviations.....	42
7.5.6. F3 Range Limits.....	42
7.5.7. F1 Range 1-4 (Change Upper Range Limits).....	43
7.5.8. F2 Change Auto Range Limits.....	43
7.5.9. F4 Alarms .....	44
7.5.10. F5 Password.....	44
7.5.11. F1 Enter Password .....	45
7.5.12. F2 Change Password .....	45
7.5.13. F3 Reset Passwords.....	45
7.5.14. F6 Linearization.....	46
7.6. F7 System Settings .....	47
7.6.1. F1 Real Time Clock .....	47
7.6.2. System Setup F2 Displays TCP/IP Address .....	49
7.6.3. Systems Setup F3 Displays Output Signal Assignments .....	49
7.6.4. System Setup F4 Displays Output Ranges .....	49
7.6.5. F5 Turns Status Line On or Off.....	50
7.6.6. F8 Measure Settings.....	50
7.6.7. F1 Set Lowpass filter .....	50
7.6.8. F2 Purge Time .....	51
7.6.9. F3 Set Temperature Compensation .....	51
7.6.10. F4 Pressure Compensation.....	51
7.6.11. F3 Low Pass Filter Time Constant.....	52
7.6.12. F10 Displays the Current Analyzer and Software Versions .....	52
7.7. F7 Remote / Manual Control .....	52
7.8. F8 Standby .....	53
<b>8. Analyzer Components.....</b>	<b>54</b>
8.1. Rear Panel.....	54
8.2. Rear Panel Connectors .....	55
8.3. Digital Outputs .....	56

8.3.1.	RS-232 (Standard 9 Pin DIN Connector).....	56
8.3.2.	TCP/IP (8 Pin RJ-47 Connector) .....	56
8.4.	Internal Component Locations.....	57
8.5.	Main Electronics Board (Potentiometers).....	58
8.6.	Mechanical Zero Adjustment.....	59
<b>9.</b>	<b>Operation.....</b>	<b>60</b>
9.1.	External Wiring .....	60
9.2.	External Piping .....	60
9.3.	Operation & Calibration .....	60
9.3.1.	Power On:.....	60
9.3.2.	Zero/San Adjustment:.....	60
9.3.3.	Start-Up .....	60
<b>10.</b>	<b>MAINTENANCE .....</b>	<b>61</b>
10.1.	Zero and Span Calibration. ....	61
10.2.	Routine                         Maintenance.....	61
10.3.	Cross Sensitivity of Gas.....	61
<b>11.</b>	<b>ADJUSTMENTS CHECKS AND REPAIRS.....</b>	<b>62</b>
11.2.	Coarse Zero Adjustment.....	62
11.3.	Span Gain Adjustment.....	62
	Table 4-1 Cross Sensitivity of Gases .....	63

11.8.	Check and Repair Tubing Trouble.....	64
11.8.1.	Problem: .....	64
11.8.2.	Symptom.....	64
11.8.3.	Check and/or replace:.....	64
11.9.	Check and Repair Amplifier Circuit.....	65
<b>12.</b>	<b>Communication Master Computer / Analyzer (AK Protocol).....</b>	<b>66</b>
12.1.	Serial Interface and AK-Commands .....	66
12.2.	Interface Specifications .....	66
12.3.	Protocol Description .....	67
12.3.1.	Instruction command .....	67
12.3.2.	Acknowledgement command.....	67
12.3.3.	Data Description .....	68
12.3.4.	Error Handling .....	68
12.4.	Scan Commands .....	69
12.4.1.	<b>AKON:</b> Measured concentration value .....	69
12.4.2.	<b>AEMB:</b> Get measuring range .....	69
12.4.3.	<b>AMBE:</b> Measuring range limit.....	69
12.4.4.	<b>AKAK:</b> Calibration gas concentrations.....	69
12.4.5.	<b>AMBU:</b> Upper and lower range switchover values for auto range.....	69
12.4.6.	<b>ASTZ:</b> Normal device status.....	70
12.4.7.	<b>ASTF:</b> Error status.....	70
12.4.8.	<b>AKEN:</b> Device identification .....	70
12.4.9.	<b>ARMU:</b> Raw value .....	70
12.4.10.	<b>ATEM:</b> Temperatures .....	71
12.4.11.	<b>ADRU:</b> Pressures/ Valve voltage .....	71
12.4.12.	<b>ADUF:</b> Flows .....	71
12.4.13.	<b>AGRД:</b> Polynom coefficients .....	71
12.4.14.	<b>AANG:</b> Deviation from zero point after autocalibration .....	71
12.4.15.	<b>AAEG:</b> Deviation from end point after autocalibration.....	71

12.4.16.	<b>AFDA:</b> Purge and Autocalibration times.....	71
12.4.17.	<b>APAR:</b> Request Autocalibration tolerance values .....	72
12.4.18.	<b>AKAL:</b> Deviations from calibration .....	72
12.4.19.	<b>ASYZ:</b> Respond System Time.....	72
12.4.20.	<b>AT90:</b> Respond Lowpass filter time.....	72
12.4.21.	<b>ADAL:</b> Diagnostic alarm limits.....	72
12.4.22.	<b>ATCP:</b> Query TCP/IP settings .....	72
12.4.23.	<b>AVER:</b> Query Software version .....	73
12.4.24.	<b>AH2O:</b> Query H <sub>2</sub> O correction parameter.....	73
12.4.25.	<b>ACO2:</b> Query CO <sub>2</sub> correction parameter .....	73
12.4.26.	<b>AUPD:</b> Query UDP data streaming parameter .....	73
12.5.	Control commands .....	74
12.5.1.	<b>SRES:</b> Reset .....	74
12.5.2.	<b>SPAU:</b> Pause .....	74
12.5.3.	<b>STBY:</b> Standby.....	74
12.5.4.	<b>SNGA:</b> Open valve for zero gas calibration.....	74
12.5.5.	<b>SEGA:</b> Open valve for end gas calibration.....	74
12.5.6.	<b>SSPL:</b> Purge Analyzer with zero gas .....	74
12.5.7.	<b>SATK:</b> Start automatic calibration .....	74
12.5.8.	<b>SEMB:</b> Set measuring range .....	74
12.5.9.	<b>SARE:</b> Auto range on .....	75
12.5.10.	<b>SARA:</b> Auto range off .....	75
12.5.11.	<b>SREM:</b> Remote mode for AK-commands .....	75
12.5.12.	<b>SMAN:</b> Manual control to control device manually .....	75
12.5.13.	<b>SMGA:</b> Start measuring .....	75
12.5.14.	<b>SNKA:</b> Saves measured value as new offset.....	75
12.5.15.	<b>SEKA:</b> Saves measured value as new span value .....	75
12.5.16.	<b>SUDP:</b> Start /Stop UDP data streaming .....	75
12.6.	Settings .....	76
12.6.1.	<b>EKAK:</b> The four span gas concentration values are set .....	76
12.6.2.	<b>EMBE:</b> The four measuring range end values are set .....	76
12.6.3.	<b>EMBU:</b> The upper and the lower range switchover for auto range are set..	76
12.6.4.	<b>EKEN:</b> Set new device identification and information .....	76
12.6.5.	<b>EGRD:</b> Set polynom coefficients .....	76

12.6.6.	<b>EFDA:</b> Set autocalibration and purge times .....	76
12.6.7.	<b>EPAR:</b> Set autocalibration tolerance values.....	76
12.6.8.	<b>ESYZ:</b> Set System Time.....	77
12.6.9.	<b>ET90:</b> Set Lowpass Filter Time .....	77
12.6.10.	<b>EDAL:</b> Diagnostic alarm limits.....	77
12.6.11.	<b>ETCP:</b> Set TCP/IP Parameters.....	77
12.6.12.	<b>EH2O</b> Set H <sub>2</sub> O correction parameters .....	77
12.6.13.	<b>ECO2</b> Set CO <sub>2</sub> correction parameters .....	77
12.6.14.	<b>EUDP</b> Set UDP Data streaming parameters .....	78
12.6.15.	Format of the streaming Data via UDP:.....	78
12.7.	Abbreviations used.....	78
<b>13.</b>	<b>Appendix 1 Electrical Block Diagram .....</b>	<b>79</b>
<b>14.</b>	<b>Appendix 2- Starting With Serial Number U06081.....</b>	<b>80</b>
	<b>Typical Relay .....</b>	<b>92</b>
	<b>Press F1 to observe the remaining seven outputs.....</b>	<b>93</b>



## 1. Introduction

### 1.1. Overview

Congratulations and thank you! You have just purchased one of the most reliable gas analyzers in the world. Before using the analyzer, please familiarize yourself with its operation by reading this manual. If you have any questions, please do not hesitate to call California Analytical Instruments for assistance. We want you to be a member of our thousands of satisfied customers.

### 1.2. Unpacking Instructions

Open the shipping container and carefully remove the analyzer from the packing materials. Inspect the instrument for any sign of damage. Remove the Top Cover retaining screws. Visually check for loose parts or connectors that are not properly seated. Verify all circuit boards and circuit board connections are secure. If all internal components look normal, re-install the cover.

### 1.3. Reporting Damage

Should there be any apparent damage either to the inside or outside of the instrument due to shipping or handling, immediately notify the shipper. The shipping container or packing materials should be retained for inspection by the shipper.

### 1.4. Contact Information

California Analytical Instruments, Inc.

1312 West Grove Avenue

Orange, CA 92865

714 974-5560

Fax 714 921-2531

Website: [www.gasanalyzers.com](http://www.gasanalyzers.com)

**1.5. *Warranty Certificate***

Subject to the exceptions and upon the conditions stated below, California Analytical Instruments (CAI) warrants that the products sold under this sales order shall be free from defects in workmanship and materials for one year after delivery of the product to the original Buyer by CAI and if any such product should prove to be defective within such one year period, CAI agrees, at its option, either (i) to correct by repair or, at CAI's election, by replacement with equivalent product any such defective product, provided that investigation and factory inspection discloses that such defect developed under normal and proper uses, or (ii) to refund the purchase price. The exceptions and conditions mentioned above are as follows:

- a. components or accessories manufactured by CAI that by their nature are not intended to or will not function for one year are warranted only to give reasonable service for a reasonable time. What constitutes reasonable time and reasonable services shall be determined solely by CAI. A complete list of such components and accessories is maintained at the factory;
- b. CAI makes no warranty with respect to components or accessories not manufactured by it; in the event of defect in any such component or accessory CAI will give reasonable assistance to Buyer in obtaining from the respective manufacturer whatever adjustment is authorized by the manufacturer's warranty;
- c. any product claimed to be defective must be returned to the factory transportation charges prepaid and CAI will return the repaired or replaced product freight collect;
- d. if the product claimed to be defective requires on-site repair, such warranty labor will be provided at no charge; however, transportation and living expenses will be charged to Buyer;
- e. if the product is a consumable or the like, it is warranted only to conform to the quantity and content and for the period (but not in excess of one year) stated on the label at the time of delivery or 90 days;
- f. CAI may from time to time provide a special printed warranty with respect to a certain product, and where applicable, such warranty shall be deemed incorporated herein by reference;
- g. CAI shall be released from all obligations under all warranties, either expressed or implied, if any product covered hereby is repaired or modified by persons other than its own authorized service personnel unless such repair by others is made with the written consent of CAI.

IT IS EXPRESSLY AGREED THAT THE ABOVE WARRANTY SHALL BE IN LIEU OF ALL WARRANTIES OF FITNESS AND OF THE WARRANTY OF MERCHANTABILITY AND THAT CAI SHALL HAVE NO LIABILITY FOR SPECIAL OR CONSEQUENTIAL DAMAGES OF ANY KIND OR FROM ANY CAUSE WHATSOEVER ARISING OUT OF THE MANUFACTURE USE, SALE, HANDLING, REPAIR, MAINTENANCE OR REPLACEMENT OF ANY OF THE PRODUCTS SOLD UNDER THIS SALES ORDER. SOME STATES DO NOT ALLOW THE EXCLUSION OR LIMITATION OF INCIDENTAL OR CONSEQUENTIAL DAMAGES, SO THAT THE ABOVE LIMITATIONS OR EXCLUSIONS MAY NOT APPLY. THIS WARRANTY GIVES YOU SPECIFIC LEGAL RIGHTS, AND YOU MAY HAVE OTHER RIGHTS, WHICH VARY FROM STATE TO STATE.

Representations and warranties made by any person, including dealers and representatives of CAI which are inconsistent or in conflict with the terms of this warranty, shall not be binding upon CAI unless produced in writing and approved by an expressly authorized officer of CAI.

**1.6. Proper Operation**

Personnel should be trained in the proper operation of this equipment before attempting to operate the equipment.

**1.7. Possible Explosion Hazard**

Do not apply power to the analyzer or attempt to energize the analyzer until determining the analyzer environment to be non-hazardous.

Use this analyzer in a **NON-HAZARDOUS** environment.

This analyzer has not been designed for use with a hazardous sample.

Tampering or use of substitute components may cause a safety hazard. Use only factory authorized replacement parts.

**1.8. Electrical Shock Hazard**

Disconnect power before removing cover. Servicing requires access to live electrical components that can cause death or serious injury. Refer servicing to qualified service personnel. For safety and proper performance, connect this instrument to a properly grounded three-wire receptacle.

**Caution****1.9. Plug Removal**

***Do not operate this analyzer UNTIL REMOVING the red plastic ¼ inch plugs from the sample inlet and exhaust fittings on the rear panel.***



**Use of this equipment in a manner not approved by California Analytical Instruments is not recommended and may cause harm to the equipment or operating personnel.**

## 2. Features

### 2.1. Description

The Paramagnetic Oxygen Analyzer (PMA) is a thermostated instrument designed primarily for, but not necessarily limited to, stationary use. It is a '19" rack mount' analyzer that is also suitable for bench top use. The operation of the analyzer is based upon the principle of the magneto-dynamic oxygen cell, which is the most accurate and reliable cell for determining the oxygen content of a gas mixture from 0-100 volume percent oxygen.


### 2.2. Features-General

The Model 600 series of analyzers have a 3 by 5 inch liquid crystal display and a 20 key data/operation input keyboard. The 16-bit microprocessor control board consists of the MSR-Card with 16 digital inputs, 16 digital outputs, 16 analog inputs and 4 analog outputs. The analyzer can be manually operated from the keypad or remotely via TCP/IP or RS-232C communications. After turning on the analyzer, it needs at least 30 seconds for initialization. During this time, the screen is illuminated.

**IMPORTANT TIP:** When the analyzer is powered up, it defaults to access level 1 (User). To operate ALL parameters, check the access level. See Section 5.


### 2.3. Principle of Operation

The paramagnetic susceptibility of oxygen is significantly greater than that of other common gases, and consequently, the molecules of oxygen are attracted much more strongly by a magnetic field than the molecules of the other gases. Most of the other gases are slightly diamagnetic, which means that their molecules are then repelled by a magnetic field.




**Figure 2-4 Magnetic Susceptibility of gases**

The principle of the magneto dynamic cell is based upon Faraday's method of determining the magnetic susceptibility of a gas. The cell consists of two nitrogen-filled quartz spheres arranged in the form of a dumbbell. A single turn of platinum wire is placed around the dumbbell that is suspended in a symmetrical non-uniform magnetic field.

**Figure 2-2 The Measuring cell in theory**

When the surrounding gas contains oxygen, the dumbbell spheres are pushed out of the magnetic field by the change in the field that is caused by the relatively strong paramagnetic oxygen. The torque acting on the dumbbell is proportional to the paramagnetic properties of the surrounding gas and, therefore, it can be used as a measure of the oxygen concentration. The distortion of the dumbbell is sensed by a light beam and projected on a mirror attached to the dumbbell whereof it is reflected to a pair of photocells. When both photocells are illuminated equally, the output will be zero. The output from the photocells is connected to an amplifier, which in turn is fed to the feedback coil of the measuring cell. If the oxygen content of the gas sample changes, the corresponding current output of the amplifier, which is proportional to the oxygen content, produces a magnetic field in the feedback coil opposing the forces and thereby causing the dumbbell to rotate.

**Figure 2-3 Principle of operation**

Since the feedback current from the amplifier is proportional to the oxygen content of the gas sample, the output signals that are produced by the amplifier will be accurate and linear. The paramagnetic susceptibility of oxygen varies inversely as the square of the absolute temperature. To provide compensation for changes in analyzer temperature, a temperature sensitive element in contact with the measuring cell assembly is included in the feedback current circuit

#### **2.5. Product Specifications Model 600P (Paramagnetic Detector)**

SAMPLE CONTACT MATERIAL: Platinum, Glass, stainless steel, Viton, Teflon\* and Tygon\*\*

RANGES: 0 - 1 % up 25 % or 0 – 5 % up to 100 % (max. of four ranges)

OUTPUTS: 0 to 10 VDC or 4 to 20 mA (0 – 20 mA), RS232 & TCP/IP (AL Protocol)

AMBIENT TEMPERATURE: 5 to 45° C

SAMPLE TEMPERATURE: 0 to 50° C

SAMPLE CONDITION: Particles < 1μ, non-corrosive dry gas

FITTINGS: ¼" tube

SAMPLE FLOW RATE: 0.5 –2.0 LPM

RESPONSE TIME: 90% full scale in 2 seconds

NOISE: Less than 1% full scale

LINEARITY: Better than 1% full scale

REPEATABILITY: Better than 1% full scale      Relative Humidity: Less than 90% R.H.\*\*\*

ZERO SPIN DRIFT: Less than 1% full scale in 24 hours

WEIGHT: 30 lbs.

Dimensions: 5 ¼" H x 19" W x 23" D

Specifications subject to change without notice

\*Teflon is a registered trademark of DuPont

\*\*Tygon is a registered trademark of the Norton Performance Plastics Corp.

\*\*\*Non-condensing

### 3. Installation

#### 3.1. General

The design of this instrument is for industrial applications. These installation instructions are for a typical site. Direct any questions regarding specific installation situations to Technical Service of California Analytical Instruments, Inc.

#### 3.2. Site and Mounting

**NOTE: The carefully observe the following precautions:**

Select a site free from direct sunlight, radiation from a high temperature surface, or abrupt temperature variations.

This analyzer is not suitable for installation outdoors.

Select a site where the air is clean. Avoid exposing the instrument to corrosive or combustible gases.

The instrument must not be subject to severe vibration. If severe vibration is present, use isolation mounts.

The instrument is designed for rack-mounting. Optional rack mount slides are available.

Do not install near equipment emitting electromagnetic interference (EMI).


**NOTE: A rear supporting brace or equivalent is required if the optional rack mount slides were not purchased.**



***The power on/off switch is accessible from the rear of the instrument only. DO NOT mount such that the power on/off switch is inaccessible.***

#### 3.3. Electrical

All wiring is connected at the rear of the instrument. The AC power is connected to the power/fuse/switch as shown below



**Figure 3-1 AC Power Switch, Connector, and Fuse.**

**NOTE: A defective ground may affect the operation of the instrument. The output voltages are connected per Table 8.1.1. Shielded wiring is recommended for output signals.**



**Replace fuses with recommended fuse size indicated on rear panel of instrument. Replacement with any other size fuse may cause damage to the instrument and possible injury to operating personnel.**

### **3.4. Analog Output Connections (Appendix)**

See Appendix for connector pinouts located on the analyzer rear panel. Remote range identification and range selection are obtained via the rear panel connections. When a range is selected, the corresponding control line is pulled low to zero VDC. Ranges not selected will remain at approximately 5 VDC. When remote range control is selected on the front panel switch, a contact closure is provided at the rear panel connector. Remote range selection is made by connection of the control line for the desired range to the analyzers zero VDC line provided in the connector. Five VDC is also provided.

### **3.5. Gases**

- 1) Nitrogen or (zero air) in pressurized cylinder.
- 2) Standard span gas(es) near full-scale concentration with a nitrogen balance, in a pressurized, certified cylinder.
- 3) Pressure regulators for zero and span gas cylinders.
- 4) Corrosive resistant gas tubing.

#### **3.5.1. Sampling System**

**Note: High-pressure oxygen is very dangerous. Virtually any material will burn in it, possibly explosively. It is essential that any person using this analyzer is aware of the dangers of oxygen, and take all appropriate precautions.**

The analyzer's sampling system consists of:

1. An internally mounted in line particulate filter
2. A sample pump and flow meter (optional)
3. A sample capillary that controls the sample flow rate to the sensor at 0.5 LPM.
4. A precision controlled relief valve.

The relief valve maintains a constant inlet pressure to the sample capillary and reduces response time by providing a bypass loop to the exhaust for excess sample.

The analyzer is designed to measure a clean dry sample gas that has been conditioned to remove moisture to prevent condensation in the analyzer. Some applications may require additional sample conditioning, dependent upon the specifications of the sample gas to be measured.

### 3.6. Gas Handling Equipment

- 1) Pressure regulators for zero gas (Air or N<sub>2</sub>), and span gas cylinders.
- 2) Corrosive resistant gas tubing.

### 3.7. Gas Connections

The tubing from the sampling system to the gas analyzer should be made from corrosive-resistant material such as Teflon or stainless steel. Even when the gases being sampled are corrosive themselves, rubber, or soft vinyl tubing should not be used since readings may be inaccurate due to gas absorption into the piping material. To obtain fast response, the tube should be as short as possible. Optimum tube internal diameter is 0.16 inch (4 mm). Couplings to the instrument are 1/4-Inch tube.

**Note: Be sure tubing and joints are clean. Dust entering the instrument may cause it to malfunction.**

### 3.8. Sampling Requirements

#### 3.8.1. Filtration

Dust must be eliminated completely. Use filters as necessary. The final filter must be capable of removing particles larger than 4 microns.

#### 3.8.2. Condensation

Dew point of the sample gases must be lower than the instrument temperature to prevent accidental condensation within the instrument. Bypass the sample through a dehumidifier to reduce the dew point to about 2 to 4° C or less. If the sample contains an acid mist, use an acid mist filter, cooler or similar device to remove all traces of the mist.

#### 3.8.3. Presence of Corrosive Gases

High concentrations of corrosive gases such as Cl<sub>2</sub>, SO<sub>2</sub>, F<sub>2</sub>, HCl, in the sample gas shorten the useful service life of the instrument.

#### 3.8.4. Gas Temperature

Do not exceed the maximum rating of the instrument 104° F (50° C) when measuring high temperature gases.

#### 3.8.5. Sample Gas Bypass Outlet (Vent)


A sample gas bypass outlet connector is located on the rear panel (1/4 Inch Tube). Pressure at this outlet should be kept at atmospheric level. **ANY** backpressure will cause an error in reading.

## 4. Basic Operation

The operation of the digital microprocessor conforms to the guidelines of the AK committee, originally developed in the German automotive industry. Via the serial port of the MSR-Card, the analyzer can be remote-controlled by a master computer. The serial communication fully corresponds to the specifications of the AK protocol. TCP/IP communication is also available.

### Display

The analyzer's LCD display includes 16 lines with 30 characters each. The display also has background lighting that can be switched on and off via the Display key on the keyboard. The following example shows the measurement screen that is formatted into 4 information areas.




**Figure 4-1 LCD Display – example showing Model 602P**


- 1) THE TOP INFORMATION AREA CONTAINS:
- 2) The AK Protocol Information. This capability is for advanced uses and may be toggled on and off in the setup screen, F5. Next to the symbol for the active operating mode, the device status is indicated. The status field is also displayed on all other screens.
- 3) SARE Auto range enabled
- 4) SMGA Measuring gas is flowing
- 5) SMAN Device is in manual operation status
- 6) SWET Device
- 7) Shown on the right is the Password Entry level with 1 to 4 horizontal lines.
- 8) THE LARGE INFORMATION AREA CONTAINS the data portion of the screen.
- 9) THE THIRD INFORMATION AREA CONTAINS the help information for the parameter selected, ranges, etc.
- 10) THE LOWER INFORMATION AREA CONTAINS The time and date and any error condition.
- 11) The symbol in the bottom right corner indicates the keyboard mode. In the example shown, the keyboard is in the function key mode. For input fields, the mode is usually switched to numerical input. Then, an

N appears in the lower right of the screen. This symbol is displayed on all screens.


#### 4.1. Menu Trees


**Model 601 Menu Tree**






F4





## 4.2. Keyboard



**Figure 4-2 Keyboard**

### 4.2.1. Operation with the Cursor Keys and the Enter Key

When operating the unit with the cursor keys, you select the various functions with the up/down cursor keys and start them with the Enter key. This method is particularly suitable for less proficient users since the system displays a short on-line help for nearly every function selected. The actual cursor position is shown as a black horizontal bar.

**TIP:** *If you are not yet familiar with the screens and their fields, just press any cursor key after a screen appears. This moves the cursor from field to field and displays the corresponding online help.*

### 4.2.2. Operation with the Function Keys

When using the function keys (F1 though F10), functions are directly accessed by pressing their corresponding function keys. This method is suitable for the advanced user since it is faster than the operation with the cursor keys.

### 4.2.3. Read/Change Parameters

To read and/or change parameters, you must switch to the parameter input mode by pressing the Enter key after calling the corresponding parameter screen. The input cursor (horizontal bar under the first character) then appears in the active edit field (black background). The cursor can be positioned with the right and left cursor keys, and the value displayed (number or letter) can be changed with the up and down cursor keys or entered directly. Every input has to be concluded by pressing the Enter key again, which causes the cursor to disappear.

## 5. Operating Structure

The analyzer's operation can be divided into 4 operating levels. The current level is always displayed as a stack of 1 to 4 horizontal bars in the top right corner of the screen. In the access level menu, you can choose between the following operating levels:

F1	User	(operating level 1)
F2	Advanced user	(operating level 2)
F3	Maintenance	(operating level 3)
F4	System user	(operating level 4)

A password can be assigned to each operating level. Only the system user, who normally has the highest operating priority, can assign the password. At the factory, the default passwords for the CAI analyzers are set as follows:

User:	111
Advanced user:	222
Maintenance:	333
System:	444

The default setting can be changed only by the system user. This manual is written to include all information for the advanced system user.

**TIP: Because of the user settings, some of the parameters shown in this manual may not appear on your analyzer. Check the access level.**

### 5.1. The Main Menu

Upon power up, the CAI logo is first displayed and then the main menu appears as below:




Figure 5-1 Main Menu on Power Up Screen

**NOTE: F6 is not available because, on initial start up, the analyzer reverts to ONLY Level 1 access. See Section 7.5.5 for Password information.**

All functions can be selected with the cursor keys and activated by pressing the Enter key, or directly with the function keys F1 through F7. A ">" to the right of a function means that one or more sub-menus are available. If this sign is missing, the function starts immediately after the activation.

**NOTE: Access level is 4.**




Figure 5-2 Main User Menu (Level 4)

## 6. Menu Structure

There are four operating levels based on the level of your password. This section shows the access rights of the single levels.

### 6.1. User Functions (Level 1)

Main Menu		Setup Menu		Password Menu	
F1	Measurements	F5	Password	F1	Enter password
F2	Purge Analyzer	F10	Version		
F3	Diagnostics				
F4	Calibrations				
F5	Setup				
F7	Standby				

### 6.2. Advanced User Functions (Level 2)

Main Menu		Setup Menu		Password Menu	
F1	Measurements	F3	Range Limits	F1	Enter password
F2	Purge Analyzer	F5	Password		
F3	Diagnostics	F10	Version		
F4	Calibrations				
F5	Setup				
F7	Standby				

### 6.3. Maintenance Functions (Level 3)

Main Menu		Setup Menu		Password Menu		System Settings Menu	
F1	Measurements	F1	Span Gas Concentration	F1	Enter password	F1	Real Time Clock
F2	Purge Analyzer	F3	Range limits	F2	Reset password	F5	Status Line on/off
F3	Diagnostics	F5	Password	F7	Auto Startup		
F4	Calibrations	F7	System Settings				
F5	Setup	F8	Measure Settings				
F7	Standby	F10	Version				

### 6.4. System User Functions (Level 4)

All Function described in this manual may be accessed from Level 4.

## 7. Main Menu Function Descriptions

### 7.1. F1 Measurements




Figure 7-1 Main Menu Screen

#### 7.1.1. F1 Measurement

The measurements screen is activated by pressing F1 on the Main Menu screen. The concentration is displayed in actual engineering units.




Figure 7-2 Measurements Screen showing Model 602P

### 7.1.2. Range Select

With the arrow keys, the ranges 1 to 4 can be selected and locked in which will disable the auto range capability. Continue pressing the arrow keys will recycle the analyzer back to auto range. The range and/or auto range is displayed on the measurement screen. If the limits are exceeded while not in the auto range mode, a warning "Over Range" appears on the screen.




Figure 7-3 Set to Auto Range (601)



Figure 7-4 Analyzer set to Range 1 (603)

## 7.2. F2 Purge Analyzer




Figure 7-5 Main Menu (User Level 4)




Figure 7-6 Purge Screen

F2 from the Main Menu activates the Purge (analyzer) function if equipped.

## 7.3. F3 Diagnostics

F3 from the Main Menu activates the Diagnostics function. F3 brings up the two diagnostics screens. The Diagnostics screens may be brought up from **EITHER** the Main Menu or the Measurements screen.




Figure 7-7 Diagnostics Screen I

#### 7.4. F4 Calibrations

F4 from the Main Menu activates the Calibrations screen. Calibrations may be automatic or manual. Deviations can also be displayed. Calibration values can be reset to default values and the range to be calibrated can be changed.




Figure 7-8 Calibration Screen

##### 7.4.1. F1 Automatic Calibration

From the Calibrations screen, F1 starts automatic calibration. If auto range is selected, the actual range in use will be calibrated. Auto calibration works as follows: First, zero gas is purged a certain time, called purge-time. Then the measurement begins. The measured value must be a minimum-time, called measuring-time, and within an upper and a lower limit to be saved as new offset value. The maximum length of measuring time is 9 seconds. If the measured value was constant during calibration time, it is checked to determine if this value deviates from the preceding value. If the deviations are too large, a warning "Deviation error!" appears and the user can choose if the new value is saved or not. At last, the zero gas is flowed a further time, verifying time, so it can be checked if the signal is still constant. All of these times can be changed. After zero gas calibration, the same happens with span gas. During auto calibration, "Calibration in progress" is displayed. It also shows which gas is flowing and which time runs. When auto calibration has finished it is displayed. If the span value of the selected range is 0 (see section 5.6.1), then it will not be calibrated. If one range is calibrated and the span value for the lower ranges is zero, calibration parameters will be copied to this range. To calibrate all ranges with the same span gas, you must enter the gas concentration in the Span Gas Calibration screen for ALL RANGES. You must also calibrate each range. Offsets and scalors are NOT copied to other ranges.




Figure 7-9 Auto Calibration Screen

#### 7.4.2. F2 Manual Calibration

From the Calibration screen, F2 starts manual calibration. If auto range is selected, calibration is not possible, and the appropriate range can be selected.



Figure 7-10 Manual calibration

In the manual calibrations menu, two options are possible:

F1 Flow zero gas  
F2 Flow span gas




Figure 7-11 Manual zero calibration





Figure 7-12 Manual span calibration

When zero or span gas is flowing, the measured value can be saved by pressing F1. If the screen is left by pressing the buttons "Main" or "Back", the measured value is not saved. Solenoids are closed by pressing F2.

#### 7.4.3. Display Deviations – from Calibration menu F3




**Figure 7-13 Display deviations**


After every calibration, the deviations are calculated for zero and for span gas.

- F1 Zero gas deviations
- F2 Span gas deviations
- F3 Deviations of zero gas during verifying
- F4 Deviations of span gas during verifying

F1 and F2 deviations are displayed in percent.



**Figure 7-14 Zero gas deviations**



**Figure 7-15 Span gas deviations**

During calibration, there is verification for zero and span gas. With option F3 and F4, you can view the deviations during the verification time. Absolute deviation is the absolute average difference from the saved value in ppm. Relative deviation is the absolute average difference in percent, related to the range limit.

#### 7.4.4. Absolute Zero Gas Deviation

Absolute zero gas deviation is zero gas content calculated by the factory polynomial related to the range limit of the calibrated range.

#### 7.4.5. Relative Zero Gas Deviation

Relative zero gas deviation is the actual deviation minus the deviation of the previous calibration related to the range limit of the calibrated range.

**7.4.6. Absolute Span Gas Deviation**

Absolute span gas deviation is span gas bottle value minus span gas value calculated by the factory-polynomial related to the range limit of the calibrated range.

**7.4.7. Relative Span Gas Deviation**

Relative span gas deviation is the actual deviation minus the deviation of the previous calibration related to the range limit of the calibrated range.

**7.4.8. F4 Check Calibration**

There is a default calibration. Pressing F4 activates an automatic zero and span check for verification.

**7.4.9. F5 Reset Calibration Values**

There is a default calibration. Pressing F5, a new screen appears and asks if the user is sure to reset calibration values to the default calibration values. F1 confirms and the calibration values are reset to default calibration values. F2 leaves this menu without resetting to default values. This function will overwrite all calibrations with factory values. In addition, the linearization polynomial will be overwritten with the factory values.

**7.4.10. F6 Range Select**

This allows a range change to be activated from the calibration menu.

### 7.5. F5 Setup

From the Main Menu, F5 brings up the setup menu. Span gas concentrations, calibration settings, range limits, alarms, password, linearization, system and measure settings can be changed. The Setup menu begins as shown below. A description of each parameter is shown in the information box. NOTE: Use the down arrow key to obtain the additional setup parameters.




Figure 7-16 Setup menu screen I




Figure 7-17 setup menu screen II

#### 7.5.1. F1 Span Gas Concentration

**Note: If you do not have a specific gas for a specific range, the calibration will use the previous ranges calibration. You must have a least one span gas.**

For calibration, it is necessary to input the concentration of the span gas in ppm. For every range, the span gas concentration can be changed. After pressing F1 in the setup menu, a screen appears in which changes can be made. Select with the cursor buttons the range to change. The selected field turns black. To change parameters, switch to parameter input mode by pressing the Enter key. The input cursor (horizontal bar under the first character) then appears in the active edit field (black background). The cursor can be positioned with the right and left cursor keys, and the value displayed (number or letter) can be changed with the up and down cursor keys or entered directly. Every input has to be concluded by pressing the Enter key again. Then the input cursor disappears and a new range can be selected. The changes are saved by leaving the screen by pressing "Main" or "Back." At the right side of the screen, the range limits of the 4 ranges are displayed. They cannot be changed in this screen.




Figure 7-18 Change Span Gas Settings

### 7.5.2. Calibration Settings




Figure 7-19 Change Auto Calibration Settings

In the calibration settings menu, times and, deviations can be changed.

#### 7.5.3. F1 Times

There are four times (in seconds) for auto calibration that can be changed. Purge, measuring, calibration and verifying time. Changes are made and saved as above.

#### 7.5.4. F2 Measuring Deviations

During auto calibration, the measured value is only saved if it is within a certain time within an upper and a lower limit. These two limits format a working window. In the setup menu, the deviation is in percent.

#### 7.5.5. F3 Deviations

Here you can change absolute and relative deviation in percent. After auto calibration, it is checked to assure the deviations are within this limit. If the deviations are not in this limit, a warning "Deviation error!" appears.

#### 7.5.6. F3 Range Limits

There are 4 different ranges. The user can define the upper range limits in ppm.



Figure 7-20 Change Range Limits

#### 7.5.7. F1 Range 1-4 (Change Upper Range Limits)

In this menu, the upper range limits can be changed. The new settings are saved by pressing MAIN or BACK. The auto range limits are automatically adapted. This means that if the upper range limit of range 1 for example has reached 90% of the upper range limit in the auto range mode, it is switched automatically to the second range.




Figure 7-21 Change Upper Range Limits

#### 7.5.8. F2 Change Auto Range Limits

Although the auto range limits are adapted automatically, it is possible to define them manually. Up means the value when the next higher range is selected in auto range mode, down the value when the next lower range is selected.




Figure 7-22 Change Auto Range Limits

### 7.5.9. F4 Alarms

Error reports are always displayed in the lowest line of the screen. There are two pressures, three temperatures, one concentration, and two voltages with alarm limits that can be defined. The user can define the range limits and, if exceeded, will display an error-message.




Figure 7-23 Alarm screen I




Figure 7-24 Alarm screen II

Set Temperature Alarms Set Concentration, Pressure and Voltage Alarms

### 7.5.10. F5 Password

After turning on the analyzer, you are in access level 1. To change the access level or to change the passwords, press F5 (Setup) in the main menu and Press F5 (Password) again. The following screen appears:



Figure 7-25 Enter / Change Password

### 7.5.11. F1 Enter Password

To change access level, press F1. The following screen appears:



Figure 7-26 Access Level Screen

F1 to F4 selects an access level. Move the cursor to the access level to be modified. You must enter the correct password for the access level desired. The passwords for the various operation levels consist of three numbers that must be entered on the numeric keypad. If the code word is incorrect, you are asked to re-enter the codeword.

**IMPORTANT TIP:** When a new analyzer is powered up, it defaults to access level 1 (User). To operate ALL parameters and gain complete access, select F4. Press the Enter key twice and enter 444.

### 7.5.12. F2 Change Password

The passwords can only be changed, if you are in access level 4. After F2, enter your new 3 digit passwords.

**IMPORTANT TIP:** You MUST remember and record this new password. If this is lost, you will need to consult the factory for the default password!

### 7.5.13. F3 Reset Passwords

The passwords can only be changed, if you are in access level 4. Reset passwords will revert to the factory defaults.

#### 7.5.14. F6 Linearization

Pressing F6 on the Setup screen brings up the Linearization screen. The analyzer can be linearized by a polynomial with 5 coefficients. By pressing F1, these 5 coefficients can be changed for each range. By pressing F2, the raw value can be displayed. This is the value before linearization and offset span correction. There are two values on the screen: The value at the top is the linearized, offset-span-corrected value, and the other value is the raw-value.




Figure 7-27 Linearization Screen




Figure 7-28 Coefficients Range Select




Figure 7-29 Change Coefficients




Figure 7-30 Linearized and raw data

## 7.6. F7 System Settings

This screen allows all the system settings to be displayed and modified.




Figure 7-31 System Setup Screen

### 7.6.1. F1 Real Time Clock

This brings up the clock time set screen; auto cal and auto cal enable screens.




Figure 7-32 Clock and Timing Setup Screen



Figure 7-33 Clock set screen

The current time may be set by using the cursor to highlight the entry and using the numeric keys to change the values.

F2 brings up the auto cal time set. As above, the date and times can be set by using the cursor to highlight the entry and using the numeric keys to change the values. F3 Sets autocalibration ranges.




Figure 7-34 Set Auto Cal Timing




Figure 7-35 Set Auto Cal Ranges







Figure 7-36 F4 Toggles Auto Cal ON of OFF

**7.6.2. System Setup F2 Displays TCP/IP Address****Figure 7-37 TCP/IP Address****7.6.3. Systems Setup F3 Displays Output Signal Assignments**

(Used to Adjust Analog Output Channels)

**Figure 7-38 Output Assignments****7.6.4. System Setup F4 Displays Output Ranges**

(Used to Adjust Scale of Analog Output Channels)

**Figure 7-39 Output Ranges**

### 7.6.5. F5 Turns Status Line On or Off

The status line displays the AK Protocol action on the top line of the display.

K3: SMGA

Figure 7-40 Status line




Figure 7-41 Status line on/off

### 7.6.6. F8 Measure Settings

This screen allows several of the system settings to be displayed and modified.




Figure 7-42 Measure setup

### 7.6.7. F1 Set Lowpass filter




Figure 7-43 Lowpass filter

#### 7.6.8. F2 Purge Time

F2 on the Menu Settings screen sets the purge time before continuing with a zero or span calibration.



Figure 7-44 Purge Time

#### 7.6.9. F3 Set Temperature Compensation




Figure 7-45 T-compensation

#### 7.6.10. F4 Pressure Compensation



Figure 7-46 P-Compensation

### 7.6.11. F3 Low Pass Filter Time Constant

F3 on the Menu Settings screen allows the software time constant to be set between 1 and 60 seconds. This is very useful in eliminating noise when measuring low-level concentrations.




Figure 7-47 Low pass filter time constant

### 7.6.12. F10 Displays the Current Analyzer and Software Versions

This displays the analyzer's information, including the factory recommended air and sample pressure settings.




Figure 7-48 Analyzer Information




Figure 7-49 Software Version

### 7.7. F7 Remote / Manual Control

The analyzer can be remote-controlled either by a master computer or via contact closures. The TCP/IP and serial communication fully corresponds to the specifications of the AK protocol. To change remote/manual control, press F6 in the main menu. This toggles between remote and manual control.




Figure 7-50 F6 Remote/Manual

#### Main Menu (User Level 4)

### 7.8. F8 Standby



Figure 7-51 F7 Standby

#### Main Menu (User Level 4)

In Standby mode, pump is turned off and the solenoids are closed. The CAI logo is displayed.





Figure 7-52 Standby Mode

## 8. Analyzer Components

### 8.1. Rear Panel

The following details the rear panel connections:



**Figure 32: Rear Panel**

- 1) **Sample Gas Inlet:** Feeds sample gas to the analyzer. 1/4 Inch Tube.
- 2) **Sample Gas Bypass Outlet (Vent):** Exhaust for sample. 1/4 Inch Tube.
- 3) **TCP/IP Connection:** Connect Network Connector.
- 4) 28 Pin output terminal strip/main 1 (standard).
- 5) 28 Pin output terminal strip/main 2 (optional).
- 6) 28 Pin output terminal strip/main 3 (optional).
- 7) **Power Entry Module:** Power connection, power switch, fuse compartment (2 Amp) **With Rear Panel Power ON/OFF Switch:** Turns ON/OFF line power to instrument
- 8) **Serial Connector:** Connect Serial Connector

## 8.2. Rear Panel Connectors

<b>28 Pin Main Connector Assignments:</b>						<b>28 Pin Auxiliary Connector Assignments:</b>			
<b>Signal Type</b>	<b>Main 1 Analog</b>		<b>Main 2 Optional Analog</b>		<b>Main 3 Optional Analog</b>		<b>Signal Type</b>	<b>600P Analog</b>	
	Pin #		Pin #		Pin #		Spare	Pin #	
A Output	1	GND (analog)	1	GND (analog)	1	GND (analog)	A Input	1	GND (analog)
A Output	2	Channel 1	2	Channel 1	2	Channel 1	A Input	2	External Analog 1
A Output	3	Channel 2	3	Channel 2	3	Channel 2	A Input	3	External Analog 2
A Output	4	Channel 3	4	Channel 3	4	Channel 3	A Input	4	Spare analog
A Output	5	Aux	5	Aux	5	Aux	A Input	5	Spare analog
	<b>Digital</b>		<b>Digital</b>		<b>Digital</b>		<b>Alarms</b>		<b>Digital</b>
D Output	6	GND (Digital)		GND (Digital)		GND (Digital)	D Output	6	GND (Alarm)
D Output	7	Sense AutoRange		Sense AutoRange		Sense AutoRange	D Output	7	General Alarm
D Output *	8	Sense Range 4		Sense Range 4		Sense Range 4	D Output	8	Ch. 1 Conc. 1 Alarm
D Output *	9	Sense Range 3		Sense Range 3		Sense Range 3	D Output	9	Ch. 1 Conc. 2 Alarm
D Output *	10	Sense Range 2		Sense Range 2		Sense Range 2	D Output	10	Ch. 2 Conc. 1 Alarm
D Output *	11	Sense Range 1		Sense Range 1		Sense Range 1	D Output	11	Ch. 2 Conc. 2 Alarm
	* Hi = True								
D Input	12	Set AutoRange	12	Set AutoRange	12	Set AutoRange	D Output	12	Ch. 3 Conc. 1 Alarm
D Input	13	Control Range 1	13	Control Range 1	13	Control Range 1	D Output	13	Ch. 3 Conc. 2 Alarm
D Input	14	Control Range 2	14	Control Range 2	14	Control Range 2	D Output	14	Reserved
D Input	15	Control Range 3	15	Control Range 3	15	Control Range 3	D Output	15	GND (Alarm)
D Input	16	Control Range 4	16	Control Range 4	16	Control Range 4	D Output	16	Calibration Alarm 1
D Input	17	Auto Cal	17	Auto Cal	17	Auto Cal	D Output	17	Calibration Alarm 2
D Input	18	Calibrate	18	Calibrate	18	Calibrate	D Output	18	Calibration Alarm 2
D Input	19	Zero	19	Zero	19	Zero	D Output	19	Reserved
D Input	20	Span	20	Span	20	Span	D Output	20	Spare
D Input	21	Pump	21	Pump	21	Pump	D Output	21	Spare
D Output	22	Zero Gas Flow	22	Zero Gas Flow	22	Zero Gas Flow	D Output	22	Spare
D Output	23	Span Gas Flow	23	Span Gas Flow	23	Span Gas Flow			Spare
D Output	24	Sample Gas Flow	24	Sample Gas Flow	24	Sample Gas Flow			Spare
D Output	25	Local/Remote	25	Local/Remote	25	Local/Remote			Spare
D Output	26	Read Cal Mode	26	Read Cal Mode	26	Read Cal Mode			Spare
D Output	27	Reserved	27	Reserved	27	Reserved			Spare
D Output	28	Reserved	28	Reserved	28	Reserved			Spare

### 8.3. Digital Outputs

#### 8.3.1. RS-232 (Standard 9 Pin DIN Connector)

Pin	Function
1	DCD Carrier Detect
2	RxD Receive Data
3	TxD Transmit Data
4	DTR Data Terminal Ready
5	Ground
6	DSR Data Set Ready
7	RTS Ready to Send
8	CTS Clear to Send
9	RI Ring Indicator


#### 8.3.2. TCP/IP (8 Pin RJ-47 Connector)


Pin	Function
1	TDX+
2	TDX-
3	RXD+
4	Open
5	Open
6	RXD-
7	LNLED
8	LNLED

**IMPORTANT TIP:** For direct connect to a PC a crossover cable is required. Connection to a hub requires a straight cable.

**8.4. Internal Component Locations**

PICTURE NOT AVAILABLE



**8.5. Main Electronics Board (Potentiometers)****Figure 8-2 Main Electronic Board Potentiometers**

- |   |     |                         |
|---|-----|-------------------------|
| 1 | RP2 | Detector voltage adjust |
| 2 | RP3 | Fine zero adjust        |
| 3 | RP4 | Not used (Fully CW)     |
| 4 | RP5 | Coarse zero adjust      |
| 5 | RP6 | Span adjust             |

**NOTE: Potentiometers are clearly labeled on both sides of the Wave PCB.**

## 8.6 Mechanical Zero Adjustment

*This adjustment may be periodically required over the life span of the analyzer, or whenever the front panel zero adjustment has reached its limit.*

**Note:** Always check for gas leaks prior to making this adjustment, especially when the zero potentiometer is at its counter-clockwise limit.

- a) Place the front panel zero adjustment to its mid-setting (5.0 on the dial).
- b) Introduce N<sub>2</sub> into the analyzer at a flow rate of 0.5-2.0 LPM.
- c) Remove the screws securing the top to the chassis and lift off the cover of the analyzer.
- d) Locate and remove the rubber 'boot' that covers the O<sub>2</sub> sensor.
- e) Loosen the locking screw that secures the adjusting screw for positing the photo-sensor on the detector assembly. Only loosen the locking screw enough to permit the necessary adjustment (see Figure 8-1).
- f) Turn the adjusting screw to give an indication of approximately 0.00 on the digital display.
- g) Carefully tighten the locking screw and replace the rubber 'boot' removed in step d).
- h) Observe the digital display and turn the front panel zero adjustment to achieve reading of 0.00 in the display.
- i) The zero adjustment should be between 4.0 and 6.0 on its dial. If not, repeat steps d) through h) until no further adjustment is required




Figure 8-1 Oxygen sensor adjustment

## 9. Operation

### 9.1. External Wiring

Make sure that the external wires have been connected as described in Section 3 Installation.

### 9.2. External Piping

Review Section 3, 3.7 and 3.8

### 9.3. Operation & Calibration

#### 9.3.1. Power On:

Turn on the power switch (located on the rear panel). The LCD display should illuminate. Allow the instrument to warm up for approximately one hour. It is preferable, but not essential, that zero gas flow through the instrument at a rate of about 1 Liter/minute (LPM).

***Note: To achieve final stability, the analyzer may require an additional warm-up period of up to four hours (depending on variables in the analyzer's environment).***

#### 9.3.2. Zero/San Adjustment:

Follow the "Quick Start Guide" at the front of this manual.

***Note: Span gas concentration should not be less than 80% of the range to be spanned.***

#### 9.3.3. Start-Up:

Prepare and check the sample system. Adjust the flow of sample gas to about 1 L/min. The instrument should show a meter indication. The infrared gas analyzer is designed for extended operation and may be left switched on continuously.

## 10. MAINTENANCE

### Warning

All replacement parts must be as supplied and/or specified by California Analytical Instruments. Failure to used specified parts may reduce the safety features of the instrument or create a hazardous condition.

#### 10.1. Zero and Span Calibration

The zero and span levels should be checked and/or calibrated daily (or as often as required.)

**Note:** On the 0-25% range of the analyzer ambient air may be used as span gas. While flowing ambient air to the analyzer adjust the span potentiometer to 20.9% O<sub>2</sub>.

#### 10.2. Routine Maintenance:

Prepare and check the sample system. Adjust the flow of sample gas to about 1 L/min. The instrument should show a meter indication. The paramagnetic oxygen analyzer is designed for extended operation and may be left switched ON continuously.

#### 10.3. Cross sensitivity of gases

The paramagnetic measuring principle is based on the very high magnetic susceptibility of oxygen. In comparison to oxygen, other gases have such a minor susceptibility that most of them are insignificant. Exceptions to this are the nitrogen oxides. However, as these gases are in most cases present in a very low concentration, the error is still negligible.

##### 10.3.1. Example 1

The residual oxygen percentage is measured in a closed carbon dioxide (CO<sub>2</sub>) atmosphere. The "zero calibration" is done by means of nitrogen (N<sub>2</sub>)-0.1080

According to the list of cross-sensitivities, the error for 100 % CO<sub>2</sub> at 20°C is 0.27%. In order to obtain a higher accuracy, this means that for the zero calibration the reading should be adjusted at +0.27% with N<sub>2</sub>, in order to compensate the error of CO<sub>2</sub>.

Since the values of cross-sensitivities are based on 100% volume of that particular gas, the error at 50% by volume CO<sub>2</sub> and 50% by volume N<sub>2</sub> is 0.135%.

##### 10.3.2. Example 2

Given the following gas composition at a temperature of 20°C:

5% volume Oxygen (O <sub>2</sub> )	+100.00 x 10 <sup>-2</sup> x 5 = +5.0000
40% volume Carbon Dioxide(CO <sub>2</sub> )	-0.27 x 10 <sup>-2</sup> x 40 = -0.1080
1% volume Ethane(C21-14)	-0.43 x 10 <sup>-2</sup> x 1 = -0.0043
54% volume Nitrogen (N <sub>2</sub> )	0.00 X 10 <sup>-2</sup> x 54 = 0.0000
Gives a reading (% by volume) of:	+4.8877

As this example shows, the total error (5.000 minus 4.8877) is 0.1123.

**Note:** see Table 4-1 below for cross sensitivity values of typical gases.

## 11. ADJUSTMENTS CHECKS AND REPAIRS

### 11.2. Coarse Zero Adjustment

With zero gas still flowing connect a DC voltmeter to TP7 (+) and TP9 (common). Switch JP8 to the on position and leave JP6 and JP7 in the off position. The target DC voltage for TP7 is 0mV. Adjust RP5 first and use RP3 for the final adjustment.

### 11.3. Span Gain Adjustment

With a full-scale span gas flowing connect a DC voltmeter to TP8 (+) and TP9 (common). The target voltage for TP8 with a full-scale span gas flowing is 8.00VDC. Turn RP4 fully clockwise (12 turns). To achieve this voltage select a resistor value, or combination of values to install into R30, R35, R36, and / or R37. Set the corresponding Jumper (JP9-JP12) to the on position for each resistor used to set the gain. The value needed may range between 1K and 100K ohms. A smaller resistor value increases the gain and achieves a larger DC voltage at TP8 when span gas is flowing. RP6 can be used to make a small gain adjustment when the target voltage is close. When finished setting the gain adjustment, flow zero gas again and repeat the previous step to achieve a good zero at TP8.

**Table 4-1 Cross Sensitivity of gases**

All values based on nitrogen 0% / oxygen 100%

Gas	Formula	20 °C	50 °C
Argon	Ar	-0.23	-0.25
Acetylene	C <sub>2</sub> H <sub>2</sub>	-0.26	-0.28
Acetone	C <sub>3</sub> H <sub>6</sub> O	-0.63	-0.69
Acetaidehyde	C <sub>2</sub> H <sub>4</sub> O	-0.31	-0.34
Ammonia	N <sub>3</sub>	-0.17	-0.19
Benzene	C <sub>6</sub> H <sub>6</sub>	-1.24	-1.34
Bromine	Br <sub>2</sub>	-1.78	-1.97
Butadiene	C <sub>4</sub> H <sub>6</sub>	-0.85	-0.93
Isobutylene	(CH <sub>3</sub> ) <sub>2</sub> CH=CH <sub>2</sub>	-0.94	-1.06
n-Butane	C <sub>4</sub> H <sub>10</sub>	-1.10	-1.22
Chlorine	Cl <sub>2</sub>	-0.83	-0.91
Hydrogen Chloride	HCl	-0.31	-0.34
Nitrous Oxide	N <sub>2</sub> O	-0.20	-0.22
Diacetylene	(CHCl) <sub>2</sub>	-1.09	-1.20
Ethane	C <sub>2</sub> H <sub>6</sub>	-0.43	-0.47
Ethylene Oxide	C <sub>2</sub> H <sub>4</sub> O <sub>2</sub>	-0.54	-0.60
Ethylene	C <sub>2</sub> H <sub>4</sub>	-0.20	-0.22
Ethylene Glycol	CH <sub>2</sub> OHCH <sub>2</sub> OH	-0.78	-0.88
Ethylbenzene	C <sub>8</sub> H <sub>10</sub>	-1.89	-2.08
Hydrogen Fluoride	HF	+0.12	+0.14
Furan	C <sub>4</sub> H <sub>4</sub> O	-0.90	-0.99
Helium	He	+0.29	+0.32
n-Hexane	C <sub>6</sub> H <sub>14</sub>	-1.78	-1.97
Krypton	Kr	-0.49	-0.54
Carbon Monoxide	CO	-0.06	-0.07
Carbon Dioxide	CO <sub>2</sub>	-0.27	-0.29
Methane	CH <sub>4</sub>	-0.16	-0.17
Methanol	CH <sub>3</sub> OH	-0.27	-0.31
Methylene Chloride	CH <sub>2</sub> Cl <sub>2</sub>	-1.00	-1.10
Neon	Ne	+0.16	+0.17
n-Octane	C <sub>8</sub> H <sub>18</sub>	-2.45	-2.70
Phenol	C <sub>6</sub> H <sub>5</sub> O	-1.40	-1.54
Propane	C <sub>3</sub> H <sub>8</sub>	-0.77	-0.85
Propylene	C <sub>3</sub> H <sub>6</sub>	-0.57	-0.62
Propene	CH <sub>3</sub> CH=CH <sub>2</sub>	-0.58	-0.64
Propylene Oxide	C <sub>3</sub> H <sub>6</sub> O	-0.90	-1.00
Propylene Chloride	C <sub>3</sub> H <sub>7</sub> Cl	-1.42	-1.44
Silane	SiH <sub>4</sub>	-0.24	-0.27
Styrene	C <sub>7</sub> H <sub>6</sub> =CH <sub>2</sub>	-1.63	-1.80
Nitrogen	N <sub>2</sub>	-0.00	-0.00
Nitrogen Monoxide	NO	+42.70	+43.00
Nitrogen Dioxide	NO <sub>2</sub>	+5.00	+16.00
Oxygen	O <sub>2</sub>	+100.00	+100.00
Sulfur Dioxide	SO <sub>2</sub>	-0.18	-0.20
Sulfur Fluoride	SF <sub>6</sub>	-0.98	-1.05
Hydrogen Sulfide	H <sub>2</sub> S	-0.41	-0.43
Toluene	C <sub>7</sub> H <sub>8</sub>	-1.57	-1.73
Trichloroethylene	C <sub>2</sub> HCl <sub>3</sub>	-1.56	-1.72
Vinyl Chloride	C <sub>2</sub> H <sub>3</sub> Cl	-0.68	-0.74
Vinyl Fluoride	CH <sub>3</sub> F	-0.49	-0.54
Water	H <sub>2</sub> O	-0.03	-0.03
Hydrogen	H <sub>2</sub>	+0.23	+0.26
Xenon	Xe	-0.95	-1.02

## **11.8. Check and Repair Tubing Trouble**

### **11.8.1. Problem:**

Tubing loosened, disconnected, contaminated, or restricted.

### **11.8.2. Symptom**

Indicator output unstable or response is too slow.

### **11.8.3. Check and/or replace:**

- 1) When tubing is disconnected or loosened, firmly reconnect it.
- 2) When tubing is contaminated or restricted, disconnect it and blow out contaminants with high-pressure air, or replace the tubing.

### **Caution**

**Cracked, broken or frayed tubing may cause a hazardous condition. Replacement tubing must be as specified by California Analytical Instruments to prevent possible damage to operating personnel or equipment.**




## 12. Communication Master Computer / Analyzer (AK Protocol)

### 12.1. Serial Interface and AK-Commands

The serial interface enables remote control of the analyzer by a master computer. It is implemented as an RS232 V24 interface and meets all requirements of the AK protocol.

A 9-pin male connector at the back of the unit is used to connect a master computer with the following pin assignment:



Pin 2 = Rxd (receive)

Pin 3 = Txd (transmit)

Pin 5 = Gnd (ground)

**Figure 12-1 Serial Interface**

### 12.2. Interface Specifications

Speed:	9600 bps
Character Length	1 start bit 8 data bits 1 stop bit
Parity:	none
Handshake	no

### 12.3. Protocol Description

#### 12.3.1. Instruction command

	Character	Explanation
1 <sup>st</sup> Byte	STX	ASCII code 02
2 <sup>nd</sup> Byte	Don't Care	Any ASCII code
3 <sup>rd</sup> Byte	Function Code 1	AK instruction e.g.: ASTF
4 <sup>th</sup> Byte	Function Code 2	
5 <sup>th</sup> Byte	Function Code 3	
6 <sup>th</sup> Byte	Function Code 4	
7 <sup>th</sup> Byte	Blank	
8 <sup>th</sup> Byte	K	
9 <sup>th</sup> Byte	0	
10 <sup>th</sup> Byte	Blank	
	D	AK instruction parameters, length is variable
	A	
	T	
	A	
nth Byte	ETX	ASCII code 03

Table 12-1 Structure of an instruction command

#### 12.3.2. Acknowledgement command

	Character	Explanation
1 <sup>st</sup> Byte	STX	ASCII code 02
2 <sup>nd</sup> Byte	Don't Care	Any ASCII code
3 <sup>rd</sup> Byte	Function Code 1	Echo of the AK instruction command
4 <sup>th</sup> Byte	Function Code 2	
5 <sup>th</sup> Byte	Function Code 3	
6 <sup>th</sup> Byte	Function Code 4	
7 <sup>th</sup> Byte	Blank	
8 <sup>th</sup> Byte	K	
9 <sup>th</sup> Byte	0	
10 <sup>th</sup> Byte	Blank	
	D	AK acknowledgement parameters, length is variable
	A	
	T	
	A	
nth Byte	ETX	ASCII code 03

Table 12-2 Structure of an acknowledgement command

### 12.3.3. Data Description

Each command begins with STX (Start of Text) in the first byte. The “don't care” byte can be any ASCII character. Generally, a blank or an underscore (\_) is used for readability reasons. The four function bytes represent the AK command. A blank comes next, followed by K and the channel number. The analyzer is a single-channel device, and because of that the channel number is usually 0. For delimiting the command parameters from the channel number, another blank follows. This may be followed by command parameters with a variable length. Every command ends with the ETX (End of Text) character. The error status byte in the acknowledgment command signals if internal errors in the analyzer occurred. It is zero when no error appeared, and it is unequal zero when one or more errors occurred. Every time a change in the errors happens the error status byte is incremented by one, no matter if one or several errors disappear or are added. If it had the value 10, it would be reset to 1. The error status byte does not indicate the real number of errors. If the analyzer does not have errors, the error status byte contains the value 0.

In general, AK commands are subdivided into three classes:

- Control commands (Sxxx)
- Inquiry commands (Axxx)
- Configuration commands (Exxx)

### 12.3.4. Error Handling

It might happen that an unknown instruction is sent, that the analyzer is busy with a function which is not the desired one, or that an error occurred in the command parameters. Table 1-4 summarizes all errors that can appear upon any master instruction.

Analyzers Acknowledgement	Explanation
???? f1	Analyzer does not know the instruction sent.
xxxx <sup>2</sup> f BS	Analyzer is busy with another function.
xxxx f SE	Syntax error within command parameters or incomplete command.
xxxx f NA	Requested function or data are not available.
xxxx f DF	Data error: The kind or number of given parameters are not valid.
xxxx f OF	Offline: Analyzer is offline, i.e. analyzer is in local Mode. Only inquiry commands and SREM (set analyzer in remote mode) are allowed.

Table 12-3 Acknowledgment response in case of error

<sup>1</sup> f stands for the error status byte.

<sup>2</sup> xxxx stands for the function code of the sent master command

#### 12.4. Scan Commands

##### 12.4.1. AKON: Measured concentration value

Command	Response	Description
_AKON_K0	_AKON_s_z.z_y.y_x.x_w.w	Measured concentration value of all channels is responded t = Timestamp (1/10 sec)
_AKON_Km	_AKON_s_z.z_t	Measured concentration value of channel m is responded t = Timestamp (1/10 sec)

##### 12.4.2. AEMB: Get measuring range

Command	Response	Description
_AEMB_K0	_AEMB_s_Mn_Mn_Mn	Current measuring range of all channels is responded
_AEMB_Km	_AKON_s_Mn	Current measuring range of channel m is responded

##### 12.4.3. AMBE: Measuring range limit

Command	Response	Description
_AMBE_K0	_AMBE_s_M1_w.w _M2_x.x _M3_y.y _M4_z.z	All existing measuring range limits of channel m are responded
_AMBE_K0_Mn	_AMBE_s_Mn_z.z	Range limit of Range Mn is responded

##### 12.4.4. AKAK: Calibration gas concentrations

Command	Response	Description
_AKAK_Km	_AKAK_s_M1_w.w _M2_x.x _M3_y.y _M4_z.z	All existing calibration gas values are responded for selected channel m
_AKAK_Km_Mn	AKAK_s_Mn_z.z	For selected channel m calibration gas value of Range Mn is responded

##### 12.4.5. AMBU: Upper and lower range switchover values for auto range

Command	Response	Description
_AMBU_Km	_AMBU_s_M1_w.w_W.W _M2_x.x_X.X _M3_y.y_Y.Y _M4_z.z_Z.Z	Lower and upper range switchover value of auto range are responded for channel m
_AMBU_Km_Mn	_AMBU_s_Mn_w.w_W.W	Lower and upper range switchover value of auto range are responded for channel m range n

**12.4.6. ASTZ:** Normal device status

Command	Response	Description	
_ASTZ_K0	_ASTZ_s_K1_State1_State2_State3 _K2_State1_State2_State3 _K3_State1_State2_State3	Respond device status for all channels	
_ASTZ_Km	_ASTZ_s_State1_State2_State3		
_ASTZ_Km	_ASTZ_s_State1_State2_State3	Respond device status only for channel m	

**Possible states:**

State 1	State 2	State 3
SREM: remote	STBY: standby	SARE: auto range on
SMAN: manual	SPAU: pause	SARA: auto range off
	SMGA: measuring gas	
	SNGA: zero gas	
	SEGA: end gas	
	SATK SNGA: zero gas during auto cal	
	SATK SEGA: end gas during auto cal	

**12.4.7. ASTF:** Error status

Command	Response	Description
_ASTF_K0	_ASTF_s_f1_f2_f3_..._f10	Current error numbers of all are responded

**Errors:**

1	Channel 1 Flow Failure	8	Channel 1 not calibrated
2	Channel 2 Flow Failure	9	Channel 2 not calibrated
3	Channel 3 Flow Failure	10	Channel 3 not calibrated
4	External Analog 1 Failure	11-13	Ch1...3: Low concentration warning
5	External Analog 2 Failure	14-16	Ch1...3: High concentration warning
6	Pressure Failure	17-19	Ch1...3: Temperature failure
7	Temperature Failure	20-22	Ch1...3: EPC Voltage failure

**12.4.8. AKEN:** Device identification

Command	Response	Description
_AKEN_K0	_AKEN_s_devicename	Device identification is responded
_AKEN_K1	_AKEN_s_model	Device model
_AKEN_K2	_AKEN_s_serialno	Device serial number
_AKEN_K3	_AKEN_s_samplepressure	Suggested input sample pressure

**12.4.9. ARMU:** Raw value

Command	Response	Description
_ARMU_K0	_ARMU_s_z.z_y.y_x.x_t	Raw value before linearization and offset-span-correction is responded for all channels t = Timestamp (1/10 sec)
		Raw value before linearization and offset-span-correction is responded for channel m t = Timestamp (1/10 sec)

**12.4.10. ATEM: Temperatures**

Command	Response	Description
_ATEM_K0	_ATEM_s_z.z	Device temperature in degrees Celsius is responded
_ATEM_Km	_ATEM_s_z.z	Detector temperature of channel m is returned in z.z

**12.4.11. ADRU: Pressures/ Valve voltage**

Command	Response	Description
_ADRU_K0	_ADRU_s_z.z	Pressure in is responded
_ADRU_Km	_ADRU_s_z.z	EPC voltage of channel m is returned in z.z

**12.4.12. ADUF: Flows**

Command	Response	Description
_ADUF_K0	_ADUF_s_z.z_y.y_x.x	Sample gas flow of all channels is responded
_ADUF_Km	_ADUF_s_z.z	Sample gas flow of channel m is responded

**12.4.13. AGRD: Polynom coefficients**

Command	Response	Description
_AGRD_Km_Mn	_AGRD_s_Mn_a0_a1_a2_a3_a4	Polynomial coefficients of channel m range Mn are responded

**12.4.14. AANG: Deviation from zero point after autocalibration**

Command	Response	Description
_AANG_Km	_AANG_s_M1_z.z_da_dr _M2_z.z_da_dr _M3_z.z_da_dr _M4_z.z_da_dr	Verifying deviations from zero point after auto calibration. Values: measured value (z.z), absolute dev (da), relative dev (dr)

**12.4.15. AAEG: Deviation from end point after autocalibration**

Command	Response	Description
_AAEG_Km	_AANG_s_M1_z.z_da_dr _M2_z.z_da_dr _M3_z.z_da_dr _M4_z.z_da_dr	Verifying deviation from end point after auto calibration Values: measured value (z.z), absolute dev (da), relative dev (dr)

**12.4.16. AFDA: Purge and Autocalibration times**

Command	Response	Description
_AFDA_Km_SATK	_AFDA_s_z_y_X_w	Auto calibration times of channel m: z: Purge time y: Calibration time x: Total calibration time w: Verify time (z, y, x, w in seconds)
_AFDA_K0_SSPL	_AFDA_s_z	Purge time will be responded

**12.4.17. APAR: Request Autocalibration tolerance values**

Command	Response	Description
_APAR_Km_SATK	_APAR_s_z.z_y.y_x.x_w.w	Autocalibration tolerance value (%): z.z: Range 1 y.y: Range 2 x.x: Range 3 w.w: Range 4

**12.4.18. AKAL: Deviations from calibration**

Command	Response	Description
_AKAL_Km_	_AKAL_s_M1_z.z_y.y_x.x_w.w _M2_z.z_y.y_x.x_w.w _M3_z.z_y.y_x.x_w.w _M4_z.z_y.y_x.x_w.w	Deviation: z.z: Zero gas relative last calibration y.y: Zero gas factory calibration. x.x: Span gas relative last calibration. w.w: Span gas factory calibration

**12.4.19. ASYZ: Respond System Time**

Command	Response	Description
_ASYZ_K0_	_ASYZ_s_yymmdd_hhmmss	System time: yymmdd: year, month, day (each 2 characters wide, no spaces) hhmmss: hour, minute, second (each 2 characters wide, no spaces)

**12.4.20. AT90: Respond Lowpass filter time**

Command	Response	Description
_AT90_K0_	_AT90_s_t	Respond low pass filter time t=filter time in seconds

**12.4.21. ADAL: Diagnostic alarm limits**

Command	Response	Description
_ADAL_K0	_ADAL_s_a1.min_a1.max... _a16.min_a16.max	All alarms limits are responded
_ADAL_K0_x	_ADAL_s_x.min_x.max	Alarm limits of x

**Alarm Limits:**

1	Sample gas flow channel 1	7	Temperature
2	Sample gas flow channel 2	8	Sample concentration channel 1
3	Sample gas flow channel 3	9	Sample concentration channel 2
4	External input 1	10	Sample concentration channel 3
5	External input 2	11-13	Temperature channel 1...3
6	Barometric -Pressure	14-16	EPC voltage channel 1...3

**12.4.22. ATCP: Query TCP/IP settings**

Command	Response	Description
_ATCP_K0	_ADAL_s_zzz.zzz.zzz.zzz _yyy.yyy.yyy.yyy _xxxx	zzz: TCP/IP Address yyy: TCP/IP subnet mask xxxx: TCP/IP port

**12.4.23. AVER:** Query Software version

Command	Response	Description
_AVER_K0	_AVER_s_3MAIN_z_3USER_y_OS MSR_x	z: Main version x.xxx.b_dd.mm.yyyy y: User version x.xxx.b_dd.mm.yyyy x: OSMSR version x.xxx_dd.mm.yyyy

**12.4.24. AH2O:** Query H<sub>2</sub>O correction parameter

Command	Response	Description
_AH2O_Km	_AH2O_s_z.z_y.y_x.x	z.z: Dry – voltage of A in with no water present y.y: 1st order coefficient x.x.: 2nd order coefficient

**12.4.25. ACO2:** Query CO<sub>2</sub> correction parameter

Command	Response	Description
_ACO2_Km	_ACO2_s_z.z_y.y_x.x_w.w	z.z: Offset – voltage of A in with no CO <sub>2</sub> present y.y: Min A in – if A in is below this value no CO <sub>2</sub> correction will be done. x.x: 1st order coefficient w.w.: 2nd order coefficient

**12.4.26. AUDP:** Query UDP data streaming parameter

Command	Response	Description
_AUDP_K0	_AUDP_s_<UDPPort>_ <DataFrequency>_[<Mode>] _  <UDP_IP>]	Port: port for open the UDP connection DataFrequency: Frequency for transmit the data in Hz Mode: A: ASCII Mode UDP_IP: Alternative IP address for open the UDP connection when it should not use the IP of connected TCP/IP client

**12.5. Control commands****12.5.1. SRES: Reset**

Command	Response	Description
_SRES_K0	_SRES_s	Reset

**12.5.2. SPAU: Pause**

Command	Response	Description
_SPAU_K0	_SPAU_s	Pause mode

**12.5.3. STBY: Standby**

Command	Response	Description
_STBY_K0	_STBY_s	Standby mode for all channels
_STBY_Km	_STBY_s	Standby mode for channel m

**12.5.4. SNGA: Open valve for zero gas calibration**

Command	Response	Description
_SNGA_K0	_SNGA_s	Open all three zero gas valves
_SNGA_Km	_SNGA_s	Open valve for zero gas calibration of actual measuring range
_SNGA_Km_Mn	_SNGA_s	Open valve for zero gas calibration of range Mn

**12.5.5. SEGA: Open valve for end gas calibration**

Command	Response	Description
_SEGA_K0	_SEGA_s	Open all three end gas valves
_SEGA_Km	_SEGA_s	Open valve for end gas calibration of actual measuring range
_SEGA_Km_Mn	_SEGA_s	Open valve for end gas calibration of range Mn

**12.5.6. SSPL: Purge Analyzer with zero gas**

Command	Response	Description
_SSPL_K0	_SSPL_s	Open zero gas valve and purge all channels

**12.5.7. SATK: Start automatic calibration**

Command	Response	Description
_SATK_Km	_SATK_s	Start automatic calibration with selected range of channel m
_SATK_Km_Mn	_SATK_s	Start automatic calibration for channel m, Range n

**12.5.8. SEMB: Set measuring range**

Command	Response	Description
_SEMB_Km_Mn	_SEMB_s	Set measuring range Auto range will be disabled

**12.5.9. SARE:** Auto range on

Command	Response	Description
_SARE_K0	_SARE_s	Set auto range on for all channels
_SARE_Km	_SARE_s	Set auto range on for channel m

**12.5.10. SARA:** Auto range off

Command	Response	Description
_SARA_K0	_SARA_s	Set auto range off for all channels
_SARA_Km	_SARA_s	Set auto range off

**12.5.11. SREM:** Remote mode for AK-commands

Command	Response	Description
_SREM_K0	_SREM_s	Set device in remote mode

**12.5.12. SMAN:** Manual control to control device manually

Command	Response	Description
_SMAN_K0	_SMAN_s	Set device in manual mode

**12.5.13. SMGA:** Start measuring

Command	Response	Description
_SMGA_K0	_SMGA_s	Start measuring Open all sample valves
_SMGA_Km	_SMGA_s	Open sample valve of channel m

**12.5.14. SNKA:** Saves measured value as new offset.

Command	Response	Description
_SNKA_K0	_SNKA_s	Saves measured value of actual range for each channel as new offset if zero valve is opened
_SNKA_Km	_SNKA_s	Saves measured value of actual range as new offset if zero valve is opened

**12.5.15. SEKA:** Saves measured value as new span value

Command	Response	Description
_SEKA_K0	_SEKA_s	Saves new span values for each channel if span valve is opened
_SEKA_Km	_SEKA_s	Saves measured value of actual range as new span value if span valve is opened

**12.5.16. SUDP:** Start /Stop UDP data streaming

Command	Response	Description
_SUDP_K0_ON	_SUDP_s	Start Data streaming via the UDP channel. You need to configure the channel before with EUDP command
_SUDP_K0_OFF	_SUDP_s	Stop streaming via the UDP channel

## 12.6. Settings

**12.6.1. EKAK:** The four span gas concentration values are set

Command	Response	Description
_EKAK_Km_M1_w.w_M2_x.x_M3_y.y_M4_z.z	_EKAK_s	Set end gas values for channel m

**12.6.2. EMBE:** The four measuring range end values are set

Command	Response	Description
_EMBE_Km_M1_w.w_M2_x.x_M3_y.y_M4_z.z	_EMBE_s	Set range limits

**12.6.3. EMBU:** The upper and the lower range switchover for auto range are set

Command	Response	Description
_EMBU_Km_M1_w.w_W.W_M2_x.x_X.X_M3_y.y_Y.Y_M4_z.z_Z.Z	_EMBU_s	Set lower and upper range switchover limits

**12.6.4. EKEN:** Set new device identification and information

Command	Response	Description
_EKEN_K0_new device-name	_EKEN_s	Set new device identification Maximum length of device name is 40 characters

**NOTE:** To change device identification, you must first rename the device to “RESET”. Now a name up to 40 characters can be given.

**NOTE:** The device name must not have any blanks between characters, e.g. “CAI CLD” is not allowed. You can use underscores, e.g., “CAI\_CLD”.

**12.6.5. EGKD:** Set polynom coefficients

Command	Response	Description
_EGKD_Km_Mn_a0_a1_a2_a3_a4	_EGKD_s	Set polynomial coefficients of range Mn on channel m

**12.6.6. EFDA:** Set autocalibration and purge times

Command	Response	Description
_EFDA_Km_SATK_z_y_x_w	_EFDA_s	Set auto cal. times for channel m: z: Purge time y: Calibration time x: Total calibration time w: Verify time (z, y, x, w in seconds)
_EFDA_K0_SSPL_z	_EFDA_s	Set analyzer purge time to z seconds

**12.6.7. EPAR:** Set autocalibration tolerance values

Command	Response	Description
_EPAR_Km_SATK_z.z_y.y_x.x_w.w	_EPAR_s	Autocalibration tolerance value (%): z.z= Range 1 y.y= Range 2 x.x= Range 3 w.w= Range 4

**12.6.8. ESYZ: Set System Time**

Command	Response	Description
_ESYZ_K0_yymmdd_hhmmss	_ESYA_s	Set system time: yymmdd: year, month, day (each 2 characters wide, no spaces) hhmmss: hour, minutes, seconds (each 2 characters, no spaces)

**12.6.9. ET90: Set Lowpass Filter Time**

Command	Response	Description
_ET90_K0_t	_ET90_s	Set lowpass filter time: t= filter time in seconds

**12.6.10. EDAL: Diagnostic alarm limits**

Command	Response	Description
_EDAL_K0_a1.min_a1.mas..._a12max	_EDAL_s	Set all alarm limits
_EDAL_K0_x_x.min_xmax	_EDAL_s	Set alarm limits of x

**Alarm Limits:**

1	Flow of channel 1	7	Temperature
2	Flow of channel 2	8	Sample concentration channel 1
3	Flow of channel 3	9	Sample concentration channel 2
4	External analog in 1	10	Sample concentration channel 3
5	External analog in 2	11-13	Temperature alarm limits channel 1...3
6	Pressure	14-16	EPC voltage alarm limits channel 1...3

**12.6.11. ETCP: Set TCP/IP Parameters**

Command	Response	Description
_ETCP_K0_zzz.zzz.zzz.zzz	_EDAL_s	zzz= TCP/IP address
_yyy.yyy.yyy.yyy		yyy= TCP/IP subnet mask
_xxxx		xxxx= TCP/IP port
		All changes take effect after next power on cycle

**12.6.12. EH2O Set H<sub>2</sub>O correction parameters**

Command	Response	Description
_EH2O_Km_z.z_y.y_x.x	_EH2O_s	z.z: dry y.y: 1st order coefficient x.x: 2nd order coefficient

**12.6.13. ECO2 Set CO<sub>2</sub> correction parameters**

Command	Response	Description
_EH2O_Km_z.z_y.y_x.x	_EH2O_s	z.z: dry y.y: 1st order coefficient x.x: 2nd order coefficient

#### 12.6.14. EUDP Set UDP Data streaming parameters

Command	Response	Description
_EUDP_K0_<UDPPort>_<DataFrequeny>_[<Mode>]_[<UDP_IP>]	_EUDP_s	<p>Configure an UDP channel for data streaming of the measuring values via Ethernet UDP.</p> <p>Port: port for open the UDP connection</p> <p>DataFrequency: Frequency for transmit the data in Hz</p> <p>Mode: A: ASCII Mode (optional)</p> <p>UDP_IP: Alternative IP address for open the UDP connection when it should not use the IP of connected TCP/IP client (optional)</p>

#### 12.6.15. Format of the streaming Data via UDP:

ASCII Mode:

The measuring values will be sent with ASCII signs. The format is:

<Sequence number>\_x.x\_y.y\_z.z

The sequence number will be incremented with every data packet, which is sent.

The measuring values x.x, y.y and z.z will be sent like in AKON K0 telegram

#### 12.7. Abbreviations used

Km	: K' + channel Number ( K1...K3
Mn	: Measuring range number
M1 .. M4	: Measuring Range 1 .. 4
w.w..zz	: Numerical value
W.W...Z.Z	: Numerical value
T	: Numeric integer value
x	: Number
a0 .. a4	: Polynomial coefficients
s	: Status

**13. Appendix 1 Electrical Block Diagram**

## 14. Appendix 2- Starting With Serial Number U06081

### 1.0 INTRODUCTION

The Model 600 Series of instruments starting with Serial Number U06081 have several new Hardware and Software features.

The Hardware includes the use of a new memory system, isolation of the analog output signals and 15 relays that are used to buffer the many new digital output signals that are now available

The available digital signals consist of a SERVICE group, that can be used to externally monitor a number of conditions for preventative maintenance and diagnostics.

A second STATUS group, is provided to define the operation of the instrument such as Spanning, Zeroing, Calibrating and the current Range(1, 2, 3, 4, AUTO)

The Software includes modifications to existing functions, changes to the Measurement screen, additional Short-Cut Keys and several New Functions that are listed as follows:

#### • MEASUREMENT

Over Range	888888
Diagnostics	F3
Zero	F5
Span	F6
Standby	F7
Range Limits	F8
Span Values	F9
Outputs	F10

Note: The operator can use these Short-Cut Keys or continue to use existing procedures.

#### • NEW FUNCTIONS

Auto Startup	F5, F7, F7
ALARMS	F5, F7,(Use F6 to toggle ON/OFF)
Offsets& Gains	F4, F3, F5
D/A Calibration	F5, F7, F8
Save Data Archiving Time	F5, F7, F1, F5 (Use ENTER to change recording time)
User Digital Outputs	F5, F9

#### • Modifications

Saved/Not good	F4, F2, F1 or F2 (To flow Zero or Span Gas)
Re-Set Calibration Values	F4, F5


**2.0 OPERATION OF MEASUREMENT KEYS:**

Note: USE the F1 & F2 Keys to view the complete list of menu items.

**2.1 Diagnostics** Use **F3** to toggle between MEASUREMENT an DIAGNOSTIC

**2.2 Zero:** Select the required channel and range then press **F5**.

Note: For instruments with an internal Zero Solenoid select Calibration By Valves. (**F5, F2, F4**)



Zero Gas will be enabled and the observed results can be used to evaluate instrument performance

Press **F3** to toggle to the Diagnostic screen for additional information

Press **F1** to save the value and complete a ZERO calibration for this channel and range

Press **MAIN, F1** to return to the MEASUREMENT screen to select other channels and ranges and repeat the process or **F2** to return to the Manual Calibration screen

**2.3 Span:** Select the required Channel and Range then press **F6**.

Note: For instruments with an internal Span Solenoid select Calibration By Valves (**F5, F2,F4**).



Span Gas for Channel1 and Range 1 will be enabled and the observed results can be used to evaluate instrument performance.

Press **F3** to toggle to the Diagnostic screen for additional information

Press **F1** to save the new value and complete the SPAN calibration for this Channel and Range.

Note that the span gas value used for this channel and range is highlighted and can be changed if necessary. Use the Enter key and the numeric keys as required.

Press **MAIN**, **F1** to return to the MEASUREMENT screen to select other Channels and Ranges and repeat the process or **F2** to return to the Manual Calibration screen.


#### 2.4 Range Limits: (F8);



The Channels and Ranges are factory defined and application specific. Consult California Analytical if any changes are required.

Use the  $\leftarrow \rightarrow$  keys to observe the other Channels.

#### 2.5 Span Values: F9



Use the  $\leftarrow \rightarrow$  keys to select the desired Channel and the  $\uparrow$  to select the Range

Note that the span gas value used for this channel and range is highlighted and can be changed if necessary.

Use the **Enter** key and the **NUMERIC** keys as required.

**2.6 Outputs: F10**

assignment	
Output	Signal
1	Channel1
2	Channel12
3	Channel13
4	Pressure

Thu Jun 28 09:25:18 2007

Use this screen to define the signals and their location that will be monitored by a remote reordering device.

### 3.0 NEW FUNCTIONS

#### 3.1 Auto Start Up: (F5, F7, F7)



All key analyzer parameters are stored in a secure memory location and retained when power is removed

In the event of an unexpected power failure it may be desirable to change some parameters until an operator can resume control.

This screen may be used to establish several desirable special instrument start-up parameters that define how the analyzer recovers from loss of AC power

When enabled this screen will define the following:

**Wait:** The time delay in minutes before proceeding

If **zero** is selected the instrument will not start until warnings are cleared

**Access Level:** The final access level

**Remote/Manual:** The final operating MODE


**Calibrations:** The number of attempts to complete a successful calibration as required in the operator defined Deviation Tables.

If calibration is not successful the instrument will continue reporting results using the last completed calibration.

The analyzer can be configured use the previous calibration by selecting zero Calibrations.

**Starting Range:** When all defined actions are completed the analyzer will return to the Measurement Screen and to the range specified.

### 3.2 Alarms On/Off: (F5, F7)




Use **F6** to toggle ON/OFF

The instrument has an extensive library of error messages that will aid in the identification of various anomalous events and are displayed at the bottom of the screen. These messages will assist in Diagnostics and indicating the need for preventative maintenance.

This screen provides an option to disable these messages during initial start-up or as may be desired for a particular application.

### 3.3 Offset & Gain: (F4, F3, F5)



This screen can be used to provide an additional means to display calibration deviations.

Use ← → keys to change channel.

The OFFSET is the value stored during zero calibration.

The GAIN is the value stored during span gas calibration using the operator defined calibration gas.

An increasing or decreasing change to the OFFSET or GAIN when used in conjunction with "Deviations" will provide insight to both short and long term changes to system performance.

### 3.4 Calibrate the Digital Outputs

F5, F7, F8 D/A Calibration

F5, F7, F3 Output Assignment

- **Overview**

The 600 CLD Series of instruments are designed to provide analog outputs that can be configured as 0-10v, 4-20 ma, or 0-20 ma.

With this version the outputs can also be configured to include an additional 1.0 volt and 5.0 volt output and a **calibration** capability.

The outputs can be calibrated to exactly match the results obtained on a PLC, Recorder, Data Logger or other remote recording device that may be connected to the analyzer.

The operator will first select the OUTPUT ASSIGNMENT screen and choose the output that is to be calibrated. All outputs of interest may be selected. When calibration is completed, the operator will return the outputs to their original assignment

The D-A CALIBRATION screen will be then be used to complete the calibration procedure.

This screen provides a section that is used to record the zero signal corrections (zero offset) and a second area to record the 100% signal corrections (Gain) for each of the four output signals that may be defined to develop a voltage or current signal..

Since this is a Digital to Analog conversion, the calibration will require the completion of a simple “trial and error” procedure. The operator will observe the results of a “zero or full scale (Gain) signal generated by the analyzer to the remote recording device and select a correction factor. The operator will save this value and then observe the results on attached the remote recording device.

The process of selection and saving for “zero” and “span” will be repeated until a satisfactory calibration is achieved. For 0-1V, 0-5V, 0-10V and a 0-20 ma outputs the Offset and Gain values are independent and do not interact... With the 4-20 ma output, the “Offset (zero)” and “Gain (span)” values interact and may require a few more trials.


The following is a table of typical values:

OUTPUT	OFFSET	GAIN
--------	--------	------

0-20 ma	0.000	0.927
4-20 ma	1.820	0.740
0-1 V	1.300	0.820
0-5 V	1.100	0.820
0-10 V	1.050	0.820

**Procedure**

**3.4.1 From the Main Menu press F5,F7,F3,to obtain following screen:**



**F5, F7, F3**

**3.4.2 Use the ↑ to highlight the outputs that require calibration.**

Note: In the above example only Output 1 will be calibrated  
Record the name of these signals, they will be restored.


**3.4.3 Press enter to provide access to all the menu of signals that are available. (Real Time, THC, CH<sub>4</sub>, Calibration, Sample pressure, etc.)**

**3.4.4 Select Calibration and press ENTER to complete the selection**

Note: Any or all of the four outputs can be selected for calibration  
This screen will not be used again until calibration has been completed.

**3.4.5 Press BACK to return to the SYSTEM SETUP screen (F5, F7 from the main menu)**

**3.4.6 Press F8 to obtain the following screen:  
Press Enter to activate highlighted area**



**F5, F7, F8**

- 3.4.7 Use the **↓** to select the desired output press **ENTER**
- 3.4.8 Press **F1** to select a ZERO signal and observe the results on the remote device
- 3.4.9 Change the offset value press **BACK** to save the new value.
- 3.4.10 Press **F8** to return to the D-A Calibration screen and note the results on the remote device.
- 3.4.11 Repeat steps 8.0 thru 10.0 until a satisfactory ZERO calibration is achieved.
- 3.4.12 Complete steps 8.0 thru 10.0 for each of the remaining outputs that require calibration.
- 3.4.13 Press **F5** to produce a full scale (100%) signal
- 3.4.14 Use the arrow keys to position the cursor at the require GAIN value
- 3.4.15 Observe the results on the remote device and make a correction to the GAIN value for the output of interest. Press **BACK** to save this new value
- 3.4.16 Press **F8** to return to the D-A calibration screen
- 3.4.17 Observe the results on the remote device and repeat steps change the GAIN value by repeating steps 3.4.14 thru 3.4.16 as needed for each output.
- 3.4.18 Return to the OUTPUT Assignment screen **F5, F7, F3** from the main menu and change the output signals from CALIBRATE to their original values as defined in step 3.2.2.

### 3.5 Save Data Archiving Time (F5, F7, F1, F5)



Use ENTER to change recording time

### 3.6 User Digital Outputs

#### Overview

The 600 Series of instruments have 15 solid state, optically coupled, isolated relays that can be programmed by the operator to indicate the status of numerous signals.

The following is a list of digital signals that are available.

- SERVICE

Signal	Displayed Message	MIN	MAX	Definition
FI1	Flow 1 ,Check	0.5	3.5	Flow 1
FI2	Flow 2 ,Check	0.5	3.5	Flow 2
FI3	Flow 3 ,Check	0.5	3.5	Flow 3
E1	Extern Analog 1 ,Check			External Analog 1
E2	Extern Analog 2 ,Check			External Analog 2
P	Pressure ,Check	10	16	Pressure
T	Temperature ,Check	20	50	Temperature
1NC	Channel 1 is not calibrated			Temperature Ch.1
2NC	Channel 2 is not calibrated			Temperature Ch.2
3NC	Channel 3 is not calibrated			Temperature Ch.3
1LoC	Ch1: Low conc. Warning	20	60	EPC Ch.1
2LoC	Ch2: Low conc. Warning	20	60	EPC Ch.2
3LoC	Ch3: Low conc. Warning	20	60	EPC Ch.3
1HiC	Ch1: High conc. Warning	2	8	
2HiC	Ch2: High conc. Warning	2	8	
3HiC	Ch3: High conc. Warning	2	8	
1DT	Ch1: Temperature !			
2DT	Ch2: Temperature !			
3DT	Ch3: Temperature !			
1EV	Ch1: EPC ,Check			
2EV	Ch2: EPC ,Check			
3EV	Ch3: EPC ,Check			
1OR	Ch1: Range overflow			
2OR	Ch2: Range overflow			
3OR	Ch3: Range overflow			
1AU	Ch1: ADC Under Range			
2AU	Ch2: ADC Under Range			
3AU	Ch3: ADC Under Range			
1AO	Ch1: ADC Over Range			
2AO	Ch2: ADC Over Range			
3AO	Ch3: ADC Over Range			
0ff				

- STATUS


1R2	1 Range 2
1R3	1 Range 3
1R4	1 Range 4
1C	1 In Calibrate
1Z	1 In Zero
1S	1 In Span
1Sa	1 In Sample
2AR	2 Auto Range
2R1	2 Range 1
2R2	2 Range 2
2R3	2 Range 3
2R4	2 Range 4
2C	2 In Calibrate
2Z	2 In Zero
2S	2 In Span
2Sa	2 In Sample
3AR	3 Auto Range
3R1	3 Range 1
3R2	3 Range 2
3R3	3 Range 3
3R4	3 Range 4
3C	3 In Calibrate
3Z	3 In Zero
3S	3 In Span
3Sa	3 In Sample

- STATUS

1R2	1 Range 2
1R3	1 Range 3
1R4	1 Range 4
1C	1 In Calibrate
1Z	1 In Zero
1S	1 In Span
1Sa	1 In Sample
2AR	2 Auto Range
2R1	2 Range 1
2R2	2 Range 2
2R3	2 Range 3
2R4	2 Range 4
2C	2 In Calibrate
2Z	2 In Zero
2S	2 In Span
2Sa	2 In Sample
3AR	3 Auto Range
3R1	3 Range 1
3R2	3 Range 2
3R3	3 Range 3
3R4	3 Range 4
3C	3 In Calibrate
3Z	3 In Zero
3S	3 In Span
3Sa	3 In Sample

---

Typical Relay



These contacts(3, 4) will drive continuously up to 500 MA using a customer voltage supply that does not exceed 60 VDC.

- Operation

Use F5, F9 to select the first eight outputs

Use the ↑ to select the desired output

Press ENTER and use ↑ to select desired item

NOTE the 600 FID can provide 35 digital output signals




Press **F1** to observe the remaining seven outputs  
Program as desired per the above



#### 4.0 Changes to existing functions


##### 4.1 Saved or NOT GOOD

During Manual Calibration the following screens will be displayed to indicate the instruments response to the value of the zero or span gas using the amount that the operator defined in the deviation table.



The above is shown using Zero/SPAN Gas

From Measurement use: **F5 or F6**  
From Main Menu use: **F4, F2, F1 or F2**




**F5, F2, F3**

Note: This screen is used by the operator to define the maximum acceptable limits of the Zero and Span gas for both Manual and Automatic Calibrating.


#### 4.2 Reset Calibration Values

When the re-set calibrations value function is used all recorded deviations

will be set to zero




F4, F5



F4 F3, F4

(Used to observe Auto Cal Results)



F4, F3, F2

(Used to observe Manual Cal results)

The above are the new deviations after the operator elects to re-set the calibration values



## TABLE OF CONTENTS

### ADDENDUM

#### STARTING WITH SERIAL #U06081

<u>Section</u>	<u>Title</u>	<u>Page</u>
1.0.	<b>Introduction -----</b>	
2.0.	<b>Operation of Measurement Keys -----</b>	
	2.1 Over Range-----	
	2.2. Diagnostics -----	
	2.3. Zero-----	
	2.4. Span-----	
	2.5. Range Limits-----	
	2.6. Span Values-----	
	2.7. Outputs -----	
3.0	<b>New Functions -----</b>	
	3.1. Auto Start Up -----	
	3.2. Alarms On/Off -----	
	3.3. Offset & Gain -----	
	3.4. Calibrate The Analog Outputs-----	
	3.5. Save Data Archiving Time-----	
	TABLE A-----	
	3.6. User Digital Outputs -----	
	TABLE B-----	
	TABLE C-----	
4.0.	<b>Changes To Existing Functions -----</b>	
	4.1. Saved or Outside Limits -----	
	4.2. Calibration Deviations-----	
	4.3. Flow Zero or Span Gas-----	
	4.4. Reset Calibration Values-----	
	TABLE D.....	

## **13.5 Starting With Serial Number U06081**

### **1.0 INTRODUCTION**

The Model 600 NDIR Series of instruments starting with Serial Number U06081 have several new Hardware and Software features.

The Hardware includes the use of a new memory system, isolation of the analog output signals and 15 relays that are used to buffer the many new digital output signals that are now available. **SEE TABLE D**

The available digital signals consist of a SERVICE group, which can be used to externally monitor a number of conditions for preventative maintenance and diagnostics.

A second STATUS group, is provided to define the operation of the instrument such as Spanning, Zeroing, Calibrating and the current Range (1, 2, 3, 4, AUTO)

Many of the various signals are duplicated because an NDIR instrument can consist of up to three different channels.

The Software includes modifications to existing functions, changes to the Measurement screen, additional Short-Cut Keys and several New Functions that are listed as follows:

- MEASUREMENT**

Over Range	<b>888888</b>
Diagnostics	<b>F3</b>
Zero	<b>F5</b>
Span	<b>F6</b>
Range Limits	<b>F8</b>
Span Values	<b>F9</b>
Outputs	<b>F10</b>

Note: The operator can use these Short-Cut Keys or continue to use existing procedures.

- NEW FUNCTIONS**

Auto Startup	<b>F5, F7, F7</b>
ALARMS	<b>F5, F7, (Use F6 to toggle ON/OFF)</b>
Offsets& Gains	<b>F4, F3, F5</b>
D/A Calibration	<b>F5, F7, F8</b>
Save Data Archiving Time	<b>F5, F7, F1, F5 (ENTER to change record time)</b>
User Digital Outputs	<b>F5, F9</b>

- Modifications**

Saved/Not good	<b>F4, F2, F1 or F2</b> (To flow Zero or Span Gas)
Re-Set Calibration Values	<b>F4, F5</b>

## 2.0 OPERATION OF MEASUREMENT KEYS

Note: USE the F1 & F2 Keys to view the complete list of menu items, from the MEASUREMENT screen


### 2.1 Over Range 888888

In the MEASUREMENT mode only, any value that exceeds the “range” by more than 10% will be displayed as 888888.

### 2.2 Diagnostics Use F3 to toggle between MEASUREMENT and DIAGNOSTIC

### 2.3 Zero: From the MEASUREMENT screen, select the required channel and range then press F5.

Note: For instruments with Zero Solenoid(s) select Calibration by Valves. (**Main, F5, F2, F4**)



2 Versions

Zero Gas will be enabled and the observed results can be used to evaluate instrument performance


Press **F3** to toggle to the Diagnostic screen for additional information

Press **F1** to save the value and complete a ZERO calibration for this channel and range.

Press **MAIN, F1** to return to the MEASUREMENT screen to select other channels and ranges and repeat the process or press **F2** to return to the Manual Calibration screen

**2.4 Span:** From the MEASUREMENT screen, select the required Channel and Range then press **F6**.

Note: For instruments with Span Solenoid(s) select Calibration by Valves (**Main**, **F5**, **F2**, and **F4**).



### 2 Versions

Span Gas for Channel1 and Range 1 will be enabled and the observed results can be used to evaluate instrument performance.


Press **F3** to toggle to the Diagnostic screen for additional information

Press **F1** to save the new value and complete the SPAN calibration for this Channel and Range.

Note: The span gas value used for this channel and range is highlighted and can be changed if necessary. Use the Enter key and the numeric keys as required.

Press **MAIN**, **F1** to return to the MEASUREMENT screen to select other Channels and Ranges and repeat the process or **F2** to return to the Manual Calibration screen.

## 2.5 Range Limits: (F8) From the MEASUREMENT screen;




### 2 Versions.

The Channels and Ranges are factory defined and application specific.  
Consult California Analytical if any changes are required.


Use the ← → keys to observe the other Channels.

## 2.6 Span Values: F9 From the MEASUREMENT screen



Use the ← → keys to select the desired Channel and the ↑ to select the Range  
Note: The span gas value used for this channel and range is highlighted and can be changed if necessary.

Use the **Enter** key and the **NUMERIC** keys as required.

**2.7 Outputs: F10 From the MEASUREMENT screen**

Use this screen to define the signals and their location that will be monitored by a remote reordering device.

Use the **↑** to select the desired Output. Press Enter to select  
Use the **↑** to select the desired Signal. Press Enter to select

Use this screen to define the signals and their location that will be monitored by a remote reordering device.

**SEE TABLE D**

### 3.0 NEW FUNCTIONS

#### 3.1 Auto Start Up: (Main, F5, F7, F7)



All key analyzer parameters are stored in a secure memory location and retained when power is removed

In the event of an unexpected power failure it may be desirable to change some parameters until an operator can resume control.

This screen may be used to establish several desirable special instrument start-up parameters that define how the analyzer recovers from loss of AC power

When enabled this screen will define the following:

**Wait:** The time delay in minutes before proceeding. If **zero** is selected the instrument will not start until all temperature warnings are cleared

**Calibrations:** The number of attempts to complete a successful calibration as required in the operator defined Deviation Tables.

If calibration is not successful the instrument will continue reporting results using the last completed calibration.

The analyzer can be configured use the previous calibration by selecting zero Calibrations.

**Starting Range:** When all defined actions are completed the analyzer will return to the Measurement Screen and to the range specified.

**Access Level:** The user level at Start Up


**Remote/Manual:** Put the analyzer in either "Remote" or "Manual" at Start Up

### 3.2 Alarms On/Off: (Main, F5, F7)

All key analyzer parameters are stored in a secure memory location and retained when power is removed. In the event of an unexpected power failure it may be desirable to change some parameters until an operator can resume control.

This screen may be used to establish several desirable special instrument start-up parameters that define how the analyzer recovers from loss of AC power.

When enabled this screen will define the following...




Use **F6** to toggle ON/OFF

The instrument has an extensive library of error messages that will aid in the identification of various anomalous events and are displayed at the bottom of the screen. These messages will assist in Diagnostics and indicating the need for preventative maintenance

This screen provides an option to disable these messages during initial start- up or as may be desired for a particular application.

### 3.3 Offset & Gain (Main, F4, F3, F5)



This screen can be used to provide an additional means to display calibration deviations.

Use ← → keys to change channel.

The OFFSET is the value stored during zero calibration.

The GAIN is the value stored during span gas calibration using the operator defined calibration gas.

An increasing or decreasing change to the OFFSET or GAIN when used in conjunction with "Deviations" will provide insight to both short and long term changes to system performance

### 3.4 Calibrate The Analog Outputs: Main, F5, F7, F8 D/A Calibration Main, F5, F7, F3 Output Assignment

- Overview


The 600 NDIR Series of instruments are designed to provide analog outputs that can be configured as 0-1 VDC, 0-5 VDC, 0-10 VDC, 4-20 mA, or 0-20 mA.

With this version the outputs can also be configured to include an additional 1.0 volt and 5.0 volt output and a **calibration** capability.

The outputs can be calibrated to exactly match the results obtained on a PLC, Recorder, Data Logger or other remote recording device that may be connected to the analyzer.

The operator will first select the **OUTPUT ASSIGNMENT (Main, F5, F7, F3)** screen and choose the output that is to be calibrated. By selecting "calibrate" as the output all outputs of interest may be selected. When calibration is completed, the operator will return the outputs to their original assignment

The D-A CALIBRATION screen will then be used to complete the calibration procedure.



This screen provides a section that is used to record the zero signal corrections (zero offset) and a second area to record the 100% signal corrections (Gain) for each of the four output signals that may be defined to develop a voltage or current signal..

Since this is a Digital to Analog conversion, the calibration will require the completion of a simple “trial and error” procedure. The operator will observe the results of a “zero” or “span” (Gain) signal generated by the analyzer to the remote recording device and select a correction factor. The operator will save this value and then observe the results on attached the remote recording device.


The process of selection and saving for “zero” and “span” will be repeated until a satisfactory calibration is achieved. For 0-1VDC, 0-5VDC, 0-10VDC and a 0-20 mA outputs the Offset and Gain values are independent and do not interact. With the 4-20 ma output, the Offset “zero” and Gain “span” values interact and may require a few more trials.

The following is a table of typical values:

<b>OUTPUT</b>	<b>OFFSET</b>	<b>GAIN</b>
0-20 ma	0.000	0.927
4-20 ma	1.820	0.740
0-1 V	1.300	0.820
0-5 V	1.100	0.820
0-10 V	1.050	0.820

- **Procedure**

**3.4.1** From the Main Menu press **F5, F7, F3**, to obtain following screen:




**Main, F5, F7, F3**

**3.4.2** Use the **↓** to highlight the outputs that require calibration.

**3.4.3** Press enter to provide access to all the menu of signals that are available.  
(Real Time). Channel 1, Channel 2, Calibration, Sample pressure, etc.)


**3.4.4** Select Calibration and press **ENTER** to complete the selection



Note: Any or all of the four outputs can be selected for calibration  
This screen will not be used again until calibration has been completed.

**3.4.5** Press **BACK** to return to the SYSTEM SETUP screen (**F5, F7** from the main menu)

**3.4.6** Press **F8** to obtain the following screen:



**MAIN, F5, F7, F8**

**3.4.7** Use the **↑** to select the desired output press **ENTER**

**3.4.8** Press **F1** to select a ZERO signal and observe the results on the remote device

**3.4.9** Change the offset value press **BACK** to save the new value.

**3.4.10** Press **F8** to return to the D-A Calibration screen and note the results on the remote device.

**3.4.11** Repeat steps **3.4.8** thru **3.4.10** until a satisfactory ZERO calibration is achieved.

**3.4.12** Complete steps **3.4.8** thru **3.4.10** for each of the remaining outputs that require calibration.

**3.4.13** Press **F5** to produce a full scale (100%) signal

**3.4.14** Use the arrow keys to position the cursor at the required GAIN value

**3.4.15** Observe the results on the remote device and make a correction to the GAIN value for the output of interest. Press **BACK** to save this new value


**3.4.16** Press **F8** to return to the D-A calibration screen

**3.4.17** Observe the results on the remote device and repeat steps change the GAIN value by repeating steps **3.4.14** thru **3.4.16** as needed for each output.

**3.4.18** Return to the OUTPUT Assignment screen **F5, F7, F3** from the main menu and change the output signals from CALIBRATE to their original values as defined in step **3.4.1**.

### **3.5 Save Data Archiving Time (F5, F7, F1, F5)**

Archive Time is the Time in seconds between each set of data points. If “zero” no data is stored in the SEC data files. The SEC data files are in .CSV format for direct import into Excel. CAI can provide the tools necessary to download these files.



Use ENTER to change recording time

**SEE TABLE A**

**TABLE A**

**600 SERIES NDIR DATA ARCHIVE FILES**

Time,  
Date,  
Month,  
Year,  
Error Index,  
TimeStamp,  
Pressure,  
Temperature,  
  
Name,  
Concentration,  
Detector Volts,  
Range,  
Auto / Manual,  
Span Value,  
Offset,  
Gain,  
Sample Pressure,  
Sample Flow,  
Sample EPC Volts,  
Detector Temperature,  
Meas Mode,  
Local / Remote,  
  
**If 2 Channel this data is added**  
Name,  
Concentration,  
Detector Volts,  
Range,  
Auto / Manual,  
Span Value,  
Offset,  
Gain,  
Sample Pressure,  
Sample Flow,  
Sample EPC Volts,  
Detector Temperature,  
Meas Mode,  
Local / Remote,

**TABLE A (cont)**

**600 SERIES NDIR DATA ARCHIVE FILES**

If 3 channel this data is added

Name,  
Concentration,  
Detector Volts,  
Range,  
Auto / Manual,  
Span Value,  
Offset,  
Gain,  
Sample Pressure,  
Sample Flow,  
Sample EPC Volts,  
Detector Temperature,  
Meas Mode,  
Local / Remote,

### 3.6 User Digital Outputs

- **Overview**

The 600 NDIR Series of instruments have 15 solid state, optically coupled, isolated relays that can be programmed by the operator to indicate the status of numerous signals.

The available digital signals consist of a SERVICE Group that can be used to externally monitor a number of conditions to aid in preventative maintenance and diagnostics. **SEE TABLE B & D**

A second STATUS group is provided and is used to define the operation of the instrument such as Spanning, Zeroing, Calibrating, and the current Range (1, 2, 3, 4 AUTO) etc. **SEE TABLE C & D**

The individual output signals can be operator selected and set to a **HOLD** or **CLEAR** mode.

In the **HOLD** mode an activated signal is retained until the operator returns to the **Digital Output Screen** and selects the appropriate output signal and performs a manual CLEAR. After performing a Clear Operation, the operator must press **F2** again to put the outputs back onto **HOLD** mode.

User DO I	K2: STBY	[ ]
1 Off	*****	
2 Off	*Set Unused*	
3 Off	*Channels Off*	
4 Off	*****	
5 Off	*****	
6 Off	*****	
7 Off	*****	
F2 Hold / Clear	Hold	
F1 8 to 15 DO's		
MAIN/BACK to SAVE		
		[ ]

In the **Clear** Mode the signal will automatically change when the microprocessor detects that the noted condition no longer exists.

User DO II	K2: SMAN	[ ]
8 Off	*****	
9 Off	*Set Unused*	
10 Off	*Channels Off*	
11 Off	*****	
12 Off	*****	
13 Off	*****	
14 Off	*****	
15 Off	*****	
F1 1 to 7 DO's		
MAIN/BACK to SAVE		
Check Ch1: Temp.		[ ]

The operator can select from the following the desired **SERVICE** or **STATUS** groups that are to be digitally monitored.


## **TABLE B**

<b>Index</b>	<b>Service Group</b>	<b>User DO Screen Display</b>
0		
1	Flow 1 Failure	F1
2	Flow 2 Failure	F2
3	Flow 3 Failure	F3
4	Extern Analog 1 Failure	E1
5	Extern Analog 2 Failure	E2
6	Pressure Failure	P
7	Temperature Failure	T
8	Channel 1 is not calibrated	1NC
9	Channel 2 is not calibrated	2NC
10	Channel 3 is not calibrated	3NC
11	Ch1: Low conc. Warning	1LoC
12	Ch2: Low conc. Warning	2LoC
13	Ch3: Low conc. Warning	3LoC
14	Ch1: High conc. Warning	1HiC
15	Ch2: High conc. Warning	2HiC
16	Ch3: High conc. Warning	3HiC
17	Ch1: Temperature!	1DT
18	Ch2: Temperature!	2DT
19	Ch3: Temperature!	3DT
20	Ch1: EPC failure	1EV
21	Ch2: EPC failure	2EV
22	Ch3: EPC failure	3EV
23	Ch1: Range overflow	1OR
24	Ch2: Range overflow	2OR
25	Ch3: Range overflow	3OR
26	Ch1: ADC Range Overflow	1AU
27	Ch2: ADC Range Overflow	2AU
28	Ch3: ADC Range Overflow	3AU
29	Ch1: ADC Range Underflow	1AO
30	Ch2: ADC Range Underflow	2AO
31	Ch3: ADC Range Underflow	3AO
32	dummy text for RTC	Off
33	General Alarm	G
35	1 Cal Alarm	1CA
36	2 Cal Alarm	2CA
37	3 Cal Alarm	3CA
34	In Remote	R

**TABLE C**

<b>Index</b>	<b><u>STATUS GROUP</u></b>	<b>User DO Screen Display</b>
38	1 AutoRange	1AR
39	1 Range 1	1R1
40	1 Range 2	1R2
41	1 Range 3	1R3
42	1 Range 4	1R4
43	1 In Calibrate	1C
44	1 In Zero	1Z
45	1 In Span	1S
46	1 In Sample	1Sa
47	2 Auto Range	2AR
48	2 Range 1	2R1
49	2 Range 2	2R2
50	2 Range 3	2R3
51	2 Range 4	2R4
52	2 In Calibrate	2C
53	2 In Zero	2Z
54	2 In Span	2S
55	2 In Sample	2Sa
56	3 Auto Range	3AR
57	3 Range 1	3R1
58	3 Range 2	3R2
59	3 Range 3	3R3
60	3 Range 4	3R4
61	3 In Calibrate	3C
62	3 In Zero	3Z
63	3 In Span	3S
64	3 In Sample	3Sa

Typical Relay



These contacts (3, 4) will drive continuously up to 500 MA using a customer voltage supply that does not exceed 60 VDC.

- **Operation**

Use (**Main, F5, F9**) to select the first eight outputs

Use the **↑** to select the desired output

Press **ENTER** and use **↑** to select desired item

Press **ENTER** to save selection

NOTE: The 600 NDIR has 14 user selectable isolated digital outputs from the list of 64 in **TABLE B & C**

User DO I	K2: STBY	[ ]
1 Off		
2 Off	*****	
3 Off	*Set Unused*	
4 Off	*Channels Off*	
5 Off	*****	
6 Off	*****	
7 Off		
F2 Hold / Clear	Hold	
F1 8 to 15 DO's		
MAIN/ BACK	to SAVE	


Press **F1** to observe the remaining seven outputs  
Program as desired per the above

User DO II	K2: SMAN	[ ]
8 Off		
9 Off		
10 Off	*****	
11 Off	*Set Unused*	
12 Off	*Channels Off*	
13 Off	*****	
14 Off	*****	
15 Off		
F1 1 to 7 DO's		
MAIN/ BACK	to SAVE	
Check Ch1: Temp.		[ ]

## 4.0 CHANGES TO EXISTING FUNCTIONS

### 4.1 Saved or Outside Limits

During Manual Calibration the following screens will be displayed to indicate the instruments response to the value of the zero or span gas using the amount that the operator defined in the deviation table.



The Above is shown using Zero and Span Cal  
From Measurement use: **F5 "zero"** or **F6 "span"**  
From Main Menu use: **F4, F2, F1 "zero"** or **F2 "span"**

**4.2 Calibration Deviations.****MAIN, F5, F2, F2 Deviations, F3 Measuring Deviations.**

<p>Setup</p> <p>Deviations[%]</p> <table border="1" style="margin-left: 20px; margin-top: 10px;"> <tr><td>absolute</td><td>relative</td></tr> <tr><td>Range1: 10.00</td><td>10.00</td></tr> <tr><td>Range2: 10.00</td><td>10.00</td></tr> <tr><td>Range3: 10.00</td><td>10.00</td></tr> <tr><td>Range4: 10.00</td><td>10.00</td></tr> </table> <p>Enter deviations for channel : 1</p> <p>CAI Tel. (800) 959 0959</p>	absolute	relative	Range1: 10.00	10.00	Range2: 10.00	10.00	Range3: 10.00	10.00	Range4: 10.00	10.00	<p>Setup K1: SMGA</p> <p>Measuring deviation[%]</p> <table border="1" style="margin-left: 20px; margin-top: 10px;"> <tr><td>Range1: 1.00</td></tr> <tr><td>Range2: 1.00</td></tr> <tr><td>Range3: 1.00</td></tr> <tr><td>Range4: 1.00</td></tr> </table> <p>Enter deviations for channel : 1</p> <p>Sun Jan 01 00:11:48 2003</p>	Range1: 1.00	Range2: 1.00	Range3: 1.00	Range4: 1.00
absolute	relative														
Range1: 10.00	10.00														
Range2: 10.00	10.00														
Range3: 10.00	10.00														
Range4: 10.00	10.00														
Range1: 1.00															
Range2: 1.00															
Range3: 1.00															
Range4: 1.00															

Note: These screens are used by the operator to define the maximum acceptable limits of the Zero and Span gas for both Manual and Automatic Calibrating.

**4.3 Flow Zero or Span****Some analyzers have the above and the ability to flow Zero and Span Gas.**


<p>Manual calib K1: SNGA</p> <table border="1" style="margin-left: 20px; margin-top: 10px;"> <tr><td>F1 Cal zero</td><td>0 . 0 2</td></tr> <tr><td>F2 Cal span</td><td>ZERO</td></tr> <tr><td>F3 Toggle zero</td><td></td></tr> <tr><td>F4 Toggle span</td><td></td></tr> <tr><td>F5 Measurements</td><td></td></tr> </table> <p>F6 Range Up ** F7 Range Down</p> <p>[CO2] CO R1-16.00 ARI-100.0 R1-25.00</p> <p>Wed May 14 16:45:58 2008</p>	F1 Cal zero	0 . 0 2	F2 Cal span	ZERO	F3 Toggle zero		F4 Toggle span		F5 Measurements		<p>Manual calib K1: SARA</p> <table border="1" style="margin-left: 20px; margin-top: 10px;"> <tr><td>F1 Cal zero</td><td>0 . 0 1</td></tr> <tr><td>F2 Cal span</td><td>SPAN</td></tr> <tr><td>F3 Toggle zero</td><td></td></tr> <tr><td>F4 Toggle span</td><td></td></tr> <tr><td>F5 Measurements</td><td></td></tr> </table> <p>F6 Range Up ** F7 Range Down</p> <p>[CO2] CO R1-16.00 ARI-100.0 R1-25.00</p> <p>Wed May 14 16:47:28 2008</p>	F1 Cal zero	0 . 0 1	F2 Cal span	SPAN	F3 Toggle zero		F4 Toggle span		F5 Measurements	
F1 Cal zero	0 . 0 2																				
F2 Cal span	ZERO																				
F3 Toggle zero																					
F4 Toggle span																					
F5 Measurements																					
F1 Cal zero	0 . 0 1																				
F2 Cal span	SPAN																				
F3 Toggle zero																					
F4 Toggle span																					
F5 Measurements																					

The above is shown using Zero/SPAN Gas


From Cal Screen use: **F2 or Main or Back**  
 From Main Menu use: **F4, F2**

#### 4.4 Reset Calibration Values


When the re-set calibrations value function is used all recorded deviations will be set to zero



**MAIN, F4, F5**



**Main, F4 F3, F4**  
(Used to observe Auto Cal Results)



**Main, F4, F3, F2**  
(Used to observe Manual Cal results)

The above are the new deviations after the operator elects to re-set the calibration values

**TABLE D**  
**600 NDIR SERIES I/O CHART**

**28 PIN MAIN CONNECTOR ASSIGNMENTS****AO = Analog Output, OC= Open Collector, SV = Solenoid Valve TTL = Transistor Logic**

OPTO I/O	Signal Type	601 Analog	602 Analog	603 Analog	<u>Levels</u>
ALG 1		pin #	pin #	pin #	
COM 0 1 2 3	A Output	1 GND (Isolated analog)	1 GND (Isolated analog)	1 GND (Isolated analog)	<u>Isolated AI</u> 1v,5v,10v,mA
	A Output	2 User Defined AO-1	2 User Defined AO-1	2 User Defined AO-1	
	A Output	3 User Defined AO-2	3 User Defined AO-2	3 User Defined AO-2	
	A Output	4 User Defined AO-3	4 User Defined AO-3	4 User Defined AO-3	
	A Output	5 User Defined AO-4	5 User Defined AO-4	5 User Defined AO-4	
DIG 1		Digital	Digital	Digital	
COM 0 1 2 3 4	D Output	6 GND (Digital)	6 GND (Digital)	6 GND (Digital)	<u>TTL-low true</u>
	D Output	7 Sense Auto Range	7 Sense Auto Range	7 Sense Auto Range	
	D Output	8 Sense Range 1	8 Sense Range 1	8 Sense Range 1	
	D Output	9 Sense Range 2	9 Sense Range 2	9 Sense Range 2	
	D Output	10 Sense Range 3	10 Sense Range 3	10 Sense Range 3	
5 6 7 8 9	D Input	12 Set Auto Range	12 Set Auto Range	12 Set Auto Range	
	D Input	13 Control Range 1	13 Control Range 1	13 Control Range 1	
	D Input	14 Control Range 2	14 Control Range 2	14 Control Range 2	
	D Input	15 Control Range 3	15 Control Range 3	15 Control Range 3	
	D Input	16 Control Range 4	16 Control Range 4	16 Control Range 4	
10 11 12 13 14 15	D Input	17 Auto Cal	17 Auto Cal	17 Auto Cal	
	D Input	18 Calibrate	18 Calibrate	18 Calibrate	
	D Input	19 Zero	19 Zero	19 Zero	
	D Input	20 Span	20 Span	20 Span	
	D Input	21 Sample	21 Sample	21 Sample	
DIG 2					
0 1 2	D Output	22 Zero Gas Flow	22 Zero Gas Flow	22 Zero Gas Flow	<u>OC (24vdc if internal SV)</u> <u>OC (24vdc if internal SV)</u> <u>OC (24vdc if internal SV)</u>
	D Output	23 Span Gas Flow	23 Span Gas Flow	23 Span Gas Flow	
	D Output	24 Sample Gas Flow	24 Sample Gas Flow	24 Sample Gas Flow	
3 4 5 6	D Output	25 Local/Remote	25 Local/Remote	25 Local/Remote	<u>TTL-low true</u> <u>TTL-low true</u>
	D Output	26 Read Cal Mode	26 Read Cal Mode	26 Read Cal Mode	
	D Output	27 Reserved	27 Reserved	27 Reserved	
	D Output	28 Reserved	28 Reserved	28 Reserved	

**TABLE D**  
**600 NDIR SERIES I/O CHART**

**28 PIN AUXILIARY CONNECTOR ASSIGNMENTS**

OPTO I/O	Signal Type	601/602/603		<b>LEVELS</b>
ALG 1	Spare	pin #		
COM	A Input	1	GND (analog)	
4	A Input	2	External Analog 1	0-10V
5	A Input	3	External Analog 2	0-10V
6	A Output	4	<b>GND (Isolated analog)</b>	
7	D Output	5	<b>Relay RTN 1</b>	<b>9,10,11,12 use RTN 1</b>
<b>DIG 3</b>	<b>Alarms</b>	<b>Digital</b>		<u>Alarms go OPEN when present</u> <u>Status go CLOSED when active</u>
COM	D Output	6	<b>Relay RTN 2</b>	<b>13,14,15,16 use RTN 2</b>
0	D Output	7	<b>Relay RTN 3</b>	<b>17,17,19,20 use RTN 3</b>
1	D Output	8	<b>Relay RTN 4</b>	<b>21,2728 use RTN 4</b>
2	D Output	9	<b>User Defined NO Relay 1</b>	
3	D Output	10	<b>User Defined NO Relay 2</b>	
4	D Output	11	<b>User Defined NO Relay 3</b>	
5	D Output	12	<b>User Defined NO Relay 4</b>	
6	D Output	13	<b>User Defined NO Relay 5</b>	
7	D Output	14	<b>User Defined NO Relay 6</b>	
8	D Output	15	<b>User Defined NO Relay 7</b>	
9	D Output	16	<b>User Defined NO Relay 8</b>	
10	D Output	17	<b>User Defined NO Relay 9</b>	
11	D Output	18	<b>User Defined NO Relay 10</b>	
12	D Output	19	<b>User Defined NO Relay 11</b>	
13	D Output	20	<b>User Defined NO Relay 12</b>	
14	D Output	21	<b>User Defined NO Relay 13</b>	
15	D Output	22	<b>Reserved Do Not Connect</b>	
<b>DIG 2</b>				
7	D Input	23	Spare	
8	D Input	24	Spare	
9	D Input	25	Spare	
10	D Input	26	<b>Set Remote</b>	
11	D Output	27	<b>User Defined NO Relay 14</b>	
12	D Output	28	<b>User Defined NO Relay 15</b>	