Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1842
CALIBRATION DATE: 19-Feb-26

SBE 37 PRESSURE CALIBRATION DATA
508 psia S/N 0201

COEFFICIENTS:

PA0 = 1.170814e-01
PA1 = 2.414063e-02
PA2 = 2.099687e-09

PTCA0 = 1.884363e+01
PTCA1 = 2.775756e-01
PTCA2 = -2.342336e-03
PTCB0 = 2.497563e+01
PTCB1 = -7.500000e-05
PTCB2 = 0.000000e+00

PRESSURE SPAN CALIBRATION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	TEMPERATURE (°C)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)
14.42	616.3	21.3	14.42	0.00
104.68	4352.9	21.3	104.67	-0.00
204.71	8489.6	21.3	204.65	-0.01
304.70	12625.4	21.3	304.68	-0.00
404.71	16758.7	21.3	404.72	0.00
504.74	20887.0	21.3	504.72	-0.01
404.69	16759.0	21.3	404.73	0.01
304.70	12626.6	21.3	304.71	0.00
204.69	8491.7	21.3	204.70	0.00
104.69	4354.6	21.3	104.71	0.00
14.43	616.5	21.3	14.43	0.00

THERMAL CORRECTION

TEMP (°C)	INSTRUMENT OUTPUT (counts)
32.50	652.33
29.00	652.38
24.00	651.34
18.50	650.64
15.00	649.35
4.50	647.02
1.00	646.50

TEMPERATURE (°C)	SPAN
-5.00	24.98
35.00	24.97

x = instrument output - PTCA0 - PTCA1 * t - PTCA2 * t^2

n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)

pressure (PSIA) = PA0 + PA1 * n + PA2 * n^2

Residual (%FSR) = (computed pressure - true pressure) * 100 / Full Scale Range