

SUNA V2

Saltwater Calibration Verification

SUNA Serial Number:	2558
RMA Number:	

Reference Solution	RS-232 output		Analog output (V)		Analog output (mA)	
	Instrument Output	Reference Range	Instrument Output	Reference Range	Instrument Output	Reference Range
Pure Water (0 μm)	0.2645	0 μm (± 2)	0.4375	0.28 ± 0.1 (0.18–0.38)	5.067	4.76 ± 0.35 (4.41–5.11)
Seawater (0 μm)	-1.078	0 μm (± 2)	0.4081	0.28 ± 0.1 (0.18–0.38)	4.859	4.76 ± 0.35 (4.41–5.11)
Low Nitrate (40 μm)	38.77	40 μm (± 4)	1.870	1.81 ± 0.16 (1.65–1.97)	10.67	10.86 ± 0.61 (10.25–11.47)

Calibration File Name	Date	Technician Initials
SNA2558A.cal	10/27/2025	SS

Final Reference Update Needed? (Yes / No):	Yes
--	------------

Calibration File Name	Date	Technician Initials
SNA2558B.CAL	10/29/2025	CV