



Sea-Bird Scientific  
13431 NE 20<sup>th</sup> Street  
Bellevue, WA 98005  
USA

+1 425-643-9866  
seabird@seabird.com  
www.seabird.com

SENSOR SERIAL NUMBER: 4275  
CALIBRATION DATE: 16-Apr-26

SBE 19plus PRESSURE CALIBRATION DATA  
1450 psia S/N 1569

#### COEFFICIENTS:

|           |               |         |               |
|-----------|---------------|---------|---------------|
| PA0 =     | 9.019983e-01  | PTCA0 = | 5.209769e+05  |
| PA1 =     | 4.435142e-03  | PTCA1 = | 1.618991e+01  |
| PA2 =     | -2.637740e-11 | PTCA2 = | -3.250334e-01 |
| PTEMPA0 = | -6.980627e+01 | PTCB0 = | 2.484725e+01  |
| PTEMPA1 = | 5.024559e+01  | PTCB1 = | -9.500000e-04 |
| PTEMPA2 = | -2.243419e-01 | PTCB2 = | 0.000000e+00  |

#### PRESSURE SPAN CALIBRATION

| PRESSURE (PSIA) | INSTRUMENT OUTPUT (counts) | THERMISTOR OUTPUT (volts) | COMPUTED PRESSURE (PSIA) | RESIDUAL (%FSR) | TEMP (°C)        | THERMISTOR OUTPUT (volts) | INSTRUMENT OUTPUT (counts) |
|-----------------|----------------------------|---------------------------|--------------------------|-----------------|------------------|---------------------------|----------------------------|
| 14.62           | 524274.4                   | 1.8                       | 14.66                    | 0.00            | 32.50            | 2.06                      | 524292.08                  |
| 301.58          | 588930.8                   | 1.8                       | 301.54                   | -0.00           | 29.00            | 1.98                      | 524296.06                  |
| 588.60          | 653670.6                   | 1.8                       | 588.57                   | -0.00           | 24.00            | 1.88                      | 524299.59                  |
| 875.68          | 718483.6                   | 1.8                       | 875.70                   | 0.00            | 18.50            | 1.77                      | 524289.99                  |
| 1162.76         | 783329.4                   | 1.8                       | 1162.76                  | 0.00            | 15.00            | 1.70                      | 524276.96                  |
| 1449.79         | 848210.0                   | 1.8                       | 1449.75                  | -0.00           | 4.50             | 1.49                      | 524180.21                  |
| 1162.80         | 783347.0                   | 1.8                       | 1162.84                  | 0.00            | 1.00             | 1.42                      | 524111.08                  |
| 875.94          | 718543.4                   | 1.8                       | 875.97                   | 0.00            |                  |                           |                            |
| 588.68          | 653693.4                   | 1.8                       | 588.67                   | -0.00           |                  |                           |                            |
| 301.54          | 588923.4                   | 1.8                       | 301.51                   | -0.00           | TEMPERATURE (°C) |                           | SPAN                       |
| 14.63           | 524271.4                   | 1.9                       | 14.64                    | 0.00            | -5.00            |                           | 24.85                      |
|                 |                            |                           |                          |                 | 35.00            |                           | 24.81                      |

y = thermistor output (volts)

$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$

$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$

$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$

pressure (PSIA) = PA0 + PA1 \* n + PA2 \* n

Residual (%FSR) = (computed pressure - true pressure) \* 100 / Full Scale Range

