



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 4275
CALIBRATION DATE: 22-Apr-26

SBE 19plus TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 1.229770e-03
a1 = 2.670368e-04
a2 = -8.559326e-07
a3 = 1.763688e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	673903.034	1.0001	0.0001
4.5000	600247.932	4.4998	-0.0002
14.9999	415479.898	15.0001	0.0002
18.5000	365441.220	18.5000	0.0000
24.0000	297253.678	23.9997	-0.0003
29.0000	245169.661	29.0001	0.0001
32.5000	213652.085	32.5000	0.0000

n = Instrument Output (counts)

MV = (n - 524288) / 1.6e+007

R = (MV * 2.900e+009 + 1.024e+008) / (2.048e+004 - MV * 2.0e+005)

Temperature ITS-90 (°C) = 1 / {a0 + a1[ln(R)] + a2[ln(R)] + a3[ln(R)]} - 273.15

Residual (°C) = instrument temperature - bath temperature

