



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0027
CALIBRATION DATE: 10-Mar-26

Prawler CTD TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -6.196733e-05
a1 = 2.973726e-04
a2 = -3.719580e-06
a3 = 1.809054e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	571198.0	1.0000	0.0000
4.5000	488390.4	4.5000	-0.0000
15.0000	311368.6	14.9999	-0.0001
18.5000	269668.2	18.5000	0.0000
24.0000	216425.0	24.0002	0.0002
29.0000	178295.6	28.9997	-0.0003
32.5000	156193.8	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

