



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0045
CALIBRATION DATE: 10-Mar-26

Prawler CTD CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.023671e+00
h = 1.489061e-01
i = -2.970557e-04
j = 4.469916e-05

CPcor = -9.5700e-008
CTcor = 3.2500e-006
WBOTC = 9.5198e-08

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2626.11	0.00000	0.00000
1.0000	34.5843	2.95802	5176.95	2.95802	0.00000
4.5000	34.5665	3.26347	5371.27	3.26347	0.00000
15.0000	34.5298	4.24017	5949.40	4.24016	-0.00001
18.5000	34.5227	4.58360	6139.50	4.58361	0.00001
24.0000	34.5159	5.13886	6434.65	5.13885	-0.00000
29.0000	34.5124	5.65812	6698.60	5.65812	0.00000
32.5000	34.5099	6.02857	6879.03	6.02549	-0.00308

$f = \text{Instrument Output(Hz)} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

$t = \text{temperature (°C)}$; $p = \text{pressure (decibars)}$; $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

$\text{Residual (Siemens/meter)} = \text{instrument conductivity} - \text{bath conductivity}$

