



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 0045
CALIBRATION DATE: 10-Mar-26

Prawler CTD TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -2.329492e-04
a1 = 3.323239e-04
a2 = -6.290958e-06
a3 = 2.488942e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	573404.6	1.0000	-0.0000
4.5000	490958.8	4.5000	0.0000
15.0000	314252.8	15.0001	0.0001
18.5000	272516.4	18.4999	-0.0001
24.0000	219145.4	23.9998	-0.0002
29.0000	180849.2	29.0003	0.0003
32.5000	158631.2	32.4998	-0.0002

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1/\{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

