



Sea-Bird Scientific
13431 NE 20th Street
Bellevue, WA 98005
USA

+1 425-643-9866
seabird@seabird.com
www.seabird.com

SENSOR SERIAL NUMBER: 1433
CALIBRATION DATE: 11-Feb-26

SBE 39 TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 7.225324e-05
a1 = 2.677905e-04
a2 = -1.998394e-06
a3 = 1.394191e-07

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
-1.5000	768014.2	-1.5000	0.0000
1.0000	684071.2	1.0000	-0.0000
4.5000	583455.4	4.5000	-0.0000
8.0000	499313.0	8.0000	-0.0000
11.5000	428699.4	11.5000	0.0000
15.0000	369239.9	15.0000	0.0000
18.5000	319006.7	18.5000	0.0000
22.0000	276432.6	22.0000	-0.0000
25.5000	240236.8	25.4999	-0.0001
29.0000	209369.8	29.0000	-0.0000
32.5000	182969.3	32.5001	0.0001

n = Instrument Output (counts)

Temperature ITS-90 (°C) = $1/\{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$

Residual (°C) = instrument temperature - bath temperature

